


User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 12/03/2001 Time: 11:00:10

Sample: AD25601
 Analysis: \$LP525 Organic Extractables
 Units: ug
 Started: 11/13/01 10:58 Ended: 11/13/01 10:58 Analyst: TB
 Page: 1


	Component Name	Viol	Result
1	DFTPP		PASSED
2	ENDRIN *		55
3	DDT		3

← flag

Chromatogram Plot File: E:\DFTB1113 Date: Nov-13-1991 22:17:16
Comment: BOHLK; METHOD 525; INST 204; 11/13/01; DFTPP/ENDRIN
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 5369056



Chromatogram Plot File: E:\DFTB1113 Date: Nov-13-1991 22:17:16
Comment: BOHLK; METHOD 525; INST 204; 11/13/01; DFTPP/ENDRIN
Scan No: 1080 Retention Time: 18:00 RIC: 52048 Mass Range: 47 - 149
Plotted: 1080 to 1120 Range: 1 to 1859 100% = 5369056



Mass List

File: E:\DFTB1113 Date: Nov-13-1991 22:17:16

Comment: BOHLK; METHOD 525; INST 204; 11/13/01; DFTPP/ENDRIN

Scan No: 1101 Retention Time: 18:21 RIC: 5369056 Mass Range: 47 - 444

Peaks: 162 Mass of Base Peak: 442 Int: 806481 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
47	3,395	0.42	84	6,481	0.80	125	4,012	0.50
49	19,135	2.37	85	4,320	0.54	127	161,111	19.98
50	133,641	16.57	86	9,567	1.19	128	16,358	2.03
51	355,555	44.09	91	5,246	0.65	129	87,345	10.83
52	13,271	1.65	92	5,246	0.65	130	8,950	1.11
53	2,469	0.31	93	41,975	5.20	134	3,395	0.42
55	5,246	0.65	97	2,160	0.27	135	9,567	1.19
56	7,716	0.96	98	36,728	4.55	136	3,395	0.42
57	14,506	1.80	99	16,975	2.10	137	4,320	0.54
61	4,629	0.57	101	8,641	1.07	141	15,123	1.88
63	9,876	1.22	103	5,864	0.73	142	4,938	0.61
65	3,395	0.42	104	8,641	1.07	143	3,086	0.38
68	5,864	0.73	105	7,407	0.92	147	8,641	1.07
69	254,938	31.61	106	3,395	0.42	148	13,888	1.72
73	4,012	0.50	107	82,098	10.18	149	3,703	0.46
74	48,148	5.97	108	10,493	1.30	155	5,246	0.65
75	36,728	4.55	109	2,777	0.34	156	8,641	1.07
76	12,962	1.61	110	135,185	16.76	157	2,469	0.31
77	246,604	30.58	111	18,827	2.33	160	4,012	0.50
78	13,888	1.72	112	2,469	0.31	161	4,320	0.54
79	29,012	3.60	117	108,641	13.47	165	3,086	0.38
80	12,962	1.61	118	4,938	0.61	166	5,864	0.73
81	18,827	2.33	122	7,098	0.88	167	25,925	3.21
82	4,012	0.50	123	7,098	0.88	168	11,111	1.38
83	3,395	0.42	124	4,938	0.61	172	3,395	0.42

Mass List

File: E:\DFTB1113 Date: Nov-13-1991 22:17:16

Comment: BOHLK; METHOD 525; INST 204; 11/13/01; DFTPP/ENDRIN

Scan No: 1101 Retention Time: 18:21 RIC: 5369056 Mass Range: 47 - 444

Peaks: 162 Mass of Base Peak: 442 Int: 806481 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
173	2,160	0.27	207	14,506	1.80	274	36,111	4.48
174	5,555	0.69	208	3,395	0.42	275	163,888	20.32 ✓
175	8,024	0.99	211	3,395	0.42	276	19,753	2.45
177	6,172	0.77	217	26,851	3.33	277	16,358	2.03
178	2,469	0.31	218	4,629	0.57	293	4,320	0.54
179	12,345	1.53	221	11,419	1.42	296	39,506	4.90
180	10,185	1.26	222	4,629	0.57	297	5,864	0.73
181	4,629	0.57	223	4,938	0.61	303	3,703	0.46
185	10,185	1.26	224	56,481	7.00	314	3,086	0.38
186	72,222	8.96	225	15,123	1.88	315	6,790	0.84
187	15,123	1.88	227	26,543	3.29	316	3,703	0.46
188	3,086	0.38	228	5,246	0.65	323	19,753	2.45
189	5,864	0.73	229	4,320	0.54	324	4,012	0.50
191	3,086	0.38	243	3,395	0.42	327	3,086	0.38
192	4,629	0.57	244	33,333	4.13	334	10,493	1.30
193	6,172	0.77	245	5,864	0.73	335	3,703	0.46
196	6,790	0.84	246	11,419	1.42	346	4,320	0.54
197	4,938	0.61	249	2,469	0.31	352	4,012	0.50
198	541,666	67.16 ✓	255	325,308	40.34	353	5,555	0.69
199	33,950	4.21 ✓	256	43,518	5.40	354	13,271	1.65
200	2,469	0.31	257	4,938	0.61	365	48,456	6.01 ✓
201	2,469	0.31	258	16,975	2.10	366	6,481	0.80
204	15,740	1.95	259	3,703	0.46	370	1,543	0.19
205	22,839	2.83	265	6,481	0.80	372	12,962	1.61
206	97,530	12.09	273	8,950	1.11	373	4,012	0.50

Mass List

File: E:\DFTB1113 Date: Nov-13-1991 22:17:16

Comment: BOHLK; METHOD 525; INST 204; 11/13/01; DFTPP/ENDRIN

Scan No: 1101 Retention Time: 18:21 RIC: 5369056 Mass Range: 47 - 444

Peaks: 162 Mass of Base Peak: 442 Int: 806481 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
173	2,160	0.27	207	14,506	1.80	274	36,111	4.48
174	5,555	0.69	208	3,395	0.42	275	163,888	20.32
175	8,024	0.99	211	3,395	0.42	276	19,753	2.45
177	6,172	0.77	217	26,851	3.33	277	16,358	2.03
178	2,469	0.31	218	4,629	0.57	293	4,320	0.54
179	12,345	1.53	221	11,419	1.42	296	39,506	4.90
180	10,185	1.26	222	4,629	0.57	297	5,864	0.73
181	4,629	0.57	223	4,938	0.61	303	3,703	0.46
185	10,185	1.26	224	56,481	7.00	314	3,086	0.38
186	72,222	8.96	225	15,123	1.88	315	6,790	0.84
187	15,123	1.88	227	26,543	3.29	316	3,703	0.46
188	3,086	0.38	228	5,246	0.65	323	19,753	2.45
189	5,864	0.73	229	4,320	0.54	324	4,012	0.50
191	3,086	0.38	243	3,395	0.42	327	3,086	0.38
192	4,629	0.57	244	33,333	4.13	334	10,493	1.30
193	6,172	0.77	245	5,864	0.73	335	3,703	0.46
196	6,790	0.84	246	11,419	1.42	346	4,320	0.54
197	4,938	0.61	249	2,469	0.31	352	4,012	0.50
198	541,666	67.16	255	325,308	40.34	353	5,555	0.69
199	33,950	4.21	256	43,518	5.40	354	13,271	1.65
200	2,469	0.31	257	4,938	0.61	365	48,456	6.01
201	2,469	0.31	258	16,975	2.10	366	6,481	0.80
204	15,740	1.95	259	3,703	0.46	370	1,543	0.19
205	22,839	2.83	265	6,481	0.80	372	12,962	1.61
206	97,530	12.09	273	8,950	1.11	373	4,012	0.50

Mass List

File: E:\DFTB1113 Date: Nov-13-1991 22:17:16

Comment: BOHLK; METHOD 525; INST 204; 11/13/01; DFTPP/ENDRIN

Scan No: 1101 Retention Time: 18:21 RIC: 5369056 Mass Range: 47 - 444

Peaks: 162 Mass of Base Peak: 442 Int: 806481 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
189	5,864	0.73	229	4,320	0.54	324	4,012	0.50
191	3,086	0.38	243	3,395	0.42	327	3,086	0.38
192	4,629	0.57	244	33,333	4.13	334	10,493	1.30
193	6,172	0.77	245	5,864	0.73	335	3,703	0.46
196	6,790	0.84	246	11,419	1.42	346	4,320	0.54
197	4,938	0.61	249	2,469	0.31	352	4,012	0.50
198	541,666	67.16	255	325,308	40.34	353	5,555	0.69
199	33,950	4.21	256	43,518	5.40	354	13,271	1.65
200	2,469	0.31	257	4,938	0.61	365	48,456	6.01
201	2,469	0.31	258	16,975	2.10	366	6,481	0.80
204	15,740	1.95	259	3,703	0.46	370	1,543	0.19
205	22,839	2.83	265	6,481	0.80	372	12,962	1.61
206	97,530	12.09	273	8,950	1.11	373	4,012	0.50
207	14,506	1.80	274	36,111	4.48	383	2,777	0.34
208	3,395	0.42	275	163,888	20.32	402	3,086	0.38
211	3,395	0.42	276	19,753	2.45	403	5,246	0.65
217	26,851	3.33	277	16,358	2.03	404	3,086	0.38
218	4,629	0.57	293	4,320	0.54	421	3,703	0.46
221	11,419	1.42	296	39,506	4.90	422	4,629	0.57
222	4,629	0.57	297	5,864	0.73	423	39,506	4.90
223	4,938	0.61	303	3,703	0.46	424	9,259	1.15
224	56,481	7.00	314	3,086	0.38	441	122,222	15.15
225	15,123	1.88	315	6,790	0.84	442	806,481	100.00
227	26,543	3.29	316	3,703	0.46	443	133,641	16.57
228	5,246	0.65	323	19,753	2.45	444	9,259	1.15

Chromatogram Plot

File: E:\DFTB1113

Date: Nov-13-1991 22:17:16

Comment: BOHLK; METHOD 525; INST 204; 11/13/01; DFTPP/ENDRIN

Scan No: 1280

Retention Time: 21:20


RIC: 45159

Mass Range: 45 - 281


Plotted: 1260 to 1280

Range: 1 to 1859


100% = 5369056



Chromatogram Plot File: E:\DFTB1113 Date: Nov-13-1991 22:17:16
 Comment: BOHLK; METHOD 525; INST 204; 11/13/01; DFTPP/ENDRIN
 Scan No: 1302 Retention Time: 21:42 RIC: 58469 Mass Range: 45 - 429
 Plotted: 1282 to 1302 Range: 1 to 1859 100% = 5369056



Chromatogram Plot File: E:\DFTB1113 Date: Nov-13-1991 22:17:16
Comment: BOHLK; METHOD 525; INST 204; 11/13/01; DFTPP/ENDRIN
Scan No: 1367 Retention Time: 22:47 RIC: 45365 Mass Range: 45 - 355
Plotted: 1347 to 1367 Range: 1 to 1859 100% = 5369056



Chromatogram Plot

File: E:\DDTB1113

Date: Nov-13-1991 22:51:40

Comment: BOHLK; METHOD 525; INST 204; 11/13/01; DDT

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0

Plotted: 1 to 1860

Range: 1 to 1860

100% = 6719856



Chromatogram Plot

File: E:\DDTB1113

Date: Nov-13-1991 22:51:40

Comment: BOHLK; METHOD 525; INST 204; 11/13/01; DDT

Scan No: 1259

Retention Time: 20:59

RIC: 37508


Mass Range: 47 - 207

Plotted: 1239 to 1259

Range: 1 to 1860

100% = 6719856

5.15%



Chromatogram Plot

File: E:\DDTB1113

Date: Nov-13-1991 22:51:40

Comment: BOHLK; METHOD 525; INST 204; 11/13/01; DDT

Scan No: 1292

Retention Time: 21:32


RIC: 43849

Mass Range: 45 - 237

Plotted: 1272 to 1292

Range: 1 to 1860

100% = 6719856



Chromatogram Plot

File: E:\DDTB1113 Date: Nov-13-1991 22:51:40

Comment: BOHLK; METHOD 525; INST 204; 11/13/01; DDT

Scan No: 1324 Retention Time: 22:04 RIC: 52839 Mass Range: 45 - 355

Plotted: 1304 to 1324

Range: 1 to 1860

100% = 6719856

