


User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 12/13/2001 Time: 18:03:10

Sample: AD25945
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 11/1/01 18:00 Ended: 11/17/01 18:00 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		2.0		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		4.6		1.0
20	Surr_Triphenyl phosphate		5.6		1.0
21	Surr_Perylene-d12	LSPEC	3.3		1.0

Chromatogram Plot File: C:\25945B Date: Nov-17-1991 20:53:56
Comment: BOHLK; METHOD 525; INST 204; 11/17/01; SAMPLE 25945
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 6112634



Integration Report

Quanfile: 25945B

Quan Entries: 13

Comment: BOHLK; METHOD 525; INST 204; 11/17/01; SAMPLE 25945

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Node	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:18	738	Auto	2,077,743	976,546
2	phenanthrene d-10 (IS)	16:56	1016	Auto	3,255,319	1,226,887
3	chrysene d-12(IS)	22:41	1361	Auto	2,028,876	914,735
4	p-terphenyl d-14(IS)	20:51	1251	Auto	3,970	2,393
5	acenaphthene d-10(SURR)	12:18	738	Auto	2,077,743	976,546
6	phenanthrene d-10(SURR)	16:56	1016	Auto	3,255,319	1,226,887
7	chrysene d-12 (SURR)	22:41	1361	Auto	2,028,876	914,735
8	1,3-dimethyl-2-nitrobenzene	7:58	478	Auto	1,199,362	619,227
9	triphenyl phosphate (S)	22:06	1326	Auto	1,624,987	685,850
10	perylene d-12 (SURR)	26:59	1619	Auto	817,810	141,938
11	bis(2-ethylhexyl)adipate	22:01	1321	Auto	46,730	2,938
12	bis(2-ethylhexyl)phthalate	22:59	1379	Auto	1,475,566	383,692
13	2,6-dinitrotoluene	11:58	718	Auto	75,096	33,237

revised IS/ surr's

Quantitation Report

Quanfile: 25945B

Quan Entries: 13

Comment: BOHLK; METHOD 525; INST 204; 11/17/01; SAMPLE 25945

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	738	12:18	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1016	16:56	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1361	22:41	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1251	20:51	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	738	12:18	BB	3443.202	UG/L
6	phenanthrene d-10(SURR	A	1016	16:56	BV	3518.199	UG/L
7	chrysene d-12 (SURR)	A	1361	22:41	BB	3287.501	UG/L
8	1,3-dimethyl-2-nitrobe	A	478	7:58	BV	4.634	UG/L
9	triphenyl phosphate (S	A	1326	22:06	BB	5.587	UG/L
10	perylene d-12 (SURR)	A	1619	26:59	BB	3.340	UG/L
16	bis(2-ethylhexyl)adipa	A	1321	22:01	MM	0.191	UG/L
17	bis(2-ethylhexyl)phtha	A	1379	22:59	BB	1.976	UG/L
20	2,6-dinitrotoluene	A	718	11:58	BB	0.559	UG/L

Comments for Sample AD25945 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-15-2001 Time: 11:18:52

The recovery for 1 of the 3 Surrogate Standards in this sample is below its acceptance range. Recoveries for 3 of the 18 compounds in the MDL and LCS Standards are outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the Spiked Sample are outside of their acceptance ranges. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank is 1.2 ug/L.