


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 12/13/2001 Time: 18:11:12

Sample: AD25949
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/1/01 18:09 Ended: 11/17/01 18:09 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		2.5		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.1		1.0
20	Surr_Triphenyl phosphate		5.8		1.0
21	Surr_Perylene-d12	LSPEC	2.7		1.0

Chromatogram Plot File: C:\25949F Date: Nov-17-1991 23:11:28
Comment: BOHLK; METHOD 525; INST 204; 11/17/01; SAMPLE 25949
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 6557762



Integration Report Quanfile: 25949F Quan Entries: 16
 Comment: BOHLK; METHOD 525; INST 204; 11/17/01; SAMPLE 25949
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:17	737	Auto	1,668,444	1,002,835
2	phenanthrene d-10 (IS)	16:55	1015	Auto	3,095,275	1,258,448
3	chrysene d-12(IS)	22:41	1361	Auto	1,391,582	503,363
4	p-terphenyl d-14(IS)	20:50	1250	Auto	3,255,467	2,064,013
5	acenaphthene d-10(SURR)	12:17	737	Auto	1,668,444	1,002,835
6	phenanthrene d-10(SURR)	16:55	1015	Auto	3,095,275	1,258,448
7	chrysene d-12 (SURR)	22:41	1361	Auto	1,391,582	503,363
8	1,3-dimethyl-2-nitrobenzene	7:58	478	Auto	1,060,173	580,740
9	triphenyl phosphate (S)	22:05	1325	Auto	1,158,696	413,741
10	perylene d-12 (SURR)	26:58	1618	Auto	455,886	72,253
11	hexachlorobenzene	15:20	920	Auto	1,808	1,155
12	simazine	16:07	967	Auto	4,118	1,497
13	bis(2-ethylhexyl)adipate	21:58	1318	Auto	75,068	16,436
14	bis(2-ethylhexyl)phthalate	22:58	1378	Auto	1,240,688	390,460
15	2,6-dinitrotoluene	11:57	717	Auto	62,318	30,830
16	2,4-dinitrotoluene	12:54	774	Auto	2,293	1,324

reversed IS/surrs

Quantitation Report

Quanfile: 25949F

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 11/17/01; SAMPLE 25949

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	737	12:17	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1015	16:55	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1361	22:41	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1250	20:50	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	737	12:17	BB	3.372	UG/L
6	phenanthrene d-10(SURR	A	1015	16:55	BV	4.000	UG/L
7	chrysene d-12 (SURR)	A	1361	22:41	BB	2.750	UG/L
8	1,3-dimethyl-2-nitrobe	A	478	7:58	BV	5.101	UG/L
9	triphenyl phosphate (S	A	1325	22:05	BB	5.000	UG/L
10	perylene d-12 (SURR)	A	1618	26:58	BB	2.715	UG/L
12	hexachlorobenzene	A	920	15:20	BB	0.000	UG/L
13	simazine	A	967	16:07	BB	0.047	UG/L
16	bis(2-ethylhexyl)adipa	A	1318	21:58	MM	0.449	UG/L
17	bis(2-ethylhexyl)phtha	A	1378	22:58	BV	2.465	UG/L
20	2,6-dinitrotoluene	A	717	11:57	BB	0.578	UG/L
21	2,4-dinitrotoluene	A	774	12:54	MN	0.020	UG/L

Comments for Sample AD25949 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-15-2001 Time: 11:29:26

The recovery for 1 of the 3 Surrogate Standards in this sample is below its acceptance range. Recoveries for 3 of the 18 compounds in the MDL and LCS Standards are outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the Spiked Sample are outside of their acceptance ranges. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank is 1.2 ug/L.