


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 12/13/2001 Time: 18:35:17

Sample: AD26080
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/1/01 18:33 Ended: 11/17/01 18:33 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.84		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		4.7		1.0
20	Surr_Triphenyl phosphate		5.5		1.0
21	Surr_Perylene-d12		3.5		1.0

Chromatogram Plot File: E:\26080P Date: Nov-18-1991 10:05:01
Comment: BOHLK; METHOD 525; INST 204; 11/17/01; SAMPLE 26080
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 5876395



Integration Report

Quanfile: 26080P

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/17/01; SAMPLE 26080

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:18	738	Auto	1,942,149	909,615
2	phenanthrene d-10 (IS)	16:56	1016	Auto	3,070,272	1,121,540
3	chrysene d-12 (IS)	22:41	1361	Auto	1,929,894	811,292
4	p-terphenyl d-14 (IS)	20:50	1250	Auto	3,167,843	1,783,963
5	acenaphthene d-10 (SURR)	12:18	738	Auto	1,942,149	909,615
6	phenanthrene d-10 (SURR)	16:56	1016	Auto	3,070,272	1,121,540
7	chrysene d-12 (SURR)	22:41	1361	Auto	1,929,894	811,292
8	1,3-dimethyl-2-nitrobenzene	7:58	478	Auto	1,148,876	605,693
9	triphenyl phosphate (S)	22:06	1326	Auto	1,534,143	638,366
10	perylene d-12 (SURR)	26:59	1619	Auto	806,796	138,944
11	simazine	16:09	969	Auto	0	0
12	bis(2-ethylhexyl)adipate	22:03	1323	Auto	47,143	3,748
13	bis(2-ethylhexyl)phthalate	22:58	1378	Auto	627,603	181,150
14	2,6-dinitrotoluene	11:57	717	Auto	25,522	11,288
15	2,4-dinitrotoluene	13:05	785	Auto	955	955

reviewed IS/SURR's

Quantitation Report

Quanfile: 26080P

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/17/01; SAMPLE 26080

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	738	12:18	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1016	16:56	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1361	22:41	BV	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1250	20:50	BB	5.000	UG/L
5	acenaphthene d-10 (SURR	A	738	12:18	BV	4.034	UG/L
6	phenanthrene d-10 (SURR	A	1016	16:56	BV	4.159	UG/L
7	chrysene d-12 (SURR)	A	1361	22:41	BV	3.919	UG/L
8	1,3-dimethyl-2-nitroben	A	478	7:58	BB	4.749	UG/L
9	triphenyl phosphate (S	A	1326	22:06	BB	5.545	UG/L
10	perylene d-12 (SURR)	A	1619	26:59	BB	3.464	UG/L
13	simazine	A	969	16:09	BB	0.001	UG/L
16	bis(2-ethylhexyl)adipa	A	1323	22:03	MM	0.203	UG/L
17	bis(2-ethylhexyl)phtha	A	1378	22:58	BV	0.845	UG/L
20	2,6-dinitrotoluene	A	717	11:57	BB	0.201	UG/L
21	2,4-dinitrotoluene	A	785	13:05	BB	0.007	UG/L

Comments for Sample AD26080 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-15-2001 Time: 11:44:55

Recoveries for 3 of the 18 compounds in the MDL and LCS Standards are outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the Spiked Sample are outside of their acceptance ranges. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank is 1.2 ug/L.