


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/30/2001 Time: 09:42:32

Sample: AD25738
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/1/01 09:37 Ended: 11/16/01 09:37 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.64		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.8		1.0
20	Surr_Triphenyl phosphate		5.7		1.0
21	Surr_Perylene-d12		3.3		1.0

Chromatogram Plot File: E:\25738D Date: Nov-17-1991 00:09:10
Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25738
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 4927312



Integration Report Quanfile: 25738D Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25738
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:18	738	Auto	1,486,330	783,407
2	phenanthrene d-10 (IS)	16:56	1016	Auto	2,375,817	888,071
3	chrysene d-12 (IS)	22:41	1361	Auto	1,558,659	664,832
4	p-terphenyl d-14 (IS)	20:50	1250	Auto	2,788,227	1,231,359
5	acenaphthene d-10 (SURR)	12:18	738	Auto	1,486,330	783,407
6	phenanthrene d-10 (SURR)	16:56	1016	Auto	2,375,817	888,071
7	chrysene d-12 (SURR)	22:41	1361	Auto	1,558,659	664,832
8	1,3-dimethyl-2-nitrobenzene	7:58	478	Auto	897,556	443,704
9	triphenyl phosphate (S)	22:06	1326	Auto	1,282,987	523,856
10	perylene d-12 (SURR)	26:59	1619	Auto	613,076	103,431
11	hexachlorocyclopentadiene	10:09	609	Auto	1,238	1,238
12	bis(2-ethylhexyl)adipate	22:06	1326	Auto	31,218	2,897
13	bis(2-ethylhexyl)phthalate	22:59	1379	Auto	386,997	98,481
14	2,6-dinitrotoluene	11:58	718	Auto	84,462	37,277

Reviewed 12/5/01

Quantitation Report

Quanfile: 25738D

Quan Entries: 14

Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25738

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	738	12:18	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1016	16:56	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1361	22:41	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1250	20:50	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	738	12:18	BB	3.508	UG/L
6	phenanthrene d-10(SURR)	A	1016	16:56	BV	3.656	UG/L
7	chrysene d-12 (SURR)	A	1361	22:41	BB	3.597	UG/L
8	1,3-dimethyl-2-nitrobe	A	478	7:58	BV	4.848	UG/L
9	triphenyl phosphate (S	A	1326	22:06	BV	5.742	UG/L
10	perylene d-12 (SURR)	A	1619	26:59	BB	3.259	UG/L
11	hexachlorocyclopentadi	A	609	10:09	BB	0.018	UG/L
16	bis(2-ethylhexyl)adipa	A	1326	22:06	MM	0.166	UG/L
17	bis(2-ethylhexyl)phtha	A	1379	22:59	BB	0.640	UG/L
20	2,6-dinitrotoluene	A	718	11:58	BB	0.888	UG/L

Comments for Sample AD25738 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-11-2001 Time: 11:02:06

Recoveries for 2 of the 18 compounds in the MDL Standard were above thier acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The recovery for 1 of the 3 Surrogate Standards in this sample was below its acceptance range.