


User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 11/30/2001 Time: 10:44:21

Sample: AD25831
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 11/1/01 09:55 Ended: 11/16/01 09:55 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.98		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.7		1.0
20	Surr_Triphenyl phosphate		5.7		1.0
21	Surr_Perylene-d12		3.0		1.0

Chromatogram Plot File: E:\25831G Date: Nov-17-1991 01:52:28
Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25831
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 10211533



Integration Report

Quanfile: 25831G

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25831

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:18	738	Auto	1,913,764	1,273,717
2	phenanthrene d-10 (IS)	16:56	1016	Auto	2,861,464	1,164,825
3	chrysene d-12(IS)	22:41	1361	Auto	1,933,393	816,278
4	p-terphenyl d-14(IS)	20:51	1251	Auto	3,344,738	2,306,275
5	acenaphthene d-10(SURR)	12:18	738	Auto	1,913,764	1,273,717
6	phenanthrene d-10(SURR)	16:56	1016	Auto	2,861,464	1,164,825
7	chrysene d-12 (SURR)	22:41	1361	Auto	1,933,393	816,278
8	1,3-dimethyl-2-nitrobe	7:59	479	Auto	1,110,775	527,258
9	triphenyl phosphate (S	22:06	1326	Auto	1,576,382	748,216
10	perylene d-12 (SURR)	26:59	1619	Auto	702,486	115,093
11	hexachlorobenzene	15:22	922	Auto	7,601	3,187
12	bis(2-ethylhexyl)adipa	22:03	1323	Auto	45,945	3,554
13	bis(2-ethylhexyl)phtha	22:59	1379	Auto	724,024	215,468
14	2,6-dinitrotoluene	11:58	718	Auto	99,945	52,784
15	2,4-dinitrotoluene	12:55	775	Auto	2,220	1,341
16	4,4'-dde	20:46	1246	Auto	5,596	3,339

Reviewed IS/SUR

Quantitation Report

Quanfile: 25831G

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25831

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	738	12:18	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1016	16:56	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1361	22:41	BB	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1251	20:51	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	738	12:18	BV	3.765	UG/L
6	phenanthrene d-10 (SURR)	A	1016	16:56	BV	3.671	UG/L
7	chrysene d-12 (SURR)	A	1361	22:41	BB	3.719	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	479	7:59	BV	4.660	UG/L
9	triphenyl phosphate (S)	A	1326	22:06	BV	5.687	UG/L
10	perylene d-12 (SURR)	A	1619	26:59	BB	3.011	UG/L
12	hexachlorobenzene	A	922	15:22	BB	0.029	UG/L
16	bis(2-ethylhexyl)adipate	A	1323	22:03	MM	0.197	UG/L
17	bis(2-ethylhexyl)phthalate	A	1379	22:59	BB	0.979	UG/L
20	2,6-dinitrotoluene	A	718	11:58	BB	0.814	UG/L
21	2,4-dinitrotoluene	A	775	12:55	BB	0.017	UG/L
28	4,4'-dichlorodiphenyl ether	A	1246	20:46	BB	0.023	UG/L

Comments for Sample AD25831 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-11-2001 Time: 11:08:28

Recoveries for 2 of the 18 compounds in the MDL Standard were above thier acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The recovery for 1 of the 3 Surrogate Standards in this sample was below its acceptance range.