


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 12/08/2001 Time: 17:10:13

Sample: AD25839
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/1/01 17:08 Ended: 11/16/01 17:08 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.94		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		4.7		1.0
20	Surr_Triphenyl phosphate		5.5		1.0
21	Surr_Perylene-d12		3.0		1.0

Chromatogram Plot File: C:\258391 Date: Nov-17-1991 04:44:35
 Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25839
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 12455652



Integration Report

Quanfile: 25839I

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25839

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:18	738	Auto	2,101,382	1,336,666
2	phenanthrene d-10 (IS)	16:56	1016	Auto	3,198,414	1,304,851
3	chrysene d-12 (IS)	22:41	1361	Auto	2,185,628	951,425
4	p-terphenyl d-14 (IS)	20:51	1251	Auto	6,176,569	4,156,962
5	acenaphthene d-10 (SURR)	12:18	738	Auto	2,101,382	1,336,666
6	phenanthrene d-10 (SURR)	16:56	1016	Auto	3,198,414	1,304,851
7	chrysene d-12 (SURR)	22:41	1361	Auto	2,185,628	951,425
8	1,3-dimethyl-2-nitrobenzene	7:58	478	Auto	1,232,444	510,783
9	triphenyl phosphate (S)	22:06	1326	Auto	1,715,055	818,078
10	perylene d-12 (SURR)	26:59	1619	Auto	791,207	144,837
11	hexachlorocyclopentadiene	10:09	609	Auto	889	889
12	hexachlorobenzene	15:22	922	Auto	9,533	3,638
13	bis(2-ethylhexyl)adipate	21:59	1319	Auto	23,267	3,175
14	bis(2-ethylhexyl)phthalate	22:59	1379	Auto	788,293	231,380
15	2,6-dinitrotoluene	11:58	718	Auto	38,532	19,945
16	4,4'-dichlorodiphenyl ether	20:46	1246	Auto	7,092	4,465

Reviewed IS/SUR

Quantitation Report

Quanfile: 25839I

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25839

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	738	12:18	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1016	16:56	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1361	22:41	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1251	20:51	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	738	12:18	BB	2.239	UG/L
6	phenanthrene d-10(SURR)	A	1016	16:56	BV	2.222	UG/L
7	chrysene d-12 (SURR)	A	1361	22:41	BB	2.277	UG/L
8	1,3-dimethyl-2-nitrobe	A	478	7:58	BV	4.709	UG/L
9	triphenyl phosphate (S	A	1326	22:06	BB	5.474	UG/L
10	perylene d-12 (SURR)	A	1619	26:59	BB	3.000	UG/L
11	hexachlorocyclopentadi	A	609	10:09	BB	0.009	UG/L
12	hexachlorobenzene	A	922	15:22	BB	0.034	UG/L
16	bis(2-ethylhexyl)adipa	A	1319	21:59	MM	0.089	UG/L
17	bis(2-ethylhexyl)phtha	A	1379	22:59	BV	0.941	UG/L
20	2,6-dinitrotoluene	A	718	11:58	BB	0.281	UG/L
28	4,4'-dde	A	1246	20:46	BB	0.026	UG/L

Comments for Sample AD25839 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-11-2001 Time: 11:10:00

Recoveries for 2 of the 18 compounds in the MDL Standard were above thier acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The recovery for 1 of the 3 Surrogate Standards in this sample was below its acceptance range.