


User: Bohk, Ted
Westchester Dept. of Labs & Research
Date: 11/30/2001 Time: 11:06:31

Sample: AD25873
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/1/01 11:04 Ended: 11/16/01 11:04 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		2.1		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen*		3.0		1.0
20	Surr_Triphenyl phosphate		5.6		1.0
21	Surr_Perylene-d12		2.9		1.0

← flag

Chromatogram Plot File: C:\258730 Date: Nov-17-1991 11:47:18
Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25873
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 5785188



Integration Report Quanfile: 258730 Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25873
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:18	738	Auto	1,753,391	915,841
2	phenanthrene d-10 (IS)	16:56	1016	Auto	2,803,238	1,062,164
3	chrysene d-12 (IS)	22:41	1361	Auto	2,021,674	889,689
4	p-terphenyl d-14 (IS)	20:51	1251	Auto	3,520,265	1,620,760
5	acenaphthene d-10 (SURR)	12:18	738	Auto	1,753,391	915,841
6	phenanthrene d-10 (SURR)	16:56	1016	Auto	2,803,238	1,062,164
7	chrysene d-12 (SURR)	22:41	1361	Auto	2,021,674	889,689
8	1,3-dimethyl-2-nitrobenzene	7:58	478	Auto	657,221	316,331
9	triphenyl phosphate (S)	22:06	1326	Auto	1,609,058	691,522
10	perylene d-12 (SURR)	26:59	1619	Auto	699,071	121,207
11	bis(2-ethylhexyl)adipate	21:58	1318	Auto	33,473	3,269
12	bis(2-ethylhexyl)phthalate	22:58	1378	Auto	1,567,681	454,865
13	2,6-dinitrotoluene	11:58	718	Auto	73,980	32,941
14	4,4'-dichlorodiphenyl ether	20:46	1246	Auto	2,331	2,331

Reviewed 15/Sun

Quantitation Report

Quanfile: 258730

Quan Entries: 14

Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25873

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	738	12:18	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1016	16:56	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1361	22:41	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1251	20:51	BV	5.000	UG/L
5	acenaphthene d-10(SURR)	A	738	12:18	BB	3.277	UG/L
6	phenanthrene d-10(SURR)	A	1016	16:56	BV	3.417	UG/L
7	chrysene d-12 (SURR)	A	1361	22:41	BB	3.695	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	478	7:58	BB	3.009	UG/L
9	triphenyl phosphate (S)	A	1326	22:06	BB	5.552	UG/L
10	perylene d-12 (SURR)	A	1619	26:59	BB	2.865	UG/L
16	bis(2-ethylhexyl)adipate	A	1318	21:58	MM	0.138	UG/L
17	bis(2-ethylhexyl)phthalate	A	1378	22:58	BB	2.118	UG/L
20	2,6-dinitrotoluene	A	718	11:58	BB	0.655	UG/L
28	4,4'-dichlorodiphenyl ether	A	1246	20:46	BB	0.010	UG/L

4 . 0 *

Comments for Sample AD25873 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-11-2001 Time: 11:59:37

Recoveries for 2 of the 18 compounds in the MDL Standard were above thier acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The recoveries for 2 of the 3 Surrogate Standards in this sample were below their acceptance ranges.