


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/30/2001 Time: 11:10:18

Sample: AD25874
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/1/01 11:07 Ended: 11/16/01 11:07 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		2.1		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.1		1.0
20	Surr_Triphenyl phosphate		5.4		1.0
21	Surr_Perylene-d12		2.9		1.0

Chromatogram Plot File: C:\25874P Date: Nov-17-1991 12:21:45
Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25874
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 6191707



Integration Report

Quanfile: 25874P

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25874

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:17	737	Auto	1,542,180	764,728
2	phenanthrene d-10 (IS)	16:55	1015	Auto	2,579,664	999,649
3	chrysene d-12 (IS)	22:41	1361	Auto	1,808,047	656,135
4	p-terphenyl d-14 (IS)	20:50	1250	Auto	3,114,031	1,940,920
5	acenaphthene d-10 (SURR)	12:17	737	Auto	1,542,180	764,728
6	phenanthrene d-10 (SURR)	16:55	1015	Auto	2,579,664	999,649
7	chrysene d-12 (SURR)	22:41	1361	Auto	1,808,047	656,135
8	1,3-dimethyl-2-nitrobenzene	7:58	478	Auto	984,855	527,370
9	triphenyl phosphate (S)	22:06	1326	Auto	1,389,490	505,310
10	perylene d-12 (SURR)	26:58	1618	Auto	637,116	111,737
11	hexachlorobenzene	15:22	922	Auto	6,797	2,391
12	bis(2-ethylhexyl)adipate	21:58	1318	Auto	27,603	2,886
13	bis(2-ethylhexyl)phthalate	22:58	1378	Auto	1,372,776	413,517
14	2,6-dinitrotoluene	11:57	717	Auto	48,941	23,954
15	4,4'-dichlorodiphenyl ether	20:46	1246	Auto	5,584	3,594

Reviewed 12/5/01

Quantitation Report

Quanfile: 25874P

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25874

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	737	12:17	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1015	16:55	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1361	22:41	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1250	20:50	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	737	12:17	BB	3.259	UG/L
6	phenanthrene d-10(SURR)	A	1015	16:55	BV	3.555	UG/L
7	chrysene d-12 (SURR)	A	1361	22:41	BB	3.735	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	478	7:58	BB	5.127	UG/L
9	triphenyl phosphate (S)	A	1326	22:06	BV	5.361	UG/L
10	perylene d-12 (SURR)	A	1618	26:58	BB	2.920	UG/L
12	hexachlorobenzene	A	922	15:22	BB	0.033	UG/L
16	bis(2-ethylhexyl)adipate	A	1318	21:58	MM	0.127	UG/L
17	bis(2-ethylhexyl)phthalate	A	1378	22:58	BV	2.070	UG/L
20	2,6-dinitrotoluene	A	717	11:57	BB	0.490	UG/L
28	4,4'-dichlorodiphenyl ether	A	1246	20:46	BB	0.026	UG/L

Comments for Sample AD25874 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-11-2001 Time: 12:00:38

Recoveries for 2 of the 18 compounds in the MDL Standard were above thier acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The recovery for 1 of the 3 Surrogate Standards in this sample was below its acceptance range.