


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 11/30/2001 Time: 11:43:45

Sample: AD25940
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/1/01 11:24 Ended: 11/16/01 11:24 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		2.3		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.2		1.0
20	Surr_Triphenyl phosphate		5.7		1.0
21	Surr_Perylene-d12		3.1		1.0

Chromatogram Plot File: C:\25940S Date: Nov-17-1991 14:05:02
Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25940
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 6964385



Integration Report

Quanfile: 25940S

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25940

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:17	737	Auto	2,004,988	1,082,258
2	phenanthrene d-10 (IS)	16:55	1015	Auto	3,533,110	1,407,241
3	chrysene d-12 (IS)	22:41	1361	Auto	1,775,514	681,446
4	p-terphenyl d-14 (IS)	20:50	1250	Auto	3,308,603	2,199,875
5	acenaphthene d-10 (SURR)	12:17	737	Auto	2,004,988	1,082,258
6	phenanthrene d-10 (SURR)	16:55	1015	Auto	3,533,110	1,407,241
7	chrysene d-12 (SURR)	22:41	1361	Auto	1,775,514	681,446
8	1,3-dimethyl-2-nitrobenzene	7:58	478	Auto	1,290,462	712,387
9	triphenyl phosphate (S)	22:05	1325	Auto	1,452,370	570,450
10	perylene d-12 (SURR)	26:58	1618	Auto	671,847	118,479
11	simazine	16:06	966	Auto	6,820	1,799
12	bis(2-ethylhexyl)adipate	22:05	1325	Auto	30,689	4,091
13	bis(2-ethylhexyl)phthalate	22:58	1378	Auto	1,469,324	506,085
14	2,6-dinitrotoluene	11:57	717	Auto	95,799	46,145
15	2,4-dinitrotoluene	13:04	784	Auto	2,347	1,369

Reviewed 15/5/01

Quantitation Report

Quanfile: 25940S

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/16/01; SAMPLE 25940

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	737	12:17	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1015	16:55	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1361	22:41	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1250	20:50	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	737	12:17	BB	3.987	UG/L
6	phenanthrene d-10(SURR	A	1015	16:55	BV	4.582	UG/L
7	chrysene d-12 (SURR)	A	1361	22:41	BB	3.453	UG/L
8	1,3-dimethyl-2-nitroben	A	478	7:58	BV	5.167	UG/L
9	triphenyl phosphate (S	A	1325	22:05	BB	5.706	UG/L
10	perylene d-12 (SURR)	A	1618	26:58	BB	3.135	UG/L
13	simazine	A	966	16:06	BB	0.068	UG/L
16	bis(2-ethylhexyl)adipa	A	1325	22:05	MM	0.144	UG/L
17	bis(2-ethylhexyl)phtha	A	1378	22:58	VB	2.273	UG/L
20	2,6-dinitrotoluene	A	717	11:57	BB	0.743	UG/L
21	2,4-dinitrotoluene	A	784	13:04	BB	0.017	UG/L

Comments for Sample AD25940 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-11-2001 Time: 12:05:15

Recoveries for 2 of the 18 compounds in the MDL Standard were above thier acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The recovery for 1 of the 3 Surrogate Standards in this sample was below its acceptance range.