


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 12/21/2001 Time: 16:16:18

Sample: AD26679
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/8/01 16:15 Ended: 11/28/01 16:15 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	MDL
1	Hexachlorocyclopentadiene		< MDL	0.10
2	Hexachlorobenzene		< MDL	0.10
3	Simazine		< MDL	0.40
4	Atrazine		< MDL	0.20
5	Alachlor		< MDL	0.20
6	bis(2-Ethylhexyl)adipate		< MDL	0.60
7	bis(2-Ethylhexyl)phthalate		< MDL	0.60
8	Benzo(a)pyrene		< MDL	0.20
9	EPTC		< MDL	1.0
10	2,6-Dinitrotoluene		< MDL	2.0
11	2,4-Dinitrotoluene		< MDL	2.0
12	Molinate		< MDL	0.90
13	Terbacil		< MDL	2.0
14	Acetochlor		< MDL	2.0
15	Metribuzin		< MDL	0.15
16	Metolachlor		< MDL	0.75
17	Butachlor		< MDL	0.40
18	4,4-DDE		< MDL	0.80
19	Surr_ 1,3-Dimethyl-2-nitrobenzen		4.7	1.0
20	Surr_ Triphenyl phosphate		5.8	1.0
21	Surr_ Perylene-d12		4.8	1.0

Chromatogram Plot
 File: E:\26679J Date: Nov-28-1991 21:30:04
 Comment: BOHLK; METHOD 525; INST 204; 11/28/01; SAMPLE 26679
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 8101520



Integration Report Quanfile: 26679J Quan Entries: 16
 Comment: BOHLK; METHOD 525; INST 204; 11/28/01; SAMPLE 26679
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:12	732	Auto	1,711,059	1,100,359
2	phenanthrene d-10 (IS)	16:47	1007	Auto	2,642,612	946,637
3	chrysene d-12 (IS)	22:35	1355	Auto	1,978,139	964,800
4	p-terphenyl d-14 (IS)	20:46	1246	Auto	3,521,064	2,648,310
5	acenaphthene d-10 (SURR)	12:12	732	Auto	1,711,059	1,100,359
6	phenanthrene d-10 (SURR)	16:47	1007	Auto	2,642,612	946,637
7	chrysene d-12 (SURR)	22:35	1355	Auto	1,978,139	964,800
8	1,3-dimethyl-2-nitrobenzene	7:54	474	Auto	957,746	418,954
9	triphenyl phosphate (S)	22:01	1321	Auto	1,261,426	597,000
10	perylene d-12 (SURR)	26:48	1608	Auto	1,095,827	227,479
11	simazine	15:52	952	Auto	0	0
12	bis(2-ethylhexyl)adipate	21:54	1314	Auto	29,211	5,164
13	bis(2-ethylhexyl)phthalate	22:52	1372	Auto	332,290	108,556
14	2,6-dinitrotoluene	11:52	712	Auto	85,885	53,297
15	2,4-dinitrotoluene	13:00	780	Auto	9,272	3,439
16	butachlor	20:19	1219	Auto	2,816	2,816

reversed IS/surrs

Quantitation Report

Quanfile: 26679J

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 11/28/01; SAMPLE 26679

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	732	12:12	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1007	16:47	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1355	22:35	BV	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1246	20:46	BV	5.000	UG/L
5	acenaphthene d-10(SURR)	A	732	12:12	BB	3.205	UG/L
6	phenanthrene d-10(SURR)	A	1007	16:47	BV	3.445	UG/L
7	chrysene d-12 (SURR)	A	1355	22:35	BV	3.356	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	474	7:54	BV	4.665	UG/L
9	triphenyl phosphate (S)	A	1321	22:01	BB	5.752	UG/L
10	perylene d-12 (SURR)	A	1608	26:48	BB	4.799	UG/L
13	simazine	A	952	15:52	BB	0.001	UG/L
16	bis(2-ethylhexyl)adipate	A	1314	21:54	MM	0.066	UG/L
17	bis(2-ethylhexyl)phthalate	A	1372	22:52	BB	0.454	UG/L
20	2,6-dinitrotoluene	A	712	11:52	BB	0.857	UG/L
21	2,4-dinitrotoluene	A	780	13:00	MM	0.082	UG/L
27	butachlor	A	1219	20:19	BB	0.015	UG/L

• **Comments for Sample AD26679 - Analysis \$525**

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-22-2001 Time: 12:58:25

The breakdown for Endrin in the Breakdown Standard was 40%. Recoveries for 4 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the LCS Standard were below their acceptance ranges.