


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 12/21/2001 Time: 16:24:32

Sample: AD26712
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/8/01 16:23 Ended: 11/28/01 16:23 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	MDL
1	Hexachlorocyclopentadiene		< MDL	0.10
2	Hexachlorobenzene		< MDL	0.10
3	Simazine		< MDL	0.40
4	Atrazine		< MDL	0.20
5	Alachlor		< MDL	0.20
6	bis(2-Ethylhexyl)adipate		< MDL	0.60
7	bis(2-Ethylhexyl)phthalate		< MDL	0.60
8	Benzo(a)pyrene		< MDL	0.20
9	EPTC		< MDL	1.0
10	2,6-Dinitrotoluene		< MDL	2.0
11	2,4-Dinitrotoluene		< MDL	2.0
12	Molinate		< MDL	0.90
13	Terbacil		< MDL	2.0
14	Acetochlor		< MDL	2.0
15	Metribuzin		< MDL	0.15
16	Metolachlor		< MDL	0.75
17	Butachlor		< MDL	0.40
18	4,4-DDE		< MDL	0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.7	1.0
20	Surr_Triphenyl phosphate		6.1	1.0
21	Surr_Perylene-d12		4.8	1.0

Chromatogram Plot File: E:\26712P Date: Nov-29-1991 04:23:03
 Comment: BOHLK; METHOD 525; INST 204; 11/28/01; SAMPLE 26712
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 5407860



Integration Report Quanfile: 26712P Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/28/01; SAMPLE 26712
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:12	732	Auto	983,478	590,581
2	phenanthrene d-10 (IS)	16:48	1008	Auto	1,718,597	675,159
3	chrysene d-12(IS)	22:35	1355	Auto	1,215,603	539,657
4	p-terphenyl d-14(IS)	20:46	1246	Auto	2,529,832	1,651,041
5	acenaphthene d-10(SURR)	12:12	732	Auto	983,478	590,581
6	phenanthrene d-10(SURR)	16:48	1008	Auto	1,718,597	675,159
7	chrysene d-12 (SURR)	22:35	1355	Auto	1,215,603	539,657
8	1,3-dimethyl-2-nitrobenzene	7:54	474	Auto	667,530	306,678
9	triphenyl phosphate (S)	22:01	1321	Auto	823,739	388,318
10	perylene d-12 (SURR)	26:47	1607	Auto	679,329	136,776
11	simazine	15:58	958	Auto	907	907
12	bis(2-ethylhexyl)adipate	22:01	1321	Auto	27,845	4,757
13	bis(2-ethylhexyl)phthalate	22:52	1372	Auto	240,764	72,237
14	2,6-dinitrotoluene	11:52	712	Auto	7,648	4,052
15	butachlor	20:19	1219	Auto	1,285	1,285

Quantitation Report

Quanfile: 26712P

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/28/01; SAMPLE 26712

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	732	12:12	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1008	16:48	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1355	22:35	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1246	20:46	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	732	12:12	BB	2.564	UG/L
6	phenanthrene d-10(SURR	A	1008	16:48	BV	3.119	UG/L
7	chrysene d-12 (SURR)	A	1355	22:35	BB	2.870	UG/L
8	1,3-dimethyl-2-nitroben	A	474	7:54	BV	5.657	UG/L
9	triphenyl phosphate (S	A	1321	22:01	BV	6.113	UG/L
10	perylene d-12 (SURR)	A	1607	26:47	BB	4.841	UG/L
13	simazine	A	958	15:58	BB	0.017	UG/L
16	bis(2-ethylhexyl)adipa	A	1321	22:01	MM	0.102	UG/L
17	bis(2-ethylhexyl)phtha	A	1372	22:52	BB	0.538	UG/L
20	2,6-dinitrotoluene	A	712	11:52	BB	0.134	UG/L
27	butachlor	A	1219	20:19	BB	0.011	UG/L

• **Comments for Sample AD26712 - Analysis \$525**

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-22-2001 Time: 13:04:43

The breakdown for Endrin in the Breakdown Standard was 40%. Recoveries for 4 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the LCS Standard were below their acceptance ranges.