

User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 01/02/2002 Time: 09:43:54

Sample: AD26829
Analysis: \$LP525 Organic Extractables
Units: ug
Started: 11/29/01 09:37 Ended: 11/29/01 09:37 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result
1	DFTPP		PASSED
2	ENDRIN	USPEC	62
3	DDT		5.3

Chromatogram Plot

File: E:\DFTA1129

Date: Nov-29-1991 14:42:04

Comment: BOHLK; METHOD 525; INST 204; 11/29/01; DFTPP/ENDRIN

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0


Plotted: 1 to 1860

Range: 1 to 1860

100% = 4242968



Chromatogram Plot File: E:\DFTA1129 Date: Nov-29-1991 14:42:04
Comment: BOHLK; METHOD 525; INST 204; 11/29/01; DFTPP/ENDRIN
Scan No: 1080 Retention Time: 18:00 RIC: 27874 Mass Range: 47 - 207
Plotted: 1080 to 1120 Range: 1 to 1860 100% = 4242968



Mass List

File: E:\DFTA1129 Date: Nov-29-1991 14:42:04

Comment: BOHLK; METHOD 525; INST 204; 11/29/01; DFTPP/ENDRIN

Scan No: 1096 Retention Time: 18:16 RIC: 4242968 Mass Range: 47 - 444

Peaks: 183 Mass of Base Peak: 442 Int: 630381 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
47	4,166	0.66	83	3,125	0.50	116	1,909	0.30
49	12,847	2.04	84	4,687	0.74	117	86,284	13.69
50	84,722	13.44	85	4,513	0.72	118	4,687	0.74
51	289,756	45.97	86	8,159	1.29	122	4,340	0.69
52	8,333	1.32	87	2,777	0.44	123	5,034	0.80
53	2,083	0.33	88	1,388	0.22	124	4,687	0.74
55	5,381	0.85	91	3,298	0.52	125	2,256	0.36
56	8,680	1.38	92	3,993	0.63	127	149,479	23.71
57	9,548	1.51	93	27,604	4.38	128	16,319	2.59
61	2,951	0.47	94	2,604	0.41	129	60,416	9.58
62	2,604	0.41	95	2,430	0.39	130	3,819	0.61
63	6,770	1.07	96	1,215	0.19	132	1,736	0.28
65	3,645	0.58	98	24,479	3.88	134	1,736	0.28
68	2,256	0.36	99	12,152	1.93	135	4,166	0.66
69	181,597	28.81	101	6,770	1.07	136	2,430	0.39
73	2,083	0.33	103	4,513	0.72	137	2,604	0.41
74	25,000	3.97	104	7,465	1.18	138	1,388	0.22
75	25,347	4.02	105	2,083	0.33	141	9,375	1.49
76	16,840	2.67	106	1,215	0.19	142	3,993	0.63
77	198,611	31.51	107	71,527	11.35	146	2,256	0.36
78	9,375	1.49	108	11,458	1.82	147	12,326	1.96
79	26,041	4.13	109	3,472	0.55	148	13,194	2.09
80	18,055	2.86	110	114,236	18.12	149	5,208	0.83
81	10,243	1.62	111	17,534	2.78	150	1,736	0.28
82	2,256	0.36	112	3,993	0.63	152	1,562	0.25

Mass List

File: E:\DFTA1129 Date: Nov-29-1991 14:42:04

Comment: BOHLK; METHOD 525; INST 204; 11/29/01; DFTPP/ENDRIN

Scan No: 1096 Retention Time: 18:16 RIC: 4242968 Mass Range: 47 - 444

Peaks: 183 Mass of Base Peak: 442 Int: 630381 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
152	1,562	0.25	196	6,597	1.05	244	34,895	5.54
153	3,993	0.63	197	3,819	0.61	245	5,034	0.80
154	3,298	0.52	198	385,069	61.09	246	10,763	1.71
155	4,340	0.69	199	31,423	4.98	247	2,430	0.39
156	8,333	1.32	204	10,243	1.62	255	261,111	41.42
160	2,951	0.47	205	20,659	3.28	256	40,625	6.44
165	5,381	0.85	206	92,881	14.73	257	1,909	0.30
166	3,125	0.50	207	11,458	1.82	258	10,416	1.65
167	29,687	4.71	208	3,472	0.55	265	3,125	0.50
168	6,423	1.02	211	4,513	0.72	272	1,215	0.19
169	3,298	0.52	216	1,562	0.25	273	5,034	0.80
172	1,215	0.19	217	11,631	1.85	274	27,951	4.43
175	5,208	0.83	219	2,256	0.36	275	143,402	22.75
177	3,993	0.63	221	2,256	0.36	276	18,750	2.97
178	1,215	0.19	223	2,083	0.33	277	14,236	2.26
179	11,805	1.87	224	36,458	5.78	278	2,430	0.39
180	9,027	1.43	225	18,923	3.00	283	3,472	0.55
181	5,729	0.91	227	21,006	3.33	284	1,388	0.22
185	5,902	0.94	228	2,083	0.33	292	1,388	0.22
186	60,243	9.56	229	3,298	0.52	293	4,513	0.72
187	13,888	2.20	231	2,951	0.47	294	2,430	0.39
188	2,430	0.39	235	1,736	0.28	296	29,861	4.74
189	2,083	0.33	236	1,388	0.22	297	2,604	0.41
192	2,256	0.36	237	1,562	0.25	301	1,041	0.17
193	5,902	0.94	243	1,562	0.25	302	2,083	0.33

Mass List

File: E:\DFTA1129 Date: Nov-29-1991 14:42:04


Comment: BOHLK; METHOD 525; INST 204; 11/29/01; DFTPP/ENDRIN

Scan No: 1096 Retention Time: 18:16 RIC: 4242968 Mass Range: 47 - 444


Peaks: 183 Mass of Base Peak: 442 Int: 630381 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
211	4,513	0.72	272	1,215	0.19	335	3,298	0.52
216	1,562	0.25	273	5,034	0.80	341	3,645	0.58
217	11,631	1.85	274	27,951	4.43	346	3,819	0.61
219	2,256	0.36	275	143,402	22.75	352	3,125	0.50
221	2,256	0.36	276	18,750	2.97	353	7,291	1.16
223	2,083	0.33	277	14,236	2.26	354	6,944	1.10
224	36,458	5.78	278	2,430	0.39	355	1,562	0.25
225	18,923	3.00	283	3,472	0.55	365	19,097	3.03 ✓
227	21,006	3.33	284	1,388	0.22	366	2,777	0.44
228	2,083	0.33	292	1,388	0.22	371	1,736	0.28
229	3,298	0.52	293	4,513	0.72	372	8,680	1.38
231	2,951	0.47	294	2,430	0.39	373	2,951	0.47
235	1,736	0.28	296	29,861	4.74	390	1,736	0.28
236	1,388	0.22	297	2,604	0.41	391	1,215	0.19
237	1,562	0.25	301	1,041	0.17	402	2,256	0.36
243	1,562	0.25	302	2,083	0.33	403	4,340	0.69
244	34,895	5.54	303	5,034	0.80	421	4,861	0.77
245	5,034	0.80	304	1,909	0.30	422	3,993	0.63
246	10,763	1.71	315	4,687	0.74	423	34,895	5.54
247	2,430	0.39	316	3,993	0.63	424	4,861	0.77
255	261,111	41.42	317	1,388	0.22	440	1,041	0.17
256	40,625	6.44	323	16,493	2.62	441	97,048	15.40 ✓
257	1,909	0.30	324	4,340	0.69	442	630,381	100.00 ✓
258	10,416	1.65	327	2,951	0.47	443	105,381	16.72 ✓
265	3,125	0.50	334	10,763	1.71	444	10,937	1.73


Chromatogram Plot File: E:\DFTA1129 Date: Nov-29-1991 14:42:04
 Comment: BOHLK; METHOD 525; INST 204; 11/29/01; DFTPP/ENDRIN
 Scan No: 1280 Retention Time: 21:20 RIC: 27709 Mass Range: 47 - 281
 Plotted: 1260 to 1280 Range: 1 to 1860 100% = 4242968



Chromatogram Plot File: E:\DFTA1129 Date: Nov-29-1991 14:42:04
Comment: BOHLK; METHOD 525; INST 204; 11/29/01; DFTPP/ENDRIN
Scan No: 1295 Retention Time: 21:35 RIC: 27925 Mass Range: 47 - 281
Plotted: 1268 to 1295 Range: 1 to 1860 100% = 4242968



Chromatogram Plot File: E:\DFTA1129 Date: Nov-29-1991 14:42:04
Comment: BOHLK; METHOD 525; INST 204; 11/29/01; DFTPP/ENDRIN
Scan No: 1367 Retention Time: 22:47 RIC: 30205 Mass Range: 47 - 281
Plotted: 1337 to 1367 Range: 1 to 1860 100% = 4242968



Chromatogram Plot

File: E:\DDTA1129

Date: Nov-29-1991 15:18:17

Comment: BOHLK; METHOD 525; INST 204; 11/29/01; DDT

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0

Plotted: 1 to 1860

Range: 1 to 1860

100% = 4385045



Chromatogram Plot

File: E:\DDTA1129

Date: Nov-29-1991 15:18:17

Comment: BOHLK; METHOD 525; INST 204; 11/29/01; DDT

Scan No: 1259

Retention Time: 20:59


RIC: 22600

Mass Range: 47 - 207


Plotted: 1239 to 1259

Range: 1 to 1860

100% = 4385045



Chromatogram Plot File: E:\DDTA1129 Date: Nov-29-1991 15:18:17
Comment: BOHLK; METHOD 525; INST 204; 11/29/01; DDT
Scan No: 1292 Retention Time: 21:32 RIC: 46752 Mass Range: 45 - 429
Plotted: 1272 to 1292 Range: 1 to 1860 100% = 4385045



Chromatogram Plot File: E:\DDTA1129 Date: Nov-29-1991 15:18:17
Comment: BOHLK; METHOD 525; INST 204; 11/29/01; DDT
Scan No: 1324 Retention Time: 22:04 RIC: 27955 Mass Range: 47 - 281
Plotted: 1300 to 1324 Range: 1 to 1860 100% = 4385045

