


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 01/02/2002 Time: 09:51:14

Sample: AD26829
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/9/01 09:49 Ended: 11/29/01 09:49 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		4.9		1.0
20	Surr_Triphenyl phosphate		5.1		1.0
21	Surr_Perylene-d12		5.2		1.0

Chromatogram Plot File: E:\26829C Date: Nov-29-1991 22:57:30
Comment: BOHLK; METHOD 525; INST 204; 11/29/01; SAMPLE 26829
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 7395095



Integration Report Quanfile: 26829C Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/29/01; SAMPLE 26829
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:13	733	Auto	1,607,073	976,607
2	phenanthrene d-10 (IS)	16:49	1009	Auto	2,596,097	932,797
3	chrysene d-12 (IS)	22:36	1356	Auto	2,040,984	915,945
4	p-terphenyl d-14 (IS)	20:47	1247	Auto	3,130,008	2,249,277
5	acenaphthene d-10 (SURR)	12:13	733	Auto	1,607,073	976,607
6	phenanthrene d-10 (SURR)	16:49	1009	Auto	2,596,097	932,797
7	chrysene d-12 (SURR)	22:36	1356	Auto	2,040,984	915,945
8	1,3-dimethyl-2-nitrobenzene	7:54	474	Auto	940,700	397,364
9	triphenyl phosphate (S)	22:02	1322	Auto	1,162,758	515,361
10	perylene d-12 (SURR)	26:49	1609	Auto	1,217,950	249,307
11	bis(2-ethylhexyl)adipate	21:54	1314	Auto	51,970	8,798
12	bis(2-ethylhexyl)phthalate	22:53	1373	Auto	427,415	144,726
13	2,6-dinitrotoluene	11:53	713	Auto	93,780	56,902
14	2,4-dinitrotoluene	13:02	782	Auto	14,527	4,752
15	butachlor	20:20	1220	Auto	1,700	1,700

reviewed IS/surrs

Quantitation Report

Quanfile: 26829C

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/29/01; SAMPLE 26829

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	733	12:13	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1009	16:49	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1356	22:36	BV	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1247	20:47	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	733	12:13	BB	3.387	UG/L
6	phenanthrene d-10(SURR	A	1009	16:49	BV	3.808	UG/L
7	chrysene d-12 (SURR)	A	1356	22:36	BV	3.895	UG/L
8	1,3-dimethyl-2-nitrobe	A	474	7:54	BV	4.879	UG/L
9	triphenyl phosphate (S	A	1322	22:02	BV	5.139	UG/L
10	perylene d-12 (SURR)	A	1609	26:49	BB	5.169	UG/L
16	bis(2-ethylhexyl)adipa	A	1314	21:54	MM	0.113	UG/L
17	bis(2-ethylhexyl)phtha	A	1373	22:53	BB	0.569	UG/L
20	2,6-dinitrotoluene	A	713	11:53	BB	0.996	UG/L
21	2,4-dinitrotoluene	A	782	13:02	MM	0.136	UG/L
27	butachlor	A	1220	20:20	BB	0.010	UG/L

Comments for Sample AD26829 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 01-03-2002 Time: 12:29:59

The recoveries for 3 of the 18 compounds in the LCS Standard were below their acceptance range. The breakdown for Endrin in the Breakdown Standard was 62%.