


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 01/02/2002 Time: 09:52:51

Sample: AD26830
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/9/01 09:51 Ended: 11/29/01 09:51 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		4.5		1.0
20	Surr_Triphenyl phosphate		6.0		1.0
21	Surr_Perylene-d12		5.3		1.0

Chromatogram Plot File: E:\26830D Date: Nov-29-1991 23:31:56
Comment: BOHLK; METHOD 525; INST 204; 11/29/01; SAMPLE 26830
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0.
Plotted: 1 to 1860 Range: 1 to 1860 100% = 7347579



Integration Report Quanfile: 26830D Quan Entries: 18
 Comment: BOHLK; METHOD 525; INST 204; 11/29/01; SAMPLE 26830
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:13	733	Auto	1,442,501	778,055
2	phenanthrene d-10 (IS)	16:48	1008	Auto	2,411,345	869,851
3	chrysene d-12(IS)	22:36	1356	Auto	1,911,986	885,997
4	p-terphenyl d-14(IS)	20:47	1247	Auto	3,372,458	2,158,636
5	acenaphthene d-10(SURR)	12:13	733	Auto	1,442,501	778,055
6	phenanthrene d-10(SURR)	16:48	1008	Auto	2,411,345	869,851
7	chrysene d-12 (SURR)	22:36	1356	Auto	1,911,986	885,997
8	1,3-dimethyl-2-nitrobenzene	7:54	474	Auto	773,689	413,082
9	triphenyl phosphate (S)	22:01	1321	Auto	1,272,852	530,537
10	perylene d-12 (SURR)	26:49	1609	Auto	1,161,823	253,164
11	hexachlorobenzene	15:14	914	Auto	876	876
12	bis(2-ethylhexyl)adipate	22:01	1321	Auto	28,462	5,795
13	bis(2-ethylhexyl)phthalate	22:53	1373	Auto	417,473	173,738
14	benzo(a)pyrene	26:49	1609	Auto	6,221	2,226
15	2,6-dinitrotoluene	11:53	713	Auto	21,670	13,624
16	2,4-dinitrotoluene	13:01	781	Auto	2,370	1,415
17	butachlor	20:20	1220	Auto	1,893	1,893
18	4,4'-dichlorodiphenyl ether	20:42	1242	Auto	1,985	1,110

reviewed IS/sur's

Quantitation Report

Quanfile: 268300

Quan Entries: 18

Comment: BOHLK; METHOD 525; INST 204; 11/29/01; SAMPLE 26830

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	733	12:13	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1008	16:48	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1356	22:36	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1247	20:47	BV	5.000	UG/L
5	acenaphthene d-10(SURR)	A	733	12:13	BB	2.821	UG/L
6	phenanthrene d-10(SURR)	A	1008	16:48	BV	3.283	UG/L
7	chrysene d-12 (SURR)	A	1356	22:36	BB	3.387	UG/L
8	1,3-dimethyl-2-nitrobe	A	474	7:54	BV	4.470	UG/L
9	triphenyl phosphate (S	A	1321	22:01	BV	6.005	UG/L
10	perylene d-12 (SURR)	A	1609	26:49	BB	5.264	UG/L
12	hexachlorobenzene	A	914	15:14	BB	0.006	UG/L
16	bis(2-ethylhexyl)adipa	A	1321	22:01	MM	0.066	UG/L
17	bis(2-ethylhexyl)phtha	A	1373	22:53	BB	0.594	UG/L
18	benzo(a)pyrene	A	1609	26:49	BB	0.016	UG/L
20	2,6-dinitrotoluene	A	713	11:53	BB	0.258	UG/L
21	2,4-dinitrotoluene	A	781	13:01	BB	0.025	UG/L
27	butachlor	A	1220	20:20	BB	0.011	UG/L
28	4,4'-dde	A	1242	20:42	BB	0.010	UG/L

• **Comments for Sample AD26830 - Analysis \$525**

Westchester Dept. of Labs & Research
User: Vinci, David L
Date: 01-03-2002 Time: 12:30:40

The recoveries for 3 of the 18 compounds in the LCS Standard were below their acceptance range. The breakdown for Endrin in the Breakdown Standard was 62%.