


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 01/03/2002 Time: 10:29:54

Sample: AD26833
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/13/01 10:27 Ended: 11/30/01 10:27 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.61		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		4.3		1.0
20	Surr_Triphenyl phosphate		6.0		1.0
21	Surr_Perylene-d12		6.4		1.0

Chromatogram Plot File: E:\26833B Date: Nov-30-1991 13:48:19
 Comment: BOHLK; METHOD 525; INST 204; 11/29/01; SAMPLE 26833
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 7319799



Integration Report Quanfile: 26833B Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 11/29/01; SAMPLE 26833
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:13	733	Auto	783,374	396,516
2	phenanthrene d-10 (IS)	16:49	1009	Auto	1,148,254	363,517
3	chrysene d-12(IS)	22:37	1357	Auto	896,764	301,772
4	p-terphenyl d-14(IS)	20:47	1247	Auto	3,553,539	2,397,289
5	acenaphthene d-10(SURR)	12:13	733	Auto	783,374	396,516
6	phenanthrene d-10(SURR)	16:49	1009	Auto	1,148,254	363,517
7	chrysene d-12 (SURR)	22:37	1357	Auto	896,764	301,772
8	1,3-dimethyl-2-nitrobenzene	7:55	475	Auto	402,444	214,097
9	triphenyl phosphate (S)	22:02	1322	Auto	594,535	289,920
10	perylene d-12 (SURR)	26:51	1611	Auto	667,628	120,253
11	bis(2-ethylhexyl)adipate	22:02	1322	Auto	15,326	3,853
12	bis(2-ethylhexyl)phthalate	22:54	1374	Auto	200,393	73,172
13	2,6-dinitrotoluene	11:53	713	Auto	30,430	15,911
14	butachlor	20:23	1223	Auto	0	0

reversed IS/surr's

Quantitation Report Quanfile: 26833B Quan Entries: 16
 Comment: BOHLK; METHOD 525; INST 204; 11/29/01; SAMPLE 26833
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	733	12:13	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1009	16:49	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1357	22:37	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1247	20:47	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	733	12:13	BB	1.454	UG/L
6	phenanthrene d-10(SURR	A	1009	16:49	BV	1.484	UG/L
7	chrysene d-12 (SURR)	A	1357	22:37	BB	1.508	UG/L
8	1,3-dimethyl-2-nitroben	A	475	7:55	BB	4.282	UG/L
9	triphenyl phosphate (S	A	1322	22:02	BB	5.980	UG/L
10	perylene d-12 (SURR)	A	1611	26:51	BB	6.449	UG/L
13	simazine	A	974	16:14	BB	0.001	UG/L
16	bis(2-ethylhexyl)adipa	A	1322	22:02	MM	0.087	UG/L
17	bis(2-ethylhexyl)phtha	A	1374	22:54	BV	0.608	UG/L
18	benzo(a)pyrene	A	1598	26:38	BB	0.063	UG/L
20	2,6-dinitrotoluene	A	713	11:53	BB	0.664	UG/L
27	butachlor	A	1223	20:23	BB	0.001	UG/L

• **Comments for Sample AD26833 - Analysis \$525**

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 01-04-2002 Time: 10:32:27

The recoveries for 5 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. The recoveries for 3 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The breakdown for Endrin in the Breakdown Standard was 48%.