


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 01/03/2002 Time: 12:19:46

Sample: AD26991
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/13/01 12:18 Ended: 11/30/01 12:18 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.99		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		4.1		1.0
20	Surr_Triphenyl phosphate		5.2		1.0
21	Surr_Perylene-d12		5.4		1.0

Chromatogram Plot File: E:\26991M Date: Dec-01-1991 01:38:20
Comment: BOHLK; METHOD 525; INST 204; 11/30/01; SAMPLE 26991
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 11716448



Integration Report Quanfile: 26991M Quan Entries: 17
 Comment: BOHLK; METHOD 525; INST 204; 11/30/01; SAMPLE 26991
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:13	733	Auto	2,324,291	1,185,583
2	phenanthrene d-10 (IS)	16:49	1009	Auto	3,496,248	1,264,447
3	chrysene d-12 (IS)	22:37	1357	Auto	2,895,026	1,360,720
4	p-terphenyl d-14 (IS)	20:47	1247	Auto	6,153,160	3,968,614
5	acenaphthene d-10 (SURR)	12:13	733	Auto	2,324,291	1,185,583
6	phenanthrene d-10 (SURR)	16:49	1009	Auto	3,496,248	1,264,447
7	chrysene d-12 (SURR)	22:37	1357	Auto	2,895,026	1,360,720
8	1,3-dimethyl-2-nitrobenzene	7:55	475	Auto	1,147,873	640,033
9	triphenyl phosphate (S)	22:02	1322	Auto	1,657,489	1,000,782
10	perylene d-12 (SURR)	26:50	1610	Auto	1,795,693	387,180
11	simazine	16:00	960	Auto	4,524	1,358
12	bis(2-ethylhexyl)adipate	22:02	1322	Auto	38,769	17,165
13	bis(2-ethylhexyl)phthalate	22:54	1374	Auto	1,034,027	460,922
14	2,6-dinitrotoluene	11:53	713	Auto	59,466	26,475
15	2,4-dinitrotoluene	13:02	782	Auto	5,326	2,535
16	terbacil	17:44	1064	Auto	14,553	1,211
17	butachlor	20:21	1221	Auto	12,766	2,118

reversed IS/SURR's

Quantitation Report Quanfile: 26991M Quan Entries: 17
 Comment: BOHLK; METHOD 525; INST 204; 11/30/01; SAMPLE 26991
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	733	12:13	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1009	16:49	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1357	22:37	VV	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1247	20:47	BV	5.000	UG/L
5	acenaphthene d-10(SURR)	A	733	12:13	BV	2.492	UG/L
6	phenanthrene d-10(SURR)	A	1009	16:49	BV	2.609	UG/L
7	chrysene d-12 (SURR)	A	1357	22:37	VV	2.811	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	475	7:55	BV	4.116	UG/L
9	triphenyl phosphate (S)	A	1322	22:02	BV	5.164	UG/L
10	perylene d-12 (SURR)	A	1610	26:50	BB	5.373	UG/L
13	simazine	A	960	16:00	BB	0.042	UG/L
16	bis(2-ethylhexyl)adipate	A	1322	22:02	MM	0.059	UG/L
17	bis(2-ethylhexyl)phthalate	A	1374	22:54	BV	0.989	UG/L
20	2,6-dinitrotoluene	A	713	11:53	BV	0.438	UG/L
21	2,4-dinitrotoluene	A	782	13:02	BB	0.035	UG/L
23	terbacil	A	1064	17:44	MM	0.182	UG/L
27	butachlor	A	1221	20:21	MM	0.051	UG/L

▼ **Comments for Sample AD26991 - Analysis \$525**

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 01-04-2002 Time: 10:44:38

The recoveries for 5 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. The recoveries for 3 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The breakdown for Endrin in the Breakdown Standard was 48%.