


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 01/03/2002 Time: 12:21:29

Sample: AD26992
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/13/01 12:19 Ended: 11/30/01 12:19 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.98		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		3.9		1.0
20	Surr_Triphenyl phosphate		5.2		1.0
21	Surr_Perylene-d12		6.2		1.0

Chromatogram Plot File: E:\26992N Date: Dec-01-1991 02:12:48
Comment: BOHLK; METHOD 525; INST 204; 11/30/01; SAMPLE 26992
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 14946652



Integration Report Quanfile: 26992N Quan Entries: 17
 Comment: BOHLK; METHOD 525; INST 204; 11/30/01; SAMPLE 26992
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:13	733	Auto	2,713,816	1,771,542
2	phenanthrene d-10 (IS)	16:49	1009	Auto	4,261,462	1,680,684
3	chrysene d-12(IS)	22:36	1356	Auto	3,451,487	1,459,623
4	p-terphenyl d-14(IS)	20:47	1247	Auto	7,041,977	5,195,420
5	acenaphthene d-10(SURR)	12:13	733	Auto	2,713,816	1,771,542
6	phenanthrene d-10(SURR)	16:49	1009	Auto	4,261,462	1,680,684
7	chrysene d-12 (SURR)	22:36	1356	Auto	3,451,487	1,459,623
8	1,3-dimethyl-2-nitrobenzene	7:55	475	Auto	1,283,016	660,975
9	triphenyl phosphate (S)	22:02	1322	Auto	1,976,081	1,069,619
10	perylene d-12 (SURR)	26:50	1610	Auto	2,464,257	559,390
11	simazine	15:58	958	Auto	0	0
12	bis(2-ethylhexyl)adipate	22:02	1322	Auto	42,896	9,769
13	bis(2-ethylhexyl)phthalate	22:53	1373	Auto	1,218,716	505,248
14	benzo(a)pyrene	26:50	1610	Auto	46,869	4,828
15	2,6-dinitrotoluene	11:53	713	Auto	101,960	62,751
16	2,4-dinitrotoluene	13:01	781	Auto	7,182	2,151
17	butachlor	20:25	1225	Auto	9,106	1,158

reversed IS/ Surrogate

Quantitation Report

Quanfile: 26992N

Quan Entries: 17

Comment: BOHLK; METHOD 525; INST 204; 11/30/01; SAMPLE 26992

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	733	12:13	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1009	16:49	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1356	22:36	BV	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1247	20:47	BV	5.000	UG/L
5	acenaphthene d-10(SURR)	A	733	12:13	BB	2.542	UG/L
6	phenanthrene d-10(SURR)	A	1009	16:49	BV	2.778	UG/L
7	chrysene d-12 (SURR)	A	1356	22:36	BV	2.928	UG/L
8	1,3-dimethyl-2-nitrobe	A	475	7:55	BB	3.940	UG/L
9	triphenyl phosphate (S	A	1322	22:02	BV	5.164	UG/L
10	perylene d-12 (SURR)	A	1610	26:50	BB	6.184	UG/L
13	simazine	A	958	15:58	BB	0.001	UG/L
16	bis(2-ethylhexyl)adipa	A	1322	22:02	MM	0.055	UG/L
17	bis(2-ethylhexyl)phtha	A	1373	22:53	BV	0.977	UG/L
18	benzo(a)pyrene	A	1610	26:50	BB	0.022	UG/L
20	2,6-dinitrotoluene	A	713	11:53	BB	0.643	UG/L
21	2,4-dinitrotoluene	A	781	13:01	BB	0.040	UG/L
27	butachlor	A	1225	20:25	MM	0.030	UG/L

O
 ↑
 TB
 1/2/02

• **Comments for Sample AD26992 - Analysis \$525**

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 01-04-2002 Time: 10:45:30

The recoveries for 5 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. The recoveries for 3 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The breakdown for Endrin in the Breakdown Standard was 48%.