


User: Bohk, Ted
Westchester Dept. of Labs & Research
Date: 01/03/2002 Time: 12:27:26

Sample: AD27009
Analysis: S525 Organic Chemicals - 525.2
Units: ug
Started: 11/13/01 12:24 Ended: 11/30/01 12:24 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		1.2		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		4.7		1.0
20	Surr_Triphenyl phosphate		5.5		1.0
21	Surr_Perylene-d12		5.2		1.0

Chromatogram Plot File: E:\27009Q Date: Dec-01-1991 03:56:07
Comment: BOHLK; METHOD 525; INST 204; 11/30/01; SAMPLE 27009
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 8873055



Quantitation Report Quanfile: 27009Q Quan Entries: 16
 Comment: BOHLK; METHOD.525; INST 204; 11/30/01; SAMPLE 27009
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	733	12:13	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1009	16:49	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1357	22:37	BV	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1247	20:47	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	733	12:13	BB	2.515	UG/L
6	phenanthrene d-10 (SURR)	A	1009	16:49	BV	2.560	UG/L
7	chrysene d-12 (SURR)	A	1357	22:37	BV	2.725	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	475	7:55	BV	4.738	UG/L
9	triphenyl phosphate (S)	A	1322	22:02	BB	5.535	UG/L
10	perylene d-12 (SURR)	A	1610	26:50	BB	5.244	UG/L
13	simazine	A	961	16:01	BB	0.001	UG/L
17	bis(2-ethylhexyl)phthalate	A	1374	22:54	VV	1.221	UG/L
18	benzo(a)pyrene	A	1611	26:51	MM	0.015	UG/L
21	2,4-dinitrotoluene	A	783	13:03	BB	0.019	UG/L
27	butachlor	A	1220	20:20	MM	0.029	UG/L
28	4,4'-dichlorodiphenyl ether	A	1243	20:43	BB	0.012	UG/L

0
 ↑
 TB
 1/202

reversed IS/SURR

Comments for Sample AD27009 - Analysis \$525

Westchester Dept. of Labs & Research
User: Vinci, David L
Date: 01-04-2002 Time: 10:30:44

The recoveries for 5 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. The recoveries for 3 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The breakdown for Endrin in the Breakdown Standard was 48%.