

Method 508 modified Date Extracted 7/20/02

Page 95

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count	
LB		500 ml		1ml			1ml	8A	7		
REF					1ml			7B			
REF					↓			8B			
3799								9B			
3800								10B			
3801	FB							10A			✓
3802								9A			
3803								10A			
3804								9B			
3805	FB							9A			
3822								7A			
*364	Re-Ext.	↓		↓			↓	10A	↓		
Remarks											
drip rate must be between 5-15 ml/min for liq-liq extractors											
drip rate -											
Extraction type	Sep fun	Solv. lot#	V42E17	NaCl lot#	V39618	Na2SO4 lot#	V46623				
STANDARD	INT STD LOG#										
CALIBRATION											
REF SPIKE	2193										
MDL/MDRF											
Q. CHECK											
INT. STD.	030102										
SURROGATE	GCMS	21606									
LPC											
ENDRIN/DDT											

Reviewed by
CB 6/5/02

Analyst EH
revision 4/12/01

Verified by _____

Reviewed by DW
Date 7/2/02

Method 508 modified Date Extracted 01-09-02

Page 18

REF

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count	
1B		500mL		1ml			1ml	2A	7		
REF					5mL			2B			
365								6B			
470								7B			
558								8A			
364								1B			✓
360								6A			
559								1A			
561								7A			
557								8L			
Remarks											
drip rate must be between 5-15 ml/min for liq-liq extractors											
drip rate -											
Extraction type -		Solv. lot# <u>V3946</u>		NaCl lot# <u>V11599</u>		Na2SO4 lot# <u>V39H00</u>					
STANDARD	INT STD LOG#										
CALIBRATION											
REF SPIKE	<u>91759</u>										
MDL/MDRF											
Q. CHECK											
INT. STD.											
SURROGATE	<u>GCMS 010302</u>										
LPC											
ENDRIN/DDT											

Analyst EA
revision 4/12/01

Verified by _____

Reviewed by DW
Date 3/22/12

3/22/02

Donna NYCD E P

AE 00364 - No charge
too long to report

AE 01153 - \$71.25

repeated error ref fail
x2

DFTPP

Data File : C:\MSDCHEM\1\DATA\031902\DFTPP2.D
Acq On : 20 Mar 2002 1:34 am
Sample :
Misc :
MS Integration Params: BENZO.P

Vial: 1
Operator: McLean
Inst : Instrumen
Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\5080302.M (RTE Integrator)
Title : Quantitation For Method 508gcscan

Spectrum Information: Scan 2287

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	61.0	469760	PASS
68	69	0.00	2	1.5	5825	PASS
69	198	0.00	100	49.7	383168	PASS
70	69	0.00	2	0.5	1876	PASS
127	198	10	80	52.1	401472	PASS
197	198	0.00	2	0.0	0	PASS
198	198	50	100	100.0	770368	PASS
199	198	5	9	6.8	52280	PASS
275	198	10	60	22.0	169216	PASS
365	198	1	100	2.8	21224	PASS
441	443	0.01	100	73.0	77160	PASS
442	198	50	100	69.5	535488	PASS
443	442	15	24	19.7	105736	PASS

DFTPP2.D 5080302.M Thu Mar 21 11:50:49 2002

File : C:\MSDCHEM\1\DATA\031902\DFTPP2.D
 Operator : McLean
 Acquired : 20 Mar 2002 1:34 am using AcqMethod 508SCAN
 Instrument : Instrumen
 Sample Name:
 Misc Info :
 Vial Number: 1

