

Method 508 met.Date Extracted 2/21/02Page 99

REF

#GCMs - 17138

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count
LB	Re-Ext	500ml		1ml			1ml	7A	7	
REF					1ml			8A		
REF					↓			8B		
31542								8A		
31581								10A		
31499								10B		
31498								7A		
31581								7B		
31591								10A		
31589								10B		
31590								8A		
31588								8B		
31593								8B		
31638	FAV	↓		↓			↓	7B	↓	
Remarks										
Reviewed by, L/B 6/5/02										
drip rate must be between 5-15 ml/min for liq-liq extractors										
drip rate -										
Extraction type <u>sed Fun</u>		Solv. lot# <u>V39466</u>		NaCl lot# <u>V39618</u>		Na2SO4 lot# <u>V46623</u>				
STANDARD	INT STD LOG#									
CALIBRATION	<u>HPD 1306</u>									
REF SPIKE	<u>2193</u>									
MDL/MORF										
Q. CHECK										
INT. STD.	<u>022002</u>									
SURROGATE	<u>GCMs 21606</u>									
LPC										
ENDRIN/DDT										

Analyst EA/B.D.

ma

Dariusz

D
2/28/2