

Data File : C:\MSDCHEM\1\DATA\022802\DFTPPB.D

Acq On : 1 Mar 2002 12:04 am

Sample :

Misc :

MS Integration Params: BENZO.P

Vial: 1

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\5080202.M (RTE Integrator)

Title : Quantitation For Method 508gcscan

Spectrum Information: Scan 2290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	59.1	340544	PASS
68	69	0.00	2	1.5	4257	PASS
69	198	0.00	100	47.7	274944	PASS
70	69	0.00	2	0.5	1379	PASS
127	198	10	80	50.9	293696	PASS
197	198	0.00	2	0.0	0	PASS
198	198	50	100	100.0	576512	PASS
199	198	5	9	7.0	40360	PASS
275	198	10	60	20.4	117384	PASS
365	198	1	100	2.4	14123	PASS
441	443	0.01	100	76.8	63376	PASS
442	198	50	100	73.2	421824	PASS
443	442	15	24	19.6	82488	PASS

DFTPPB.D 5080202.M

Wed Mar 13 06:58:24 2002

File : C:\MSDCHEM\1\DATA\022802\DFTPPB.D
 Operator : McLean
 Acquired : 1 Mar 2002 12:04 am using AcqMethod 508SCAN
 Instrument : Instrumen
 Sample Name:
 Misc Info :
 Vial Number: 1

