

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 310

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1				Seq: 2300	17:22:18	06 Feb 02	HG	
Hg	.000	ppb	4527					
*** Standard: 2 Rep: 1				Seq: 2301	17:23:57	06 Feb 02	HG	
Hg	.400	ppb	37178					
*** Standard: 3 Rep: 1				Seq: 2302	17:25:39	06 Feb 02	HG	
Hg	1.00	ppb	74584					
*** Standard: 4 Rep: 1				Seq: 2303	17:27:42	06 Feb 02	HG	
Hg	3.00	ppb	231716					
*** Standard: 5 Rep: 1				Seq: 2304	17:29:19	06 Feb 02	HG	
Hg	5.00	ppb	366400					
*** Standard: 6 Rep: 1				Seq: 2305	17:31:06	06 Feb 02	HG	
Hg	10.0	ppb	730285					
*** Check Standard: 2 Ck21.0 qc				Seq: 2306	17:33:16	06 Feb 02	HG	
Line Flag %Rcv. Found True Units SD/RSD								
Hg	95.5	✓	.955	1.00	ppb	.000		
*** Sample ID: 2761 pot				Seq: 2307	17:35:33	06 Feb 02	HG	
Hg	-.061	ppb	.000	-.061				
*** Sample ID: 2761 spk				Seq: 2308	17:37:09	06 Feb 02	HG	
Hg	4.97	ppb	.000	4.97				
*** Sample ID: 2761d				Seq: 2309	17:39:40	06 Feb 02	HG	
Hg	.014	ppb	.000	.014				
*** Sample ID: 2761 spk				Seq: 2310	17:41:40	06 Feb 02	HG	
Hg	4.87	ppb	.000	4.87				
*** Sample ID: 2782				Seq: 2311	17:43:16	06 Feb 02	HG	
Hg	-.036	ppb	.000	-.036				
*** Sample ID: 2783				Seq: 2312	17:44:52	06 Feb 02	HG	
Hg	-.084	ppb	.000	-.084				
*** Sample ID: 2784				Seq: 2313	17:46:42	06 Feb 02	HG	
Hg	-.098	ppb	.000	-.098				

Reviewed by,
JBS 6/29/02

