

westchester county labs

Folder: RUNS1

Page 321

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq:	2342	15:37:48	08 Feb 02	HG
Hg	.000	ppb	792					=
*** Standard: 2 Rep: 1								
				Seq:	2343	15:39:27	08 Feb 02	HG
Hg	.400	ppb	28847					=
*** Standard: 3 Rep: 1								
				Seq:	2344	15:41:48	08 Feb 02	HG
Hg	1.00	ppb	72735					=
*** Standard: 4 Rep: 1								
				Seq:	2345	15:44:06	08 Feb 02	HG
Hg	3.00	ppb	206795					=
*** Standard: 5 Rep: 1								
				Seq:	2346	15:45:46	08 Feb 02	HG
Hg	5.00	ppb	339659					=
*** Standard: 6 Rep: 1								
				Seq:	2347	15:48:09	08 Feb 02	HG
Hg	10.0	ppb	712550					=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		106.0	1.06	1.00	ppb	.000		=
*** Sample ID: 2966								
				Seq:	2357	16:07:37	08 Feb 02	HG
Hg	-.017	ppb	.000					=
*** Sample ID: 2967								
				Seq:	2358	16:09:14	08 Feb 02	HG
Hg	-.008	ppb	.000					=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		94.5	.945	1.00	ppb	.000		=
*** Sample ID: 2968								
				Seq:	2360	16:12:40	08 Feb 02	HG
Hg	.020	ppb	.000					=
*** Sample ID: 2969								
				Seq:	2361	16:15:06	08 Feb 02	HG
Hg	.091	ppb	.000					=
*** Sample ID: 2970								
				Seq:	2362	16:16:54	08 Feb 02	HG
Hg	.118	ppb	.000					=
*** Sample ID: 2971								
				Seq:	2363	16:18:52	08 Feb 02	HG
Hg	.029	ppb	.000					=

Reviewed by,

westchester county labs

Folder: RUNS1

Page 323

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 3073								
				Seq: 2377	16:44:23	08 Feb 02	HG	
Hg	-.017	ppb	.000	-.017				
*** Sample ID: 3074								
				Seq: 2378	16:46:41	08 Feb 02	HG	
Hg	-.014	ppb	.000	-.014				
*** Sample ID: 3075								
				Seq: 2379	16:48:19	08 Feb 02	HG	
Hg	.049	ppb	.000	.049				
*** Sample ID: 3076								
				Seq: 2380	16:50:30	08 Feb 02	HG	
Hg	.087	ppb	.000	.087				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		101. ✓	1.01	1.00	ppb	.000		
*** Sample ID: 3077								
				Seq: 2382	16:56:08	08 Feb 02	HG	
Hg	-.021	ppb	.000	-.021				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		103. ✓	1.03	1.00	ppb	.000		

