

westchester county labs

Folder: RUNS1

Page 324

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq: 2415	15:11:34	13 Feb 02	HG	
Hg	.000	ppb	2675					
*** Standard: 2 Rep: 1								
				Seq: 2416	15:14:01	13 Feb 02	HG	
Hg	.400	ppb	24831					
*** Standard: 3 Rep: 1								
				Seq: 2417	15:16:08	13 Feb 02	HG	
Hg	1.00	ppb	58452					
*** Standard: 4 Rep: 1								
				Seq: 2418	15:18:16	13 Feb 02	HG	
Hg	3.00	ppb	191751					
*** Standard: 5 Rep: 1								
				Seq: 2419	15:20:05	13 Feb 02	HG	
Hg	5.00	ppb	298205					
*** Standard: 6 Rep: 1								
				Seq: 2420	15:21:52	13 Feb 02	HG	
Hg	10.0	ppb	642901					
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		88.3	.883	1.00	ppb	.000		
*** Sample ID: 3179 pot								
				Seq: 2422	15:25:19	13 Feb 02	HG	
Hg	.102	ppb	.000	.102				
*** Sample ID: 3179 spk								
				Seq: 2423	15:27:18	13 Feb 02	HG	
Hg	4.71	ppb	.000	4.71				
*** Sample ID: 3179d								
				Seq: 2424	15:29:04	13 Feb 02	HG	
Hg	.153	ppb	.000	.153				
*** Sample ID: 3179d spk								
				Seq: 2425	15:31:20	13 Feb 02	HG	
Hg	4.81	ppb	.000	4.81				
*** Sample ID: 3180								
				Seq: 2426	15:33:16	13 Feb 02	HG	
Hg	.093	ppb	.000	.093				
*** Sample ID: 3181								
				Seq: 2427	15:35:02	13 Feb 02	HG	
Hg	.024	ppb	.000	.024				
*** Sample ID: 3182								
				Seq: 2428	15:37:00	13 Feb 02	HG	
Hg	.054	ppb	.000	.054				

Reviewed by:

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 325

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 3183				Seq: 2429	15:38:36	13 Feb 02	HG	
Hg	.058	ppb	.000	.058				=
*** Sample ID: 3184				Seq: 2430	15:40:13	13 Feb 02	HG	
Hg	.120	ppb	.000	.120				=
*** Sample ID: 3185				Seq: 2431	15:41:50	13 Feb 02	HG	
Hg	.095	ppb	.000	.095				=
*** Check Standard: 2 Ck21.0 qc				Seq: 2432	15:43:27	13 Feb 02	HG	
Line Flag %Rcv.		Found	True	Units	SD/RSD			
Hg	99.8	✓	.998	1.00	ppb	.000		=
*** Sample ID: 3245				Seq: 2433	15:45:13	13 Feb 02	HG	
Hg	.050	ppb	.000	.050				=
*** Sample ID: 3246				Seq: 2434	15:47:10	13 Feb 02	HG	
Hg	.032	ppb	.000	.032				=
*** Sample ID: 3247				Seq: 2435	15:48:46	13 Feb 02	HG	
Hg	.002	ppb	.000	.002				=
*** Sample ID: 3249				Seq: 2436	15:50:43	13 Feb 02	HG	
Hg	.040	ppb	.000	.040				=
*** Sample ID: 3250				Seq: 2437	15:52:24	13 Feb 02	HG	
Hg	.078	ppb	.000	.078				=
*** Sample ID: 3251				Seq: 2438	15:54:14	13 Feb 02	HG	
Hg	.105	ppb	.000	.105				=
*** Sample ID: 3253				Seq: 2439	15:55:52	13 Feb 02	HG	
Hg	.118	ppb	.000	.118				=
*** Sample ID: 3254				Seq: 2440	15:58:04	13 Feb 02	HG	
Hg	.065	ppb	.000	.065				=
*** Sample ID: 3255				Seq: 2441	15:59:43	13 Feb 02	HG	
Hg	.010	ppb	.000	.010				=

westchester county labs

Folder: RUNS1

Page 326

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
------	-------	-------	--------	---	---	---	---	---

*** Sample ID: 3256

Seq: 2442 16:01:25 13 Feb 02 HG

Hg	.039	ppb	.000	.039				
----	------	-----	------	------	--	--	--	--

*** Check Standard: 2 Ck21.0 qc Seq: 2443 16:03:14 13 Feb 02 HG

Line	Flag	%Rcv.	Found	True	Units	SD/RSD
Hg		92.2	.922	1.00	ppb	.000

