

westchester county labs

Folder: RUNS1

Page 335

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq: 2444	16:07:35	13 Feb 02	HG	
Hg	.000	ppb	5214					
*** Standard: 2 Rep: 1								
				Seq: 2445	16:09:13	13 Feb 02	HG	
Hg	.400	ppb	25530					
*** Standard: 3 Rep: 1								
				Seq: 2446	16:11:05	13 Feb 02	HG	
Hg	1.00	ppb	57325					
*** Standard: 4 Rep: 1								
				Seq: 2447	16:12:44	13 Feb 02	HG	
Hg	3.00	ppb	186997					
*** Standard: 5 Rep: 1								
				Seq: 2448	16:14:21	13 Feb 02	HG	
Hg	5.00	ppb	301436					
*** Standard: 6 Rep: 1								
				Seq: 2449	16:16:08	13 Feb 02	HG	
Hg	10.0	ppb	653402					
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		103.	1.03	1.00	ppb	.000		
*** Sample ID: 3257								
				Seq: 2460	16:39:34	13 Feb 02	HG	
Hg	.051	ppb	.000	.051				
*** Sample ID: 3258								
				Seq: 2461	16:41:24	13 Feb 02	HG	
Hg	.061	ppb	.000	.061				
*** Sample ID: 3259								
				Seq: 2462	16:43:13	13 Feb 02	HG	
Hg	.063	ppb	.000	.063				
*** Sample ID: 3263								
				Seq: 2463	16:44:49	13 Feb 02	HG	
Hg	.138	ppb	.000	.138				
*** Sample ID: 3264								
				Seq: 2464	16:46:27	13 Feb 02	HG	
Hg	.196	ppb	.000	.196				
*** Sample ID: 3281								
				Seq: 2465	16:48:08	13 Feb 02	HG	
Hg	.093	ppb	.000	.093				
*** Sample ID: 3282								
				Seq: 2466	16:49:45	13 Feb 02	HG	
Hg	.054	ppb	.000	.054				

Reviewed by,


westchester county labs

Folder: RUNS1

Page 336

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 3380								
				Seq: 2467	16:51:32	13 Feb 02	HG	
Hg	.053	ppb	.000	.053				
*** Sample ID: 3381								
				Seq: 2468	16:53:19	13 Feb 02	HG	
Hg	.141	ppb	.000	.141				
*** Sample ID: 3382								
				Seq: 2469	16:55:09	13 Feb 02	HG	
Hg	.053	ppb	.000	.053				
*** Check Standard 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		103. ✓	1.03	1.00	ppb	.000		
*** Sample ID: 3384								
				Seq: 2471	16:59:08	13 Feb 02	HG	
Hg	.097	ppb	.000	.097				
*** Sample ID: 3385								
				Seq: 2472	17:00:47	13 Feb 02	HG	
Hg	.057	ppb	.000	.057				
*** Sample ID: 3386								
				Seq: 2473	17:03:45	13 Feb 02	HG	
Hg	.058	ppb	.000	.058				
*** Sample ID: 3388								
				Seq: 2474	17:05:27	13 Feb 02	HG	
Hg	.055	ppb	.000	.055				
*** Sample ID: 3389								
				Seq: 2475	17:07:19	13 Feb 02	HG	
Hg	.103	ppb	.000	.103				
*** Sample ID: 3390								
				Seq: 2476	17:09:27	13 Feb 02	HG	
Hg	.109	ppb	.000	.109				
*** Sample ID: 3390 spk 123.4% (second spk O.K.)								
Hg	6.17	ppb	.000	6.17				
*** Sample ID: 3390d								
				Seq: 2478	17:12:58	13 Feb 02	HG	
Hg	.022	ppb	.000	.022				
*** Sample ID: 3390d spk 4.95 (second spk O.K.)								
Hg	4.95	ppb	.000	4.95				

westchester county labs

Folder: RUNS1

Page 337

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
------	-------	-------	--------	---	---	---	---	---

*** Check Standard: 2 / Ck21.0 qc Seq: 2481 17:18:21 13 Feb 02 HG

Line	Flag	%Rcv.	Found	True	Units	SD/RSD
Hg		96.8	.968	1.00	ppb	.000

