

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 383

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1				Seq: 2843	16:07:30	05 Mar 02	HG	
Hg	.000	ppb	3976					
*** Standard: 2 Rep: 1				Seq: 2844	16:10:21	05 Mar 02	HG	
Hg	.400	ppb	27534					
*** Standard: 3 Rep: 1				Seq: 2845	16:12:48	05 Mar 02	HG	
Hg	1.00	ppb	78645					
*** Standard: 4 Rep: 1				Seq: 2846	16:14:48	05 Mar 02	HG	
Hg	3.00	ppb	199756					
*** Standard: 5 Rep: 1				Seq: 2847	16:16:51	05 Mar 02	HG	
Hg	5.00	ppb	319926					
*** Standard: 6 Rep: 1				Seq: 2848	16:18:30	05 Mar 02	HG	
Hg	10.0	ppb	620041					
*** Check Standards: 2 Ck21.0 qc				Seq: 2849	16:20:21	05 Mar 02	HG	
Line Flag %Rcv Found True Units SD/RSD								
Hg	104.	✓	1.04	1.00	ppb	.000		
*** Sample ID: 4847				Seq: 2850	16:21:59	05 Mar 02	HG	
Hg	-.036	ppb	.000	-.036				
*** Sample ID: 4848				Seq: 2851	16:23:49	05 Mar 02	HG	
Hg	-.023	ppb	.000	-.023				
*** Sample ID: 4849				Seq: 2852	16:26:15	05 Mar 02	HG	
Hg	.005	ppb	.000	.005				
*** Sample ID: 4941				Seq: 2853	16:28:07	05 Mar 02	HG	
Hg	-.031	ppb	.000	-.031				
*** Sample ID: 4942				Seq: 2854	16:29:44	05 Mar 02	HG	
Hg	-.074	ppb	.000	-.074				
*** Sample ID: 4943				Seq: 2855	16:31:21	05 Mar 02	HG	
Hg	-.104	ppb	.000	-.104				
*** Sample ID: 4944				Seq: 2856	16:33:01	05 Mar 02	HG	
Hg	-.073	ppb	.000	-.073				

Reviewed by,
SPB/29/02

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 4945								
				Seq:	2857	16:34:38	05 Mar 02	HG
Hg	-.200	ppb	.000					
*** Sample ID: 4946								
				Seq:	2858	16:36:40	05 Mar 02	HG
Hg	.027	ppb	.000					
*** Sample ID: 4947								
				Seq:	2859	16:38:18	05 Mar 02	HG
Hg	.027	ppb	.000					
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Seq:	2860	16:40:15	05 Mar 02
Hg		111.	1.11	1.00	Units		SD/RSD	HG
					ppb		.000	
*** Sample ID: 5031								
				Seq:	2861	16:42:05	05 Mar 02	HG
Hg	-.058	ppb	.000					
*** Sample ID: 5032								
				Seq:	2862	16:44:15	05 Mar 02	HG
Hg	-.111	ppb	.000					
*** Sample ID: 5033								
				Seq:	2863	16:45:55	05 Mar 02	HG
Hg	-.083	ppb	.000					
*** Sample ID: 5035								
				Seq:	2864	16:48:12	05 Mar 02	HG
Hg	-.106	ppb	.000					
*** Sample ID: 5035 spk								
				Seq:	2865	16:49:51	05 Mar 02	HG
Hg	5.15	ppb	.000					
*** Sample ID: 5035d								
				Seq:	2866	16:51:53	05 Mar 02	HG
Hg	-.001	ppb	.000					
*** Sample ID: 5035d spk								
				Seq:	2867	16:53:32	05 Mar 02	HG
Hg	5.12	ppb	.000					
*** Sample ID: 5036								
				Seq:	2868	16:55:39	05 Mar 02	HG
Hg	-.051	ppb	.000					
*** Sample ID: 5037								
				Seq:	2869	16:57:38	05 Mar 02	HG
Hg	-.104	ppb	.000					

westchester county labs

Folder: RUNS1

Page 385

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 5050								
				Seq: 2870	16:59:19	05 Mar 02	HG	
Hg	-.101	ppb	.000	-.101				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		107.	1.07	1.00	ppb	.000		
*** Sample ID: 5051								
				Seq: 2872	17:02:59	05 Mar 02	HG	
Hg	-.010	ppb	.000	-.010				
*** Sample ID: 5052								
				Seq: 2873	17:04:48	05 Mar 02	HG	
Hg	.008	ppb	.000	.008				
*** Sample ID: 5053								
				Seq: 2874	17:06:27	05 Mar 02	HG	
Hg	-.027	ppb	.000	-.027				
*** Sample ID: 5054								
				Seq: 2875	17:08:04	05 Mar 02	HG	
Hg	-.051	ppb	.000	-.051				
*** Sample ID: 5055								
				Seq: 2876	17:09:51	05 Mar 02	HG	
Hg	-.081	ppb	.000	-.081				
*** Sample ID: 5056								
				Seq: 2877	17:12:17	05 Mar 02	HG	
Hg	-.063	ppb	.000	-.063				
*** Sample ID: 5062								
				Seq: 2878	17:13:57	05 Mar 02	HG	
Hg	-.159	ppb	.000	-.159				
*** Sample ID: 5063								
				Seq: 2879	17:15:34	05 Mar 02	HG	
Hg	.008	ppb	.000	.008				
*** Sample ID: 5151								
				Seq: 2880	17:17:42	05 Mar 02	HG	
Hg	.027	ppb	.000	.027				
*** Sample ID: 5152								
				Seq: 2881	17:19:29	05 Mar 02	HG	
Hg	-.029	ppb	.000	-.029				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		104.	1.04	1.00	ppb	.000		

WinHg Database 1.3

File Utility Help

Protocol **hg-cv** Dataset/Proto **RUNS1 /hg-cv**

Protocol | **Line info** | **Cal Curve** | **Report** | **Ctrl Chart** | **Viewer**

Reset

Calb Coeffs

New Cal

Update Coeffs

Spike Coeffs

A

B

C

Rho

Type

Accept

Calibrated ☒

Accepted ☒

Rel Abs

Accepted ☐

New ☐

Include ☐ ST ☒ Rep 1 ☐ ☐ ☐ ☐ ☐

Conc 10.0

S	Conc	Calc	Dev	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3
01	0.0000	-.092	-.092	3977	1	3976		
02	.40000	.291	-.109	27534	0%	27534		
03	1.0000	1.12	.123	78645	0%	78645		
04	3.0000	3.09	.094	199757	0%	199756		
05	5.0000	5.05	.050	319927	0%	319926		
06	10.000	9.93	-.066	620042	0%	620041		

Ready

NUM