

westchester county labs

Folder: RUNS2

Page 391

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
	Seq: 1			14:41:56 08 Mar 02				HG
Hg	.000	ppb	8293					
*** Standard: 2 Rep: 1								
	Seq: 2			14:43:33 08 Mar 02				HG
Hg	.400	ppb	25922					
*** Standard: 3 Rep: 1								
	Seq: 3			14:45:42 08 Mar 02				HG
Hg	1.00	ppb	54699					
*** Standard: 4 Rep: 1								
	Seq: 4			14:47:19 08 Mar 02				HG
Hg	3.00	ppb	153685					
*** Standard: 5 Rep: 1								
	Seq: 5			14:48:57 08 Mar 02				HG
Hg	5.00	ppb	295145					
*** Standard: 6 Rep: 1								
	Seq: 6			14:50:54 08 Mar 02				HG
Hg	10.0	ppb	599096					
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv	Found	True	Units	SD/RSD		
Hg		108. ✓	1.08	1.00	ppb	.000		
*** Sample ID: 5206 pot								
	Seq: 8			14:54:22 08 Mar 02				HG
Hg	.269	ppb	.000	.269				
*** Sample ID: 5209								
	Seq: 9			14:56:08 08 Mar 02				HG
Hg	.224	ppb	.000	.224				
*** Sample ID: 5211								
	Seq: 10			14:58:18 08 Mar 02				HG
Hg	.173	ppb	.000	.173				
*** Sample ID: 5212								
	Seq: 11			15:00:35 08 Mar 02				HG
Hg	.167	ppb	.000	.167				
*** Sample ID: 5228								
	Seq: 12			15:02:27 08 Mar 02				HG
Hg	.157	ppb	.000	.157				
*** Sample ID: 5230								
	Seq: 13			15:04:14 08 Mar 02				HG
Hg	.172	ppb	.000	.172				
*** Sample ID: 5231								
	Seq: 14			15:06:12 08 Mar 02				HG
Hg	.194	ppb	.000	.194				

NO DEP

Reviewed by,
JBS 6/20/02

westchester county labs

Folder: RUNS2

Page 392

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 5231 spk								
				Seq: 15	15:07:49	08 Mar 02	HG	
Hg	5.48	ppb	.000	5.48				
*** Sample ID: 5231d								
				Seq: 16	15:10:36	08 Mar 02	HG	
Hg	.257	ppb	.000	.257				
*** Sample ID: 5231d spk								
				Seq: 17	15:12:17	08 Mar 02	HG	
Hg	5.58	ppb	.000	5.58				
*** Check Standard: 2 Ck21.0 qc								
				Seq: 18	15:14:28	08 Mar 02	HG	
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg	H	122.	1.22	1.00	ppb	.000		

