

westchester county labs

Folder: RUNS2
Protocol: hg-cv
POST-RUN REPORT

Page 397

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq: 19		15:18:21	08 Mar 02	HG
Hg	.000	ppb	7530					
*** Standard: 2 Rep: 1								
				Seq: 20		15:19:57	08 Mar 02	HG
Hg	.400	ppb	30917					
*** Standard: 3 Rep: 1								
				Seq: 21		15:21:39	08 Mar 02	HG
Hg	1.00	ppb	68645					
*** Standard: 4 Rep: 1								
				Seq: 22		15:23:18	08 Mar 02	HG
Hg	3.00	ppb	183940					
*** Standard: 5 Rep: 1								
				Seq: 23		15:25:10	08 Mar 02	HG
Hg	5.00	ppb	325056					
*** Standard: 6 Rep: 1								
				Seq: 24		15:27:09	08 Mar 02	HG
Hg	10.0	ppb	663214					
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		104. ✓	1.04	1.00	ppb	.000		
*** Sample ID: 5232								
				Seq: 26		15:31:50	08 Mar 02	HG
Hg	.086	ppb	.000	.086				
*** Sample ID: 5233								
				Seq: 27		15:33:38	08 Mar 02	HG
Hg	.131	ppb	.000	.131				
*** Sample ID: 5235								
				Seq: 28		15:35:14	08 Mar 02	HG
Hg	.144	ppb	.000	.144				
*** Sample ID: 5236								
				Seq: 29		15:36:55	08 Mar 02	HG
Hg	.127	ppb	.000	.127				
*** Sample ID: 5237								
				Seq: 30		15:39:55	08 Mar 02	HG
Hg	.053	ppb	.000	.053				
*** Sample ID: 5239								
				Seq: 31		15:41:47	08 Mar 02	HG
Hg	.052	ppb	.000	.052				
*** Sample ID: 5240								
				Seq: 32		15:43:24	08 Mar 02	HG
Hg	.121	ppb	.000	.121				

Reviewed by,
JB 6/25/02

westchester county labs

Folder: RUNS2

Page 398

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 5241								
				Seq: 33	15:45:21	08 Mar 02	HG	
Hg	.189	ppb	.000	.189				
*** Sample ID: 5244								
				Seq: 34	15:47:54	08 Mar 02	HG	
Hg	.141	ppb	.000	.141				
*** Sample ID: 5246								
				Seq: 35	15:49:43	08 Mar 02	HG	
Hg	.096	ppb	.000	.096				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		107. ✓	1.07	1.00	ppb	.000		
*** Sample ID: 5248								
				Seq: 37	15:53:09	08 Mar 02	HG	
Hg ✓	.074	ppb	.000	.074				
*** Sample ID: 5250								
				Seq: 38	15:55:26	08 Mar 02	HG	
Hg ✓	.133	ppb	.000	.133				
*** Sample ID: 5286								
				Seq: 39	15:57:04	08 Mar 02	HG	
Hg ✓	.137	ppb	.000	.137				
*** Sample ID: 5301								
				Seq: 40	15:58:53	08 Mar 02	HG	
Hg ✓	.145	ppb	.000	.145				
*** Sample ID: 5302								
				Seq: 41	16:00:29	08 Mar 02	HG	
Hg ✓	.103	ppb	.000	.103				
*** Sample ID: 5303								
				Seq: 42	16:02:16	08 Mar 02	HG	
Hg ✓	.093	ppb	.000	.093				
*** Sample ID: 5304								
				Seq: 43	16:04:03	08 Mar 02	HG	
Hg ✓	.090	ppb	.000	.090				
*** Sample ID: 5305								
				Seq: 44	16:05:41	08 Mar 02	HG	
Hg ✓	.078	ppb	.000	.078				
*** Sample ID: 5306								
				Seq: 45	16:07:20	08 Mar 02	HG	
Hg ✓	.109	ppb	.000	.109				

westchester county labs

Folder: RUNS2

Page 399

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 5307								
				Seq: 46	16:09:17	08 Mar 02	HG	
Hg	✓ .148	ppb	.000	.148				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		113. ✓	1.13	1.00	ppb	.000		
*** Sample ID: 5403								
				Seq: 48	16:12:36	08 Mar 02	HG	
Hg	✓ .080	ppb	.000	.080				
*** Sample ID: 5405								
				Seq: 49	16:14:33	08 Mar 02	HG	
Hg	✓ .071	ppb	.000	.071				
*** Sample ID: 5436								
				Seq: 50	16:16:35	08 Mar 02	HG	
Hg	✓ .089	ppb	.000	.089				
*** Sample ID: 5512								
				Seq: 51	16:18:22	08 Mar 02	HG	
Hg	✓ .096	ppb	.000	.096				
*** Sample ID: 5513								
				Seq: 52	16:20:33	08 Mar 02	HG	
Hg	✓ .140	ppb	.000	.140				
*** Sample ID: 5514								
				Seq: 53	16:22:11	08 Mar 02	HG	
Hg	✓ .143	ppb	.000	.143				
*** Sample ID: 5515								
				Seq: 54	16:23:58	08 Mar 02	HG	
Hg	✓ .142	ppb	.000	.142				
*** Sample ID: 5516								
				Seq: 55	16:25:57	08 Mar 02	HG	
Hg	✓ .074	ppb	.000	.074				
*** Sample ID: 5517								
				Seq: 56	16:27:36	08 Mar 02	HG	
Hg	✓ .082	ppb	.000	.082				
*** Sample ID: 5518								
				Seq: 57	16:29:25	08 Mar 02	HG	
Hg	✓ .083	ppb	.000	.083				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		105. ✓	1.05	1.00	ppb	.000		

westchester county labs

Folder: RUNS2

Page 400

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 5518spk Seq: 59 16:33:10 08 Mar 02 HG								
Hg	5.13	ppb	.000	5.13				
*** Sample ID: 5518d Seq: 60 16:34:47 08 Mar 02 HG								
Hg	.091	ppb	.000	.091				
*** Sample ID: 5518d spk Seq: 61 16:36:26 08 Mar 02 HG								
Hg	5.22	ppb	.000	5.22				
*** Check Standard: 2 Ck21.0 qc Seq: 69 16:51:15 08 Mar 02 HG								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg	H	117. ✓	1.17	1.00	ppb	.000		

LMDL
OK
(o.c. to report c.mpl values)

