

westchester county labs

Folder: RUNS2

Page 436

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq: 338	16:46:39	19 Mar 02	HG	
Hg	.000	ppb	761					=
*** Standard: 2 Rep: 1								
				Seq: 339	16:48:59	19 Mar 02	HG	
Hg	.400	ppb	23777					=
*** Standard: 3 Rep: 1								
				Seq: 340	16:50:42	19 Mar 02	HG	
Hg	1.00	ppb	59712					=
*** Standard: 4 Rep: 1								
				Seq: 341	16:52:41	19 Mar 02	HG	
Hg	3.00	ppb	196734					=
*** Standard: 5 Rep: 1								
				Seq: 342	16:56:15	19 Mar 02	HG	
Hg	5.00	ppb	334364					=
*** Standard: 6 Rep: 1								
				Seq: 343	16:58:02	19 Mar 02	HG	
Hg	10.0	ppb	674399					=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		89.5	.895	1.00	ppb	.000		=
*** Sample ID: 6092 pot								
				Seq: 345	17:01:39	19 Mar 02	HG	
Hg	L .052	ppb	.000	.052				=
*** Sample ID: 6092 spk								
				Seq: 346	17:03:15	19 Mar 02	HG	
Hg	5.21	ppb	.000	5.21				=
*** Sample ID: 6092d								
				Seq: 347	17:05:02	19 Mar 02	HG	
Hg	.088	ppb	.000	.088				=
*** Sample ID: 6092d spk								
				Seq: 348	17:07:23	19 Mar 02	HG	
Hg	4.74	ppb	.000	4.74				=
*** Sample ID: 6301								
				Seq: 349	17:09:10	19 Mar 02	HG	
Hg	L .136	ppb	.000	.136				=
*** Sample ID: 6302								
				Seq: 350	17:10:56	19 Mar 02	HG	
Hg	L .084	ppb	.000	.084				=
*** Sample ID: 6303								
				Seq: 351	17:12:52	19 Mar 02	HG	
Hg	L .098	ppb	.000	.098				=

Reviewed by:

westchester county labs

Folder: RUNS2

Page 437

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 6305								
				Seq: 352	17:14:34	19 Mar 02	HG	
Hg	L .037	ppb	.000	.037				
*** Sample ID: 6306								
				Seq: 353	17:16:16	19 Mar 02	HG	
Hg	L .053	ppb	.000	.053				
*** Sample ID: 6307								
				Seq: 354	17:18:13	19 Mar 02	HG	
Hg	L .065	ppb	.000	.065				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		91.5	.915	1.00	ppb	.000		
*** Sample ID: 6333								
				Seq: 356	17:21:49	19 Mar 02	HG	
Hg	L .120	ppb	.000	.120				
*** Sample ID: 6334								
				Seq: 357	17:23:26	19 Mar 02	HG	
Hg	L .074	ppb	.000	.074				
*** Sample ID: 6335								
				Seq: 358	17:25:03	19 Mar 02	HG	
Hg	L .043	ppb	.000	.043				
*** Sample ID: 6336								
				Seq: 359	17:26:50	19 Mar 02	HG	
Hg	L .042	ppb	.000	.042				
*** Sample ID: 6337								
				Seq: 360	17:28:27	19 Mar 02	HG	
Hg	L .039	ppb	.000	.039				
*** Sample ID: 6337								
				Seq: 361	17:30:13	19 Mar 02	HG	
Hg	-.087	ppb	.000	-.087				
*** Sample ID: 6338								
				Seq: 362	17:32:11	19 Mar 02	HG	
Hg	L .114	ppb	.000	.114				
*** Sample ID: 6339								
				Seq: 363	17:34:44	19 Mar 02	HG	
Hg	L .109	ppb	.000	.109				
*** Sample ID: 6406								
				Seq: 364	17:36:20	19 Mar 02	HG	
Hg	L .086	ppb	.000	.086				

westchester county labs

Folder: RUNS2

Page 438

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 6407				Seq: 365	17:38:01	19 Mar 02	HG	
Hg	<i>L</i> .037	ppb	.000	.037				
*** Check Standard: 2	Ck21.0 qc			Seq: 366	17:39:38	19 Mar 02	HG	
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		86.9	✓.869	1.00	ppb	.000		
*** Sample ID: 6408				Seq: 367	17:41:19	19 Mar 02	HG	
Hg	<i>L</i> .048	ppb	.000	.048				
*** Sample ID: 6409				Seq: 368	17:42:57	19 Mar 02	HG	
Hg	<i>L</i> .035	ppb	.000	.035				
*** Sample ID: 6410				Seq: 369	17:45:37	19 Mar 02	HG	
Hg	<i>L</i> .120	ppb	.000	.120				
*** Sample ID: 6411				Seq: 370	17:47:25	19 Mar 02	HG	
Hg	<i>L</i> .103	ppb	.000	.103				
*** Sample ID: 6412				Seq: 371	17:49:01	19 Mar 02	HG	
Hg	<i>L</i> .083	ppb	.000	.083				
*** Sample ID: 6412 spk				Seq: 372	17:51:39	19 Mar 02	HG	
Hg	5.12	ppb	.000	5.12				
*** Sample ID: 6412d				Seq: 373	17:53:15	19 Mar 02	HG	
Hg	.035	ppb	.000	.035				
*** Sample ID: 6412d spk				Seq: 374	17:54:57	19 Mar 02	HG	
Hg	5.07	ppb	.000	5.07				
*** Sample ID: 6423				Seq: 375	17:56:35	19 Mar 02	HG	
Hg	<i>L</i> .018	ppb	.000	.018				
*** Sample ID: 6424				Seq: 376	17:59:13	19 Mar 02	HG	
Hg	<i>L</i> .104	ppb	.000	.104				
*** Check Standard: 2	Ck21.0 qc			Seq: 377	18:02:11	19 Mar 02	HG	
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		94.9	✓.949	1.00	ppb	.000		

