

westchester county labs

Folder: RUNS2
Protocol: hg-cv
POST-RUN REPORT

Page 444

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq:	395	14:37:19	22 Mar 02	HG
Hg	.000	ppb	6346					=
*** Standard: 2 Rep: 1								
				Seq:	396	14:39:58	22 Mar 02	HG
Hg	.400	ppb	25817					=
*** Standard: 3 Rep: 1								
				Seq:	397	14:41:50	22 Mar 02	HG
Hg	1.00	ppb	65789					=
*** Standard: 4 Rep: 1								
				Seq:	398	14:43:31	22 Mar 02	HG
Hg	3.00	ppb	198243					=
*** Standard: 5 Rep: 1								
				Seq:	399	14:45:11	22 Mar 02	HG
Hg	5.00	ppb	346897					=
*** Standard: 6 Rep: 1								
				Seq:	400	14:46:48	22 Mar 02	HG
Hg	10.0	ppb	677098					=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		97.9	.979	1.00	ppb	.000		=
*** Sample ID: 6425 pot								
				Seq:	402	14:51:09	22 Mar 02	HG
Hg	.046	ppb	.000	.046				=
*** Sample ID: 6425 spk								
				Seq:	403	14:53:49	22 Mar 02	HG
Hg	5.23	ppb	.000	5.23				=
*** Sample ID: 6425d								
				Seq:	404	14:55:35	22 Mar 02	HG
Hg	.126	ppb	.000	.126				=
*** Sample ID: 6425d spk								
				Seq:	405	14:57:21	22 Mar 02	HG
Hg	5.13	ppb	.000	5.13				=
*** Sample ID: 6427								
				Seq:	406	14:59:21	22 Mar 02	HG
Hg	.058	ppb	.000	.058				=
*** Sample ID: 6428								
				Seq:	407	15:01:03	22 Mar 02	HG
Hg	-.038	ppb	.000	-.038				=
*** Sample ID: 6429								
				Seq:	408	15:02:45	22 Mar 02	HG
Hg	-.008	ppb	.000	-.008				=

Reviewed by,
6/5/02

westchester county labs

Folder: RUNS2

Page 445

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 6449								
				Seq: 409	15:04:24	22 Mar 02	HG	
Hg	-.099	ppb	.000	-.099				
*** Sample ID: 6450								
				Seq: 410	15:06:33	22 Mar 02	HG	
Hg	.109	ppb	.000	.109				
*** Sample ID: 6451								
				Seq: 411	15:08:52	22 Mar 02	HG	
Hg	.041	ppb	.000	.041				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		92.6	.926	1.00	ppb	.000		
*** Sample ID: 6452								
				Seq: 413	15:12:09	22 Mar 02	HG	
Hg	.004	ppb	.000	.004				
*** Sample ID: 6453								
				Seq: 414	15:13:51	22 Mar 02	HG	
Hg	-.008	ppb	.000	-.008				
*** Sample ID: 6454								
				Seq: 415	15:15:51	22 Mar 02	HG	
Hg	.010	ppb	.000	.010				
*** Sample ID: 6455								
				Seq: 416	15:17:41	22 Mar 02	HG	
Hg	.037	ppb	.000	.037				
*** Sample ID: 6457								
				Seq: 417	15:19:27	22 Mar 02	HG	
Hg	.067	ppb	.000	.067				
*** Sample ID: 6458								
				Seq: 418	15:21:06	22 Mar 02	HG	
Hg	.041	ppb	.000	.041				
*** Sample ID: 6529								
				Seq: 419	15:23:45	22 Mar 02	HG	
Hg	.013	ppb	.000	.013				
*** Sample ID: 6530								
				Seq: 420	15:25:33	22 Mar 02	HG	
Hg	-.047	ppb	.000	-.047				
*** Sample ID: 6563								
				Seq: 421	15:27:10	22 Mar 02	HG	
Hg	-.003	ppb	.000	-.003				

westchester county labs

Folder: RUNS2

Page 446

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 6565								
				Seq: 422	15:29:27	22 Mar 02	HG	
Hg	L .048	ppb	.000	.048				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		98.8	.988	1.00	ppb	.000		
*** Sample ID: 6567								
				Seq: 424	15:34:19	22 Mar 02	HG	
Hg	L .049	ppb	.000	.049				
*** Sample ID: 6580								
				Seq: 425	15:36:09	22 Mar 02	HG	
Hg	-.036	ppb	.000	-.036				
*** Sample ID: 6581								
				Seq: 426	15:38:07	22 Mar 02	HG	
Hg	-.011	ppb	.000	-.011				
*** Sample ID: 6582								
				Seq: 427	15:39:57	22 Mar 02	HG	
Hg	L .034	ppb	.000	.034				
*** Sample ID: 6582 spk								
				Seq: 428	15:42:07	22 Mar 02	HG	
Hg	5.37	ppb	.000	5.37				
*** Sample ID: 6582d								
				Seq: 429	15:44:17	22 Mar 02	HG	
Hg	L .079	ppb	.000	.079				
*** Sample ID: 6582d spk								
				Seq: 430	15:46:05	22 Mar 02	HG	
Hg	5.42	ppb	.000	5.42				
*** Sample ID: 6584								
				Seq: 431	15:47:43	22 Mar 02	HG	
Hg	-.001	ppb	.000	-.001				
*** Sample ID: 6585								
				Seq: 432	15:49:20	22 Mar 02	HG	
Hg	-.060	ppb	.000	-.060				
*** Sample ID: 6586								
				Seq: 433	15:50:59	22 Mar 02	HG	
Hg	-.012	ppb	.000	-.012				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		97.6	.976	1.00	ppb	.000		

westchester county labs

Folder: RUNS2

Page 447

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 6588								
				Seq: 435	15:54:37	22 Mar 02	HG	
Hg	L.035	ppb	.000	.035				
*** Sample ID: 6589								
				Seq: 436	15:56:20	22 Mar 02	HG	
Hg	L.023	ppb	.000	.023				
*** Sample ID: 6590								
				Seq: 437	15:58:18	22 Mar 02	HG	
Hg	-.018	ppb	.000	-.018				
*** Sample ID: 6630								
				Seq: 438	15:59:56	22 Mar 02	HG	
Hg	-.031	ppb	.000	-.031				
*** Sample ID: 6631								
				Seq: 439	16:01:47	22 Mar 02	HG	
Hg	L.009	ppb	.000	.009				
*** Sample ID: 6632								
				Seq: 440	16:03:26	22 Mar 02	HG	
Hg	L.039	ppb	.000	.039				
*** Sample ID: 6633								
				Seq: 441	16:05:05	22 Mar 02	HG	
Hg	L.052	ppb	.000	.052				
*** Sample ID: 6634								
				Seq: 442	16:07:04	22 Mar 02	HG	
Hg	L.012	ppb	.000	.012				
*** Sample ID: 6635								
				Seq: 443	16:08:52	22 Mar 02	HG	
Hg	L.004	ppb	.000	.004				
*** Sample ID: 6636								
				Seq: 444	16:11:20	22 Mar 02	HG	
Hg	-.050	ppb	.000	-.050				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		93.8	.938	1.00	ppb	.000		
*** Sample ID: 6678								
				Seq: 446	16:15:09	22 Mar 02	HG	
Hg	L.066	ppb	.000	.066				
*** Sample ID: 6681								
				Seq: 447	16:18:44	22 Mar 02	HG	
Hg	L.001	ppb	.000	.001				

westchester county labs

Folder: RUNS2

Page 448

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 6617 np								
				Seq: 448	16:20:34	22 Mar 02	HG	
Hg	-.026	ppb	.000	-.026				
*** Sample ID: 6618 np								
				Seq: 449	16:22:17	22 Mar 02	HG	
Hg	-.042	ppb	.000	-.042				
*** Check Standard: 2 Ck21.0 qc								
				Seq: 450	16:25:18	22 Mar 02	HG	
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		95.2	.952	1.00	ppb	.000		

WinHg Runner 1.3

File Edit Utility Help

DB

Protocol: hg-cv State: IDLE Error: Clear

Dataset: RUNS2 Seq: 451 Batch: 032202 Action: 0

Gas: Off LPM Pump: Off mL/min

Main Control Standard Sample Output Report Utility

Identification

Site Name: westchester county labs

User Name: JIMMARSHALL

Statistics

Total Hours: 1015.05 Reset

Samples Run: 3361 Reset

Start

Rack Editor

Database

Printer Control

☐ Real-Time Printing

Next Page: 3299 Reset

Lines Per Page: 70

Lines Waiting: 0 Form feed

Notepad

NUM