

westchester county labs

Folder: RUNS2
Protocol: hg-cv
POST-RUN REPORT

Page 515

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq: 871	14:24:17	12 Apr 02	HG	
Hg	.000	ppb	-3623					
*** Standard: 2 Rep: 1								
				Seq: 872	14:26:17	12 Apr 02	HG	
Hg	.400	ppb	22624					
*** Standard: 3 Rep: 1								
				Seq: 873	14:28:36	12 Apr 02	HG	
Hg	1.00	ppb	58925					
*** Standard: 4 Rep: 1								
				Seq: 874	14:30:14	12 Apr 02	HG	
Hg	3.00	ppb	203468					
*** Standard: 5 Rep: 1								
				Seq: 875	14:31:56	12 Apr 02	HG	
Hg	5.00	ppb	353062					
*** Standard: 6 Rep: 1								
				Seq: 876	14:33:38	12 Apr 02	HG	
Hg	10.0	ppb	698614					
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		92.9	.929	1.00	ppb	.000		
*** Sample ID: 8425 pot								
				Seq: 878	14:37:07	12 Apr 02	HG	
Hg	.053	ppb	.000	.053				
*** Sample ID: 8425 spk								
				Seq: 879	14:38:43	12 Apr 02	HG	
Hg	5.40	ppb	.000	5.40				
*** Sample ID: 8425d								
				Seq: 880	14:41:21	12 Apr 02	HG	
Hg	.073	ppb	.000	.073				
*** Sample ID: 8425d spk								
				Seq: 881	14:43:11	12 Apr 02	HG	
Hg	5.46	ppb	.000	5.46				
*** Sample ID: 8426								
				Seq: 882	14:44:57	12 Apr 02	HG	
Hg	.050	ppb	.000	.050				
*** Sample ID: 8427								
				Seq: 883	14:46:43	12 Apr 02	HG	
Hg	.044	ppb	.000	.044				
*** Sample ID: 8429								
				Seq: 884	14:49:00	12 Apr 02	HG	
Hg	.101	ppb	.000	.101				

Reviewed by,
LB gbp/r

westchester county labs

Folder: RUNS2

Page 516

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 8430								
				Seq: 885	14:50:38	12 Apr 02	HG	
Hg	L.064	ppb	.000	.064				
*** Sample ID: 8431								
				Seq: 886	14:52:20	12 Apr 02	HG	
Hg	L.078	ppb	.000	.078				
*** Sample ID: 8433								
				Seq: 887	14:54:02	12 Apr 02	HG	
Hg	L.074	ppb	.000	.074				
*** Check Standard 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		102.	1.02	1.00	ppb	.000		
*** Sample ID: 8434								
				Seq: 889	14:57:32	12 Apr 02	HG	
Hg	L.075	ppb	.000	.075				
*** Sample ID: 8435								
				Seq: 890	14:59:12	12 Apr 02	HG	
Hg	L.057	ppb	.000	.057				
*** Check Standard 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		101.	1.01	1.00	ppb	.000		

NO DET

