

4/2/02 IT

NYC Sampler.

B1K

2 STD

IPC+CB

4383 Lab Dup + spk

4385 ✓

4386 ✓

4387 ✓

4410 ✓

4411 ✓

4412 ✓

4413 ✓

4414 ✓

4415 ✓

IPC+CB

4416 Lab Dup + spk

4418 ✓

4419 ✓

4449 ✓

4450 ✓

4509 ✓

4510 ✓

4512 ✓

4513 ✓

4514 ✓

IPC+CB

B1K

5 STD

IPC+CB

4FB+CRB

6528 Na

6567 Lab Dup + spk

6563

6565

6629

6678

6681

6683

6685

6687

IPC+CB

→ DE OR 5311, 5313, 5315, 5317, 5319
check

5321

5323

5325

5327

5329

IPC+CB

Reviewed by
JB 7/8/02

Si only


Standardization Rpt.

04/02/02 10:12:50 AM

page 1

Method: NYCW Standard: STD1-Blank

Run Time: 04/02/02 10:11:08



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	6.809	6.625	8.353	.0020	2.579	1.326
SDev	.012	.027	.028	.0016	.004	.007
%RSD	.1719	.4075	.3366	81.65	.1598	.5368
#1	6.796	6.634	8.358	.0000	2.576	1.320
#2	6.804	6.652	8.368	.0020	2.582	1.326
#3	6.822	6.588	8.374	.0040	2.576	1.336
#4	6.816	6.628	8.312	.0020	2.584	1.322

Standardization Rpt.

04/02/02 10:18:43 AM

page 1

Method: NYCW Standard: STD2
Run Time: 04/02/02 10:17:21



Elem	Ca3179	Co2286	Mg2790
Avge	73.38	187.2	8.877
SDev	.26	.9	.033
%RSD	.3488	.4579	.3754
#1	73.71	188.3	8.916
#2	73.23	186.5	8.846
#3	73.45	187.3	8.892
#4	73.13	186.5	8.852

Analysis Report

04/02/02 10:21:01 AM

page 1

Method: NYCW Sample Name: ipc

Run Time: 04/02/02 10:19:18

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	1.043	1.010	.9990	1.029	.9969	.9941
SDev	.013	.003	.0020	.008	.0066	.0153
%RSD	1.258	.2627	.1988	.7662	.6569	1.543
#1	1.050	1.011	1.000	1.019	.9882	.9736
#2	1.027	1.011	.9991	1.034	1.004	.9930
#3	1.057	1.006	1.000	1.025	.9977	.9998
#4	1.040	1.011	.9961	1.036	.9977	1.010

Analysis Report

04/02/02 10:27:24 AM

page 1

Method: NYCW Sample Name: cb

Run Time: 04/02/02 10:25:41

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0086	.0002	<.0000	<.0000	<.0000	.0161
SDev	.0084	.0022	.0027	.0026	.0138	.0050
%RSD	97.50	1403.	82.30	688.4	108.7	31.03
#1	.0104	.0028	<.0000	.0030	<.0000	.0161
#2	.0191	.0010	<.0000	<.0000	.0008	.0093
#3	<.0000	<.0000	<.0000	<.0000	<.0000	.0178
#4	.0061	<.0000	<.0000	<.0000	<.0000	.0211

LSO

LSO

Analysis Report

04/02/02 10:38:06 AM

page 1

Method: NYCW Sample Name: lfb

Operator: it

Run Time: 04/02/02 10:36:22

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.2237	.0087	.1007	.2671	<.0000	.1029
SDev	.0048	.0044	.0008	.0029	.0149	.0048
%RSD	2.133	50.60	.7832	1.078	56.67	4.673
#1	.2291	.0100	.1013	.2690	<.0000	.1036
#2	.2262	.0139	.1014	.2629	<.0000	.0959
#3	.2190	.0034	.1002	.2690	<.0000	.1061
#4	.2204	.0075	.0998	.2675	<.0000	.1061

90%

101

107

103

Analysis Report

04/02/02 10:40:32 AM

page 1

Method: NYCW Sample Name: lrb

Run Time: 04/02/02 10:38:45

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	<.0000	<.0000	<.0000	<.0000	<.0000
SDev	.0030	.0040	.0019	.0038	.0154	.0080
%RSD	14.71	91.73	67.34	508.5	69.39	316.1
#1	<.0000	<.0000	<.0000	.0000	<.0000	<.0000
#2	<.0000	<.0000	<.0000	<.0000	<.0000	.0030
#3	<.0000	.0007	<.0000	.0000	<.0000	.0030
#4	<.0000	<.0000	<.0000	.0030	<.0000	<.0000

Analysis Report

04/02/02 10:51:21 AM

page 1

Method: NYCW Sample Name: 4383

Run Time: 04/02/02 10:49:38

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0118	6.479	<.0000	.0355	1.476	<.0000
SDev	.0193	.020	.0014	.0033	.011	.0059
%RSD	163.7	.3146	150.3	9.249	.7456	135.4
#1	.0031	6.488	<.0000	.0400	1.486	<.0000
#2	.0291	6.471	.0001	.0324	1.480	<.0000
#3	.0262	6.501	.0004	.0355	1.480	.0037
#4	<.0000	6.454	<.0000	.0340	1.461	<.0000

Analysis Report

04/02/02 10:53:22 AM

page 1

Method: NYCW Sample Name: 4383lab dup

Operator: it

Run Time: 04/02/02 10:51:39

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.421	<.0000	.0355	1.436	<.0000
SDev	.0159	.021	.0031	.0039	.011	.0037
%RSD	181.5	.3307	264.5	11.05	.7749	43.82
#1	<.0000	6.434	<.0000	.0340	1.432	<.0000
#2	<.0000	6.397	.0011	.0370	1.442	<.0000
#3	.0104	6.409	<.0000	.0309	1.422	<.0000
#4	<.0000	6.442	.0017	.0400	1.448	<.0000

0.76 RPD

Analysis Report

04/02/02 11:05:18 AM

page 1

Method: NYCW Sample Name: 4383spk

Operator: it

Run Time: 04/02/02 11:03:35

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.2568	6.407	.0999	.2923	1.448	.0924
SDev	.0198	.052	.0039	.0048	.012	.0074
%RSD	7.700	.8046	3.887	1.642	.8012	8.048
#1	.2802	6.440	.1026	.2939	1.461	.1009
#2	.2528	6.461	.1037	.2984	1.454	.0958
#3	.2615	6.373	.0976	.2893	1.435	.0840
#4	.2326	6.354	.0957	.2878	1.441	.0891

96

100

102

92%

Analysis Report

04/02/02 11:08:17 AM

page 1

Method: NYCW Sample Name: 4385

Run Time: 04/02/02 11:06:34

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0129	6.446	.0003	.0343	1.467	<.0000
SDev	.0097	.025	.0021	.0023	.012	.0037
%RSD	75.13	.3885	800.7	6.639	.7914	65.13
#1	.0262	6.463	.0026	.0370	1.454	<.0000
#2	.0118	6.411	<.0000	.0324	1.461	<.0000
#3	.0104	6.445	<.0000	.0324	1.480	<.0000
#4	.0031	6.465	.0012	.0355	1.473	<.0000

Analysis Report

04/02/02 11:10:46 AM

page 1

Method: NYCW Sample Name: 4386

Run Time: 04/02/02 11:09:03

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0107	6.455	<.0000	.0370	1.451	<.0000
SDev	.0213	.013	.0015	.0048	.015	.0046
%RSD	199.1	.1941	251.2	12.99	1.043	63.36
#1	.0262	6.457	.0004	.0355	1.467	<.0000
#2	.0306	6.439	<.0000	.0400	1.442	<.0000
#3	.0002	6.470	.0003	.0309	1.435	<.0000
#4	<.0000	6.454	<.0000	.0416	1.461	<.0000

Analysis Report

04/02/02 11:13:04 AM

page 1

Method: NYCW Sample Name: 4387

Operator: it

Run Time: 04/02/02 11:11:21

Comment:

Mode: CONC Corr. Factor: 1



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	24.20	<.0000	.0547	8.183	<.0000
SDev	.0216	.11	.0024	.0033	.040	.0021
%RSD	219.6	.4341	278.0	5.992	.4926	24.14
#1	.0006	24.21	<.0000	.0563	8.172	<.0000
#2	<.0000	24.14	<.0000	.0578	8.159	<.0000
#3	.0150	24.34	.0012	.0547	8.242	<.0000
#4	<.0000	24.11	<.0000	.0502	8.156	<.0000

Analysis Report

04/02/02 11:15:26 AM

page 1

Method: NYCW Sample Name: 4410

Operator: it

Run Time: 04/02/02 11:13:43

Comment:

Mode: CONC Corr. Factor: 1



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0085	6.541	.0009	.0423	1.484	<.0000
SDev	.0149	.022	.0016	.0029	.025	.0037
%RSD	175.0	.3304	180.4	6.894	1.700	34.22
#1	.0118	6.522	.0011	.0446	1.489	<.0000
#2	.0060	6.528	.0005	.0416	1.461	<.0000
#3	<.0000	6.542	<.0000	.0446	1.470	<.0000
#4	.0262	6.570	.0030	.0385	1.518	<.0000

Analysis Report

04/02/02 11:18:26 AM

page 1

Method: NYCW Sample Name: 4411

Run Time: 04/02/02 11:16:43

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.493	<.0000	.0385	1.468	<.0000
SDev	.0126	.024	.0037	.0028	.019	.0040
%RSD	260.8	.3686	194.0	7.203	1.280	49.43
#1	<.0000	6.521	<.0000	.0370	1.442	<.0000
#2	.0075	6.482	.0034	.0400	1.473	<.0000
#3	<.0000	6.466	<.0000	.0355	1.486	<.0000
#4	.0017	6.501	<.0000	.0416	1.470	<.0000

Analysis Report

04/02/02 11:20:54 AM

page 1

Method: NYCW Sample Name: 4412

Run Time: 04/02/02 11:19:11

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.526	<.0000	.0389	1.446	<.0000
SDev	.0198	.023	.0007	.0034	.011	.0034
%RSD	531.4	.3502	52.38	8.658	.7286	25.48
#1	<.0000	6.496	<.0000	.0355	1.435	<.0000
#2	.0233	6.539	<.0000	.0400	1.461	<.0000
#3	<.0000	6.549	<.0000	.0370	1.445	<.0000
#4	<.0000	6.521	<.0000	.0431	1.445	<.0000

Analysis Report

04/02/02 11:22:50 AM

page 1

Method: NYCW Sample Name: 4413

Run Time: 04/02/02 11:21:07

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	7.202	<.0000	.0369	1.747	<.0000
SDev	.0214	.025	.0026	.0033	.016	.0033
%RSD	139.1	.3498	82.34	8.902	.8959	25.86
#1	.0146	7.168	.0006	.0384	1.766	<.0000
#2	<.0000	7.201	<.0000	.0369	1.750	<.0000
#3	<.0000	7.215	<.0000	.0399	1.747	<.0000
#4	<.0000	7.226	<.0000	.0323	1.728	<.0000

Analysis Report

04/02/02 11:25:12 AM

page 1

Method: NYCW Sample Name: 4414

Run Time: 04/02/02 11:23:29

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	10.28	<.0000	.0473	2.874	<.0000
SDev	.0146	.04	.0021	.0049	.014	.0029
%RSD	84.82	.4173	100.4	10.29	.4725	25.20
#1	<.0000	10.24	.0004	.0454	2.868	<.0000
#2	<.0000	10.34	<.0000	.0545	2.878	<.0000
#3	<.0000	10.27	<.0000	.0454	2.891	<.0000
#4	<.0000	10.29	<.0000	.0439	2.859	<.0000

Analysis Report

04/02/02 11:28:16 AM

page 1

Method: NYCW Sample Name: 4415

Run Time: 04/02/02 11:26:33

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0056	6.534	<.0000	.0499	1.453	<.0000
SDev	.0168	.032	.0025	.0029	.008	.0047
%RSD	297.1	.4930	113.3	5.839	.5504	39.66
#1	.0190	6.577	<.0000	.0476	1.461	<.0000
#2	.0176	6.530	.0002	.0507	1.454	<.0000
#3	<.0000	6.531	<.0000	.0476	1.454	<.0000
#4	.0031	6.499	<.0000	.0537	1.442	<.0000

Analysis Report

04/02/02 11:32:33 AM

page 1

Method: NYCW Sample Name: ipc
Run Time: 04/02/02 11:30:50
Comment:
Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	1.028	.9947	.9915	1.026	.9874	.9534
SDev	.020	.0075	.0047	.004	.0082	.0075
%RSD	1.976	.7507	.4724	.3529	.8311	.7851
#1	1.018	.9892	.9955	1.029	.9842	.9487
#2	1.011	.9877	.9849	1.027	.9906	.9453
#3	1.027	.9985	.9940	1.029	.9779	.9606
#4	1.057	1.003	.9916	1.021	.9969	.9589

Analysis Report

04/02/02 11:36:05 AM

page 1

Method: NYCW Sample Name: cb

Run Time: 04/02/02 11:34:22

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	<.0000	<.0000	<.0000	<.0000	.0080
SDev	.0268	.0043	.0028	.0042	.0270	.0046
%RSD	412.8	57.46	250.7	368.4	188.9	56.77
#1	<.0000	<.0000	<.0000	.0030	<.0000	.0097
#2	<.0000	<.0000	.0018	<.0000	.0119	.0097
#3	.0227	<.0000	.0006	.0015	<.0000	.0114
#4	<.0000	<.0000	<.0000	<.0000	<.0000	.0013



Analysis Report

04/02/02 12:51:52 PM

page 1

Method: NYCW Sample Name: 4416
Run Time: 04/02/02 12:50:07
Comment:
Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0169	6.532	.0005	.0393	1.452	<.0000
SDev	.0279	.020	.0034	.0032	.035	.0043
%RSD	165.6	.3126	700.5	8.032	2.439	313.8
#1	.0468	6.559	.0050	.0400	1.485	.0035
#2	.0064	6.535	.0011	.0431	1.479	<.0000
#3	.0309	6.520	<.0000	.0385	1.415	.0009
#4	<.0000	6.513	<.0000	.0355	1.428	<.0000

Analysis Report

04/02/02 12:59:57 PM

page 1

Method: NYCW Sample Name: 4416lab dup

Operator: it

Run Time: 04/02/02 12:55:33

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0212	6.548	<.0000	.0374	1.472	<.0000
SDev	.0137	.011	.0020	.0036	.017	.0035
%RSD	64.79	.1614	362.1	9.609	1.150	298.8
#1	.0281	6.560	<.0000	.0400	1.453	.0009
#2	.0064	6.552	.0015	.0324	1.463	.0026
#3	.0136	6.536	<.0000	.0370	1.492	<.0000
#4	.0367	6.542	.0006	.0400	1.479	<.0000

CDL

37.039
ok

Analysis Report

04/02/02 01:06:09 PM

page 1

Method: NYCW Sample Name: 4416spk

Operator: it

Run Time: 04/02/02 13:04:18

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.2823	6.470	.1031	.2966	1.450	.0960
SDev	.0107	.025	.0019	.0047	.012	.0065
%RSD	3.784	.3861	1.834	1.586	.8597	6.767
#1	.2837	6.468	.1056	.2924	1.434	.0930
#2	.2808	6.505	.1011	.2970	1.456	.0921
#3	.2953	6.460	.1030	.3031	1.463	.1057
#4	.2693	6.447	.1028	.2940	1.447	.0930

106

103

103

966

Analysis Report

04/02/02 01:08:45 PM

page 1

Method: NYCW Sample Name: 4418

Operator: it

Run Time: 04/02/02 13:07:01

Comment:

Mode: CONC Corr. Factor: 1



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0197	6.491	<.0000	.0412	1.446	.0003
SDev	.0072	.025	.0023	.0045	.022	.0025
%RSD	36.52	.3794	191.2	11.01	1.529	804.5
#1	.0295	6.467	.0022	.0401	1.457	.0009
#2	.0179	6.503	<.0000	.0400	1.466	.0009
#3	.0194	6.520	<.0000	.0477	1.447	.0026
#4	.0122	6.475	<.0000	.0370	1.415	<.0000

Analysis Report

04/02/02 01:11:16 PM

page 1

Method: NYCW Sample Name: 4419

Run Time: 04/02/02 13:09:33

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0454	6.623	<.0000	.0385	1.503	<.0000
SDev	.0068	.021	.0014	.0033	.007	.0046
%RSD	14.93	.3167	359.2	8.526	.4365	388.8
#1	.0512	6.645	<.0000	.0355	1.498	<.0000
#2	.0512	6.620	.0002	.0431	1.511	.0043
#3	.0410	6.634	<.0000	.0385	1.507	.0001
#4	.0382	6.596	.0013	.0370	1.498	<.0000

Analysis Report

04/02/02 01:32:44 PM

page 1

Method: NYCW Sample Name: 4449

Operator: it

Run Time: 04/02/02 13:31:01

Comment:

Mode: CONC Corr. Factor: 1



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0486	6.642	.0037	.0381	1.509	.0005
SDev	.0183	.032	.0013	.0053	.013	.0021
%RSD	37.69	.4801	36.57	13.96	.8870	419.8
#1	.0382	6.603	.0025	.0355	1.514	<.0000
#2	.0439	6.637	.0053	.0324	1.507	.0001
#3	.0757	6.648	.0042	.0446	1.492	.0035
#4	.0367	6.680	.0026	.0400	1.523	.0001

Analysis Report

04/02/02 01:35:14 PM

page 1

Method: NYCW Sample Name: 4450

Run Time: 04/02/02 13:33:31

Comment:

Mode: CONC Corr. Factor: 1



Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0371	6.596	.0002	.0374	1.506	.0016
SDev	.0303	.033	.0019	.0019	.006	.0040
%RSD	81.83	.4942	1102.	5.117	.4046	257.0
#1	<.0000	6.555	<.0000	.0370	1.501	<.0000
#2	.0656	6.614	.0008	.0370	1.514	.0051
#3	.0555	6.629	.0022	.0400	1.501	.0035
#4	.0295	6.587	<.0000	.0355	1.507	.0018

Analysis Report

04/02/02 01:40:04 PM

page 1.

Method: NYCW Sample Name: 4509

Operator: it

Run Time: 04/02/02 13:38:21

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0244	6.552	.0004	.0351	1.485	<.0000
SDev	.0255	.046	.0005	.0034	.027	.0036
%RSD	104.2	.7080	132.5	9.614	1.793	670.9
#1	.0122	6.529	.0003	.0340	1.482	.0010
#2	<.0000	6.547	.0002	.0324	1.463	<.0000
#3	.0497	6.618	<.0000	.0339	1.523	.0035
#4	.0411	6.514	.0011	.0400	1.472	<.0000

Analysis Report

04/02/02 01:43:37 PM

page 1

Method: NYCW Sample Name: 4510

Operator: it

Run Time: 04/02/02 13:41:53

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0342	24.64	.0022	.0535	8.299	.0047
SDev	.0220	.06	.0016	.0023	.049	.0048
%RSD	64.50	.2613	74.71	4.267	.5879	102.4
#1	.0298	24.65	.0022	.0562	8.309	.0005
#2	.0371	24.57	.0022	.0547	8.268	.0022
#3	.0616	24.72	.0042	.0516	8.363	.0047
#4	.0082	24.60	.0002	.0517	8.255	.0115

Analysis Report

04/02/02 01:45:49 PM

page 1

Method: NYCW Sample Name: 4512

Run Time: 04/02/02 13:44:06

Comment:

Mode: CONC Corr. Factor: 1

✓ Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0414	6.591	.0006	.0347	1.495	.0009
SDev	.0103	.027	.0022	.0029	.026	.0043
%RSD	24.89	.4092	390.5	8.397	1.765	455.8
#1	.0425	6.580	<.0000	.0324	1.517	<.0000
#2	.0353	6.627	.0013	.0324	1.504	.0001
#3	.0324	6.564	<.0000	.0355	1.457	.0001
#4	.0555	6.594	.0032	.0385	1.501	.0068

Analysis Report

04/02/02 01:55:02 PM

page 1

Method: NYCW Sample Name: 4513

Run Time: 04/02/02 13:53:16

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0071	6.533	<.0000	.0366	1.490	.0026
SDev	.0184	.062	.0028	.0036	.012	.0014
%RSD	258.7	.9424	192.4	9.814	.7913	52.63
#1	<.0000	6.566	<.0000	.0415	1.475	.0043
#2	.0075	6.560	<.0000	.0339	1.498	.0026
#3	.0046	6.441	<.0000	.0339	1.501	.0026
#4	.0306	6.567	.0019	.0370	1.485	.0009

Analysis Report

04/02/02 01:57:44 PM

page 1

Method: NYCW Sample Name: 4514

Run Time: 04/02/02 13:55:58

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	24.43	<.0000	.0543	8.220	.0035
SDev	.0089	.03	.0011	.0042	.033	.0016
%RSD	81.31	.1228	119.3	7.703	.3983	46.37
#1	<.0000	24.47	<.0000	.0592	8.225	.0022
#2	<.0000	24.43	.0007	.0501	8.257	.0056
#3	<.0000	24.42	<.0000	.0562	8.219	.0022
#4	<.0000	24.40	<.0000	.0517	8.177	.0039

Analysis Report

04/02/02 02:01:38 PM


page 1

Method: NYCW Sample Name: ipc
Run Time: 04/02/02 13:59:53
Comment:
Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	1.021	.9848	.9900	1.027	.9722	.9710
SDev	.014	.0031	.0054	.008	.0213	.0072
%RSD	1.394	.3184	.5501	.7547	2.191	.7414
#1	1.017	.9838	.9837	1.016	.9818	.9655
#2	1.042	.9889	.9968	1.034	.9818	.9680
#3	1.019	.9851	.9885	1.030	.9404	.9689
#4	1.008	.9814	.9910	1.027	.9850	.9816



Analysis Report

04/02/02 02:07:11 PM

page 1

Method: NYCW Sample Name: cb

Run Time: 04/02/02 14:05:26

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0127	<.0000	.0003	<.0000	<.0000	.0180
SDev	.0055	.0018	.0022	.0049	.0074	.0017
%RSD	43.74	8.770	803.2	1290.	132.0	9.397
#1	.0141	<.0000	.0021	.0015	.0008	.0171
#2	.0083	<.0000	.0017	.0015	<.0000	.0171
#3	.0083	<.0000	<.0000	<.0000	<.0000	.0171
#4	.0199	<.0000	.0002	.0030	.0008	.0205

