

NYC Samples.

4/9/02

Blk

5516 ✓

2 STP

IPC+CB

IPC+CB

~~Blk~~

Reviewed by,

LFB+LRB

5517 ICB+ICB

9/5 7/8/02

3617 ✓

IPC+CB

5232 ✓ (5232 lob Dub + Spk)

5233 ✓

xip Blk

103.3

5236 ✓

Spk

 3.993×50

= 1976.7

5237 ✓

4.087 1010

104.1

5239 ✓

Dup

 3.993×50

= 1993.6

5240 ✓

8135 $5.951 \times 50 = 297.6$

5241 ✓

8136 $9.425 \times 50 = 147.4$

5301 ✓

.3333

5302 ✓

8137 $3.12 \times 50 = 327.1$

IPC+CB ✓

5303 + Lab Dup + Spk

8138 $25.75 \times 50 = 2686.8$

5304 ✓

8139 $2.951 \times 50 = 252.96$

5305 ✓

8140 $19.42 \times 50 = 1783.6$

5306 ✓

8141 $4.86 \times 50 = 277.7$

5307 ✓

5512 ✓

8142 $3.374 \times 50 = 303.6$

5513 ✓

8143 $2.757 \times 50 = 137.85$

5514 ✓

8144 $.2083 \times 50 = 10.4$

5515 ✓

Standardization Rpt.

04/09/02 08:48:50 AM

page 1

Method: NYCW Standard: STD1-Blank

Run Time: 04/09/02 08:47:08



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	6.252	5.996	7.656	.0025	2.352	1.172
SDev	.016	.016	.024	.0030	.010	.006
%RSD	.2483	.2737	.3071	120.0	.4302	.5024
#1	6.272	6.018	7.646	.0040	2.354	1.178
#2	6.244	5.998	7.630	.0060	2.340	1.166
#3	6.236	5.988	7.666	.0000	2.348	1.168
#4	6.254	5.980	7.684	.0000	2.364	1.176

Standardization Rpt.

04/09/02 08:54:09 AM

page 1

Method: NYCW Standard: STD3

Run Time: 04/09/02 08:52:28

Elem	Al3082	Fe2599	Mo2020
Avge	29.50	24.62	47.42
SDev	.11	.09	.16
%RSD	.3791	.3753	.3434


#1	29.57	24.71	47.55
#2	29.43	24.57	47.32
#3	29.38	24.52	47.23
#4	29.61	24.69	47.56

Standardization Rpt.

04/09/02 08:56:50 AM

page 1

Method: NYCW Standard: STD2
Run Time: 04/09/02 08:55:28



Elem	Ca3179	Co2286	Mg2790
Avge	129.2	324.0	12.27
SDev	.2	.6	.02
%RSD	.1693	.1874	.2016

#1	129.1	323.6	12.24
#2	129.1	323.7	12.25
#3	129.1	323.8	12.27
#4	129.5	324.9	12.30

Analysis Report

04/09/02 08:59:45 AM

page 1

Method: NYCW Sample Name: ipc

Operator: it

Run Time: 04/09/02 08:58:02

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.9831	1.014	.9919	.9918	.9988	.9821
SDev	.0032	.004	.0030	.0051	.0057	.0066
%RSD	.3305	.4264	.3010	.5115	.5713	.6693
#1	.9790	1.020	.9962	.9865	.9988	.9817
#2	.9842	1.009	.9918	.9906	.9947	.9800
#3	.9868	1.013	.9896	.9914	.9947	.9756
#4	.9825	1.013	.9900	.9987	1.007	.9912

Analysis Report

04/09/02 09:01:52 AM

page 1

Method: NYCW Sample Name: cb

Operator: it

Run Time: 04/09/02 09:00:09

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	<.0000	<.0000	.0018	<.0000	.0067
SDev	.0098	.0015	.0013	.0014	.0118	.0026
%RSD	39.45	34.50	60.26	76.66	45.02	38.77
#1	<.0000	<.0000	<.0000	.0031	<.0000	.0104
#2	<.0000	<.0000	<.0000	.0022	<.0000	.0065
#3	<.0000	<.0000	<.0000	.0022	<.0000	.0056
#4	<.0000	<.0000	<.0000	<.0000	<.0000	.0043

Analysis Report

04/09/02 09:03:48 AM

page 1

Method: NYCW Sample Name: lfb

Operator: it

Run Time: 04/09/02 09:02:05

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.2235	.0121	.0985	.2504	<.0000	.0985
SDev	.0108	.0011	.0012	.0016	.0106	.0022
%RSD	4.816	8.914	1.258	.6215	46.58	2.206
#1	.2230	.0120	.0976	.2492	<.0000	.0981
#2	.2316	.0136	.1004	.2508	<.0000	.1015
#3	.2308	.0118	.0982	.2492	<.0000	.0981
#4	.2084	.0110	.0980	.2525	<.0000	.0964

gg

qo

150

qg

Analysis Report

04/09/02 09:09:46 AM

page 1

Method: NYCW Sample Name: lrb

Operator: it

Run Time: 04/09/02 09:08:03

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	<.0000	<.0000	.0008	<.0000	<.0000
SDev	.0073	.0031	.0011	.0024	.0126	.0020
%RSD	39.16	98.17	197.0	296.8	119.0	99.64
#1	<.0000	<.0000	<.0000	.0014	<.0000	<.0000
#2	<.0000	.0003	.0011	.0030	.0025	<.0000
#3	<.0000	<.0000	<.0000	<.0000	<.0000	<.0000
#4	<.0000	<.0000	<.0000	.0014	<.0000	.0009

Analysis Report

04/09/02 09:15:46 AM

page 1

Method: NYCW Sample Name: 3617

Run Time: 04/09/02 09:14:03

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.846	<.0000	.0288	1.433	<.0000
SDev	.0103	.025	.0008	.0018	.013	.0005
%RSD	57.91	.3688	46.72	6.090	.8926	12.66
#1	<.0000	6.829	<.0000	.0264	1.434	<.0000
#2	<.0000	6.868	<.0000	.0296	1.446	<.0000
#3	<.0000	6.867	<.0000	.0288	1.438	<.0000
#4	<.0000	6.820	<.0000	.0305	1.416	<.0000

Analysis Report

04/09/02 09:19:50 AM

page 1

Method: NYCW Sample Name: 5232

Operator: it

Run Time: 04/09/02 09:18:07

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.566	<.0000	.0305	1.457	<.0000
SDev	.0056	.016	.0015	.0018	.016	.0015
%RSD	22.99	.2468	42.39	5.771	1.081	32.78
#1	<.0000	6.563	<.0000	.0297	1.442	<.0000
#2	<.0000	6.552	<.0000	.0329	1.446	<.0000
#3	<.0000	6.559	<.0000	.0305	1.462	<.0000
#4	<.0000	6.589	<.0000	.0288	1.476	<.0000

Analysis Report

04/09/02 09:21:53 AM

page 1

Method: NYCW Sample Name: 5232lab dup Operator: it
Run Time: 04/09/02 09:20:10
Comment:
Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.550	<.0000	.0293	1.467	<.0000
SDev	.0099	.029	.0023	.0027	.009	.0013
%RSD	46.01	.4352	86.82	9.196	.6096	30.47
#1	<.0000	6.567	<.0000	.0321	1.480	<.0000
#2	<.0000	6.573	<.0000	.0297	1.464	<.0000
#3	<.0000	6.510	<.0000	.0256	1.464	<.0000
#4	<.0000	6.549	<.0000	.0297	1.460	<.0000

29³⁰
9²⁰
Juf

Analysis Report

04/09/02 09:43:25 AM

page 1


Method: NYCW Sample Name: 5232spk

Operator: it

Run Time: 04/09/02 09:41:42

Comment:

Mode: CONC Corr. Factor: 1



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.2435	6.509	.1010	.2848	1.459	.0929
SDev	.0092	.012	.0027	.0016	.008	.0031
%RSD	3.789	.1801	2.690	.5708	.5221	3.384
#1	.2540	6.524	.1038	.2840	1.460	.0958
#2	.2316	6.510	.0977	.2840	1.448	.0898
#3	.2428	6.508	.0999	.2840	1.466	.0906
#4	.2454	6.496	.1026	.2872	1.460	.0954

977

1017

1027

936

Analysis Report

04/09/02 09:46:34 AM

page 1

Method: NYCW Sample Name: 5233

Run Time: 04/09/02 09:44:51

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	24.73	<.0000	.0579	8.349	<.0000
SDev	.0027	.17	.0008	.0027	.052	.0014
%RSD	11.25	.7039	33.07	4.587	.6181	101.1
#1	<.0000	24.71	<.0000	.0561	8.342	<.0000
#2	<.0000	24.98	<.0000	.0617	8.418	.0003
#3	<.0000	24.62	<.0000	.0577	8.344	<.0000
#4	<.0000	24.60	<.0000	.0561	8.293	<.0000

Analysis Report

04/09/02 09:48:56 AM

page 1

Method: NYCW Sample Name: 5236

Run Time: 04/09/02 09:47:13

Operator: it

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.540	<.0000	.0339	1.460	<.0000
SDev	.0084	.028	.0008	.0014	.013	.0022
%RSD	60.98	.4324	27.56	4.081	.9192	57.52
#1	<.0000	6.507	<.0000	.0321	1.444	<.0000
#2	<.0000	6.531	<.0000	.0345	1.462	<.0000
#3	<.0000	6.574	<.0000	.0337	1.476	<.0000
#4	<.0000	6.549	<.0000	.0354	1.456	<.0000

Analysis Report

04/09/02 09:51:16 AM

page 1

Method: NYCW Sample Name: 5237

Run Time: 04/09/02 09:49:33

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	24.52	<.0000	.0541	8.261	<.0000
SDev	.0062	.10	.0013	.0008	.028	.0017
%RSD	22.37	.4002	44.51	1.498	.3389	39.92
#1	<.0000	24.63	<.0000	.0545	8.301	<.0000
#2	<.0000	24.54	<.0000	.0545	8.251	<.0000
#3	<.0000	24.50	<.0000	.0529	8.255	<.0000
#4	<.0000	24.39	<.0000	.0545	8.237	<.0000

Analysis Report

04/09/02 09:53:37 AM

page 1

Method: NYCW Sample Name: 5239
Run Time: 04/09/02 09:51:54
Comment:
Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.557	<.0000	.0311	1.471	<.0000
SDev	.0103	.020	.0011	.0014	.010	.0014
%RSD	108.8	.3114	48.90	4.458	.6854	43.85
#1	.0054	6.563	<.0000	.0329	1.474	<.0000
#2	<.0000	6.566	<.0000	.0313	1.482	<.0000
#3	<.0000	6.571	<.0000	.0297	1.470	<.0000
#4	<.0000	6.526	<.0000	.0305	1.458	<.0000

Analysis Report

04/09/02 09:57:20 AM

page 1

Method: NYCW Sample Name: 5240

Run Time: 04/09/02 09:55:37

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	24.86	<.0000	.0569	8.397	<.0000
SDev	.0123	.15	.0006	.0015	.040	.0022
%RSD	57.23	.6049	25.86	2.579	.4756	47.56
#1	<.0000	25.00	<.0000	.0585	8.441	<.0000
#2	<.0000	24.82	<.0000	.0577	8.380	<.0000
#3	<.0000	24.66	<.0000	.0553	8.350	<.0000
#4	<.0000	24.94	<.0000	.0561	8.416	<.0000

Analysis Report

04/09/02 10:00:05 AM

page 1

Method: NYCW Sample Name: 5241

Run Time: 04/09/02 09:58:22

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.637	<.0000	.0347	1.498	<.0000
SDev	.0085	.027	.0012	.0033	.009	.0013
%RSD	119.8	.4057	68.98	9.430	.6159	29.76
#1	<.0000	6.619	<.0000	.0321	1.484	<.0000
#2	<.0000	6.674	<.0000	.0345	1.501	<.0000
#3	.0019	6.638	<.0000	.0329	1.505	<.0000
#4	<.0000	6.616	<.0000	.0394	1.503	<.0000

Analysis Report

04/09/02 10:02:29 AM

page 1

Method: NYCW Sample Name: 5301

Run Time: 04/09/02 10:00:46

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.685	<.0000	.0392	1.510	<.0000
SDev	.0044	.026	.0008	.0010	.010	.0031
%RSD	60.18	.3861	24.59	2.603	.6486	65.64
#1	<.0000	6.665	<.0000	.0394	1.505	<.0000
#2	<.0000	6.696	<.0000	.0378	1.505	<.0000
#3	<.0000	6.716	<.0000	.0402	1.525	<.0000
#4	<.0000	6.662	<.0000	.0394	1.507	<.0000

Analysis Report

04/09/02 10:05:03 AM

page 1

Method: NYCW Sample Name: 5302

Run Time: 04/09/02 10:03:21

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0024	6.646	<.0000	.0380	1.501	<.0000
SDev	.0131	.023	.0007	.0028	.007	.0013
%RSD	555.0	.3507	31.10	7.276	.4440	29.58
#1	<.0000	6.621	<.0000	.0370	1.497	<.0000
#2	<.0000	6.673	<.0000	.0402	1.505	<.0000
#3	.0174	6.633	<.0000	.0402	1.509	<.0000
#4	.0080	6.656	<.0000	.0345	1.495	<.0000

Analysis Report

04/09/02 10:07:30 AM

page 1

Method: NYCW Sample Name: ipc

Run Time: 04/09/02 10:05:47

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.9775	1.014	.9986	.9983	.9720	.9803
SDev	.0050	.005	.0086	.0108	.0112	.0154
%RSD	.5149	.4580	.8630	1.082	1.154	1.566
#1	.9722	1.009	.9863	.9825	.9584	.9579
#2	.9799	1.020	1.005	1.005	.9766	.9830
#3	.9833	1.015	1.004	1.000	.9846	.9886
#4	.9747	1.014	.9989	1.005	.9685	.9917



Analysis Report

04/09/02 10:09:28 AM

page 1

Method: NYCW Sample Name: cb
Run Time: 04/09/02 10:07:44
Comment:
Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	<.0000	<.0000	.0024	<.0000	.0111
SDev	.0044	.0021	.0007	.0015	.0134	.0030
%RSD	13.85	30.91	46.38	62.83	42.92	27.40
#1	<.0000	<.0000	<.0000	.0022	<.0000	.0147
#2	<.0000	<.0000	<.0000	.0047	<.0000	.0121
#3	<.0000	<.0000	<.0000	.0014	<.0000	.0078
#4	<.0000	<.0000	<.0000	.0014	<.0000	.0095



Analysis Report

04/09/02 10:12:14 AM

page 1

Method: NYCW Sample Name: 5303

Operator: it

Run Time: 04/09/02 10:10:31

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	25.21	<.0000	.0637	8.524	<.0000
SDev	.0065	.10	.0006	.0019	.037	.0005
%RSD	24.94	.4023	16.27	3.062	.4339	461.5
#1	<.0000	25.19	<.0000	.0625	8.541	.0003
#2	<.0000	25.32	<.0000	.0617	8.568	<.0000
#3	<.0000	25.25	<.0000	.0649	8.501	.0003
#4	<.0000	25.08	<.0000	.0658	8.487	<.0000

Analysis Report

04/09/02 10:18:42 AM

page 1

Method: NYCW Sample Name: 5303lab dup

Operator: it

Run Time: 04/09/02 10:16:59

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	25.09	<.0000	.0674	8.475	<.0000
SDev	.0080	.11	.0012	.0011	.029	.0026
%RSD	27.13	.4577	32.88	1.694	.3383	51.38
#1	<.0000	25.06	<.0000	.0682	8.463	<.0000
#2	<.0000	24.97	<.0000	.0658	8.485	<.0000
#3	<.0000	25.24	<.0000	.0682	8.509	<.0000
#4	<.0000	25.10	<.0000	.0674	8.443	<.0000

3.7
64.5
670

Analysis Report

04/09/02 10:24:01 AM

page 1

Method: NYCW Sample Name: 5303spk

Operator: it

Run Time: 04/09/02 10:22:18

Comment:

Mode: CONC Corr. Factor: 1



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.2293	24.79	.0999	.3234	8.368	.0948
SDev	.0070	.11	.0014	.0110	.031	.0021
%RSD	3.060	.4444	1.376	3.392	.3741	2.216
#1	.2385	24.66	.0987	.3161	8.358	.0928
#2	.2230	24.83	.1004	.3201	8.386	.0954
#3	.2247	24.74	.0990	.3396	8.330	.0936
#4	.2308	24.92	.1017	.3177	8.400	.0975

92%

100%

104%

95%

Analysis Report

04/09/02 10:26:08 AM

page 1

Method: NYCW Sample Name: 5304

Run Time: 04/09/02 10:24:25

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	7.730	<.0000	.0473	1.855	<.0000
SDev	.0101	.024	.0014	.0029	.006	.0022
%RSD	64.38	.3047	36.69	6.112	.3123	72.85
#1	<.0000	7.706	<.0000	.0449	1.854	<.0000
#2	<.0000	7.733	<.0000	.0514	1.852	<.0000
#3	<.0000	7.761	<.0000	.0457	1.864	<.0000
#4	<.0000	7.721	<.0000	.0473	1.852	<.0000

Analysis Report

04/09/02 10:28:23 AM

page 1

Method: NYCW Sample Name: 5305

Run Time: 04/09/02 10:26:40

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	23.56	<.0000	.0817	7.883	<.0000
SDev	.0106	.08	.0006	.0004	.027	.0019
%RSD	39.58	.3502	15.05	.4984	.3372	48.00
#1	<.0000	23.48	<.0000	.0815	7.858	<.0000
#2	<.0000	23.54	<.0000	.0815	7.874	<.0000
#3	<.0000	23.55	<.0000	.0823	7.880	<.0000
#4	<.0000	23.68	<.0000	.0815	7.920	<.0000

Analysis Report

04/09/02 10:31:27 AM

page 1

Method: NYCW Sample Name: 5306

Operator: it

Run Time: 04/09/02 10:29:44

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.821	<.0000	.0439	1.454	<.0000
SDev	.0057	.024	.0003	.0017	.014	.0015
%RSD	32.39	.3568	9.008	3.844	.9407	31.95
#1	<.0000	6.857	<.0000	.0459	1.474	<.0000
#2	<.0000	6.808	<.0000	.0418	1.448	<.0000
#3	<.0000	6.803	<.0000	.0435	1.444	<.0000
#4	<.0000	6.816	<.0000	.0443	1.450	<.0000

Analysis Report

04/09/02 10:33:28 AM

page 1

Method: NYCW Sample Name: 5307

Run Time: 04/09/02 10:31:45

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.693	<.0000	.0396	1.458	<.0000
SDev	.0068	.023	.0008	.0035	.008	.0019
%RSD	59.23	.3488	23.78	8.914	.5296	56.55
#1	<.0000	6.665	<.0000	.0402	1.456	<.0000
#2	<.0000	6.717	<.0000	.0410	1.464	<.0000
#3	<.0000	6.684	<.0000	.0345	1.448	<.0000
#4	<.0000	6.706	<.0000	.0427	1.464	<.0000

Analysis Report

04/09/02 10:35:42 AM

page 1

Method: NYCW Sample Name: 5512

Run Time: 04/09/02 10:33:59

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.548	<.0000	.0347	1.460	<.0000
SDev	.0131	.048	.0008	.0017	.013	.0013
%RSD	99.75	.7356	25.39	4.801	.8735	33.49
#1	.0002	6.510	<.0000	.0345	1.464	<.0000
#2	<.0000	6.590	<.0000	.0370	1.476	<.0000
#3	<.0000	6.589	<.0000	.0345	1.452	<.0000
#4	<.0000	6.503	<.0000	.0329	1.448	<.0000

Analysis Report

04/09/02 10:40:12 AM

page 1

Method: NYCW Sample Name: 5513

Run Time: 04/09/02 10:38:29

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.541	<.0000	.0333	1.457	<.0000
SDev	.0112	.033	.0017	.0023	.008	.0022
%RSD	51.91	.5110	50.68	7.027	.5708	50.86
#1	<.0000	6.550	<.0000	.0329	1.448	<.0000
#2	<.0000	6.523	<.0000	.0305	1.452	<.0000
#3	<.0000	6.508	<.0000	.0337	1.464	<.0000
#4	<.0000	6.584	<.0000	.0362	1.464	<.0000

Analysis Report

04/09/02 10:42:23 AM

page 1

Method: NYCW Sample Name: 5514

Run Time: 04/09/02 10:40:39

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	9.912	<.0000	.0412	2.656	<.0000
SDev	.0107	.050	.0018	.0021	.013	.0018
%RSD	73.09	.5088	48.49	5.078	.5075	63.84
#1	<.0000	9.895	<.0000	.0404	2.658	<.0000
#2	<.0000	9.853	<.0000	.0388	2.644	<.0000
#3	<.0000	9.973	<.0000	.0436	2.648	<.0000
#4	<.0000	9.926	<.0000	.0420	2.674	<.0000

Analysis Report

04/09/02 10:44:22 AM

page 1

Method: NYCW Sample Name: 5515

Run Time: 04/09/02 10:42:39

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	7.325	<.0000	.0383	1.733	<.0000
SDev	.0051	.015	.0013	.0018	.012	.0025
%RSD	78.35	.2000	31.14	4.702	.7199	79.48
#1	<.0000	7.324	<.0000	.0376	1.731	<.0000
#2	<.0000	7.332	<.0000	.0409	1.749	<.0000
#3	.0010	7.340	<.0000	.0368	1.735	<.0000
#4	<.0000	7.306	<.0000	.0376	1.718	<.0000

Analysis Report

04/09/02 10:46:47 AM

page 1

Method: NYCW Sample Name: 5516

Operator: it

Run Time: 04/09/02 10:45:04

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	11.67	<.0000	.0985	3.370	<.0000
SDev	.0095	.02	.0006	.0030	.009	.0012
%RSD	127.5	.1328	14.89	3.011	.2541	32.67
#1	<.0000	11.66	<.0000	.0953	3.382	<.0000
#2	<.0000	11.66	<.0000	.0969	3.368	<.0000
#3	.0021	11.69	<.0000	.1018	3.368	<.0000
#4	<.0000	11.67	<.0000	.1001	3.362	<.0000

Analysis Report

04/09/02 10:49:25 AM

page 1

Method: NYCW Sample Name: ipc

Run Time: 04/09/02 10:47:42

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.9786	1.016	1.002	1.003	.9750	.9927
SDev	.0162	.005	.007	.004	.0143	.0079
%RSD	1.654	.5393	.6610	.3853	1.467	.7954
#1	.9687	1.013	.9967	1.000	.9645	.9839
#2	.9808	1.017	1.002	1.005	.9927	.9973
#3	.9644	1.010	.9980	.9987	.9624	.9886
#4	1.001	1.023	1.011	1.007	.9806	1.001

Analysis Report

04/09/02 10:51:16 AM

page 1

Method: NYCW Sample Name: cb

Operator: it

Run Time: 04/09/02 10:49:33

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	<.0000	<.0000	.0004	<.0000	.0071
SDev	.0087	.0008	.0010	.0036	.0120	.0046
%RSD	28.18	10.69	42.41	857.2	32.73	65.39
#1	<.0000	<.0000	<.0000	.0039	<.0000	.0130
#2	<.0000	<.0000	<.0000	.0031	<.0000	.0082
#3	<.0000	<.0000	<.0000	<.0000	<.0000	.0026
#4	<.0000	<.0000	<.0000	<.0000	<.0000	.0044

