

6/5/02 IT

NYC samples

6/6/

BIK	12153	BIK
2 STD	IPC+CB	2 STD
IPC+CB	BIK	IPC
LFB+LRB	5 STD	LFB
11575 rerun	ICV+ICB	1215
11906 Lab Dup + s/k	Soi' BIK	1215
11907	LCSS F=201.5	12157
11926	11662 F=746.2	1215
11927	11662 Dup F=718.9	1216
11928	(10858 K check F=722.8)	1216
11929	QC 10ppm K	1216
11930	12330 F=670.4	1216
11931	12330 DF=618.9	1216
11932	12753 F=474.2	1216
IPC+CB	12754 F=282.4	IPC
12092+Lab Dup + s/k	12755 F=266.5	12168
12093	12756 F=321.2	1216
12094	12757 F=322	1227
12095	12757	1227
12096	12757 s/k	1228
12097	LCSS2 F=201.5	1228
12098	CCV+CCB	1228
12151		1228
12152		12286

Reviewed by,
 [Signature] 7/13/02

Standardization Rpt.

06/05/02 09:39:15 AM

page 1

Method: NYCW Standard: STD1-Blank

Run Time: 06/05/02 09:37:34

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	5.827	5.510	6.662	.0010	2.146	.9530
SDev	.009	.018	.026	.0011	.011	.0011
%RSD	.1548	.3265	.3975	115.5	.4931	.1212
#1	5.826	5.524	6.694	.0020	2.132	.9540
#2	5.828	5.526	6.634	.0000	2.144	.9540
#3	5.838	5.504	6.648	.0020	2.156	.9520
#4	5.816	5.488	6.672	.0000	2.152	.9520

Standardization Rpt.

06/05/02 09:53:58 AM

page 1

Method: NYCW Standard: STD3

Run Time: 06/05/02 09:52:16



Elem	Al3082	Fe2599	Mo2020
Avge	29.06	25.43	40.35
SDev	.11	.12	.31
%RSD	.3661	.4638	.7618

#1	28.95	25.36	40.20
#2	29.00	25.32	40.24
#3	29.08	25.45	40.16
#4	29.20	25.59	40.81

Standardization Rpt.

06/05/02 09:56:01 AM

page 1

Method: NYCW Standard: STD2

Run Time: 06/05/02 09:54:40

Elem	Ca3179	Co2286	Mg2790
Avge	135.1	313.4	11.78
SDev	.4	1.0	.03
%RSD	.2857	.3083	.2391

#1	135.5	314.0	11.82
#2	134.8	312.3	11.76
#3	134.8	312.9	11.75
#4	135.4	314.4	11.78

Analysis Report

06/05/02 10:02:25 AM

page 1

Method: NYCW Sample Name: ipc

Operator: it

Run Time: 06/05/02 10:00:38

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	1.046	1.045	1.026	1.034	1.049	.9743
SDev	.012	.002	.004	.003	.005	.0120
%RSD	1.185	.2067	.4058	.3197	.4978	1.236

#1	1.041	1.043	1.021	1.030	1.050	.9568
#2	1.032	1.044	1.026	1.038	1.050	.9778
#3	1.061	1.048	1.031	1.035	1.054	.9782
#4	1.049	1.044	1.026	1.034	1.041	.9844

✓

✓

✓

✓

Analysis Report

06/05/02 10:04:49 AM

page 1

Method: NYCW Sample Name: cb

Run Time: 06/05/02 10:03:02

Comment:


Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	<.0000	.0005	<.0000	<.0000	<.0000
SDev	.0008	.0007	.0012	.0013	.0079	.0054
%RSD	22.55	29.84	218.8	342.9	139.3	55.45

#1	<.0000	<.0000	.0009	<.0000	<.0000	<.0000
#2	<.0000	<.0000	<.0000	<.0000	.0005	<.0000
#3	<.0000	<.0000	.0020	.0014	.0005	<.0000
#4	<.0000	<.0000	<.0000	<.0000	<.0000	<.0000



Analysis Report

06/05/02 10:06:45 AM

page 1

Method: NYCW Sample Name: lfb

Operator: it

Run Time: 06/05/02 10:05:02

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.2731	.0213	.1050	.2812	.0087	.1011
SDev	.0019	.0019	.0008	.0059	.0139	.0028
%RSD	.6994	8.800	.7990	2.087	159.0	2.814
#1	.2716	.0226	.1045	.2743	.0212	.1023
#2	.2725	.0231	.1051	.2822	.0170	.1039
#3	.2759	.0191	.1043	.2798	.0066	.0972
#4	.2725	.0205	.1061	.2885	<.0000	.1008

109

105

112

101

Analysis Report

06/05/02 10:08:46 AM

page 1

Method: NYCW Sample Name: lrb

Operator: it

Run Time: 06/05/02 10:07:01

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	<.0000	<.0000	.0000	<.0000	<.0000
SDev	.0078	.0020	.0013	.0024	.0035	.0024
%RSD	606.7	59.91	90.96	86580.	42.69	6.459
#1	<.0000	<.0000	<.0000	.0014	<.0000	<.0000
#2	<.0000	<.0000	<.0000	<.0000	<.0000	<.0000
#3	.0073	<.0000	.0002	<.0000	<.0000	<.0000
#4	.0004	<.0000	<.0000	.0022	<.0000	<.0000

Analysis Report

06/05/02 10:21:31 AM

page 1

Method: NYCW Sample Name: 11575

Run Time: 06/05/02 10:19:47

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0078	6.415	<.0000	.0256	1.461	<.0000
SDev	.0065	.022	.0006	.0023	.015	.0018
%RSD	84.00	.3468	58.94	9.075	1.001	3.742
#1	.0125	6.441	<.0000	.0224	1.458	<.0000
#2	.0030	6.410	<.0000	.0263	1.464	<.0000
#3	.0013	6.388	<.0000	.0279	1.443	<.0000
#4	.0142	6.423	<.0000	.0256	1.479	<.0000

Analysis Report

06/05/02 10:24:06 AM

page 1

Method: NYCW Sample Name: 11906

Operator: it

Run Time: 06/05/02 10:22:08

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0390	6.590	.0019	.0328	1.516	<.0000
SDev	.0059	.026	.0015	.0026	.021	.0022
%RSD	15.12	.3984	78.15	7.901	1.386	4.833
#1	.0306	6.555	<.0000	.0295	1.485	<.0000
#2	.0392	6.585	.0031	.0326	1.522	<.0000
#3	.0435	6.611	.0019	.0334	1.531	<.0000
#4	.0426	6.609	.0027	.0357	1.526	<.0000

Analysis Report

06/05/02 10:26:23 AM

page 1

Method: NYCW Sample Name: 11906lab dup

Operator: it

Run Time: 06/05/02 10:24:34

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0140	6.524	<.0000	.0320	1.479	<.0000
SDev	.0091	.037	.0012	.0019	.021	.0013
%RSD	64.92	.5635	142.6	5.797	1.390	2.639
#1	.0168	6.475	.0001	.0318	1.452	<.0000
#2	.0013	6.517	<.0000	.0295	1.489	<.0000
#3	.0228	6.560	<.0000	.0334	1.499	<.0000
#4	.0151	6.542	<.0000	.0334	1.477	<.0000

✓

✓

good

✓

Analysis Report

06/05/02 10:36:08 AM

page 1

Method: NYCW Sample Name: 11906spk

Operator: it

Run Time: 06/05/02 10:31:48

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.2781	6.178	.1058	.2941	1.407	.0892
SDev	.0079	.031	.0017	.0017	.008	.0038
%RSD	2.835	.5079	1.601	.5771	.5941	4.303
#1	.2794	6.133	.1034	.2925	1.398	.0859
#2	.2880	6.185	.1075	.2965	1.412	.0920
#3	.2759	6.190	.1060	.2941	1.404	.0859
#4	.2690	6.205	.1061	.2933	1.416	.0930

94

104

104

89

Analysis Report

06/05/02 10:40:41 AM

page

Method: NYCW Sample Name: 11907

Operator: it

Run Time: 06/05/02 10:38:58

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.396	<.0000	.0234	1.462	<.0000
SDev	.0053	.036	.0006	.0018	.014	.0021
%RSD	95.77	.5600	48.83	7.778	.9830	24.01
#1	<.0000	6.342	<.0000	.0250	1.441	<.0000
#2	<.0000	6.414	<.0000	.0218	1.466	<.0000
#3	<.0000	6.417	<.0000	.0250	1.474	<.0000
#4	<.0000	6.411	<.0000	.0218	1.466	<.0000

Analysis Report

06/05/02 10:43:00 AM

page 1

Method: NYCW Sample Name: 11926

Run Time: 06/05/02 10:41:17

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0041	6.477	<.0000	.0265	1.479	<.0000
SDev	.0023	.047	.0005	.0013	.018	.0009
%RSD	55.28	.7248	34.27	4.816	1.235	8.170
#1	.0063	6.481	<.0000	.0265	1.470	<.0000
#2	.0037	6.532	<.0000	.0265	1.491	<.0000
#3	.0054	6.479	<.0000	.0281	1.497	<.0000
#4	.0011	6.417	<.0000	.0250	1.458	<.0000

Analysis Report

06/05/02 10:46:59 AM

page 1

Method: NYCW Sample Name: 11927

Operator: it

Run Time: 06/05/02 10:45:16

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0140	6.481	<.0000	.0269	1.486	<.0000
SDev	.0088	.020	.0011	.0008	.019	.0024
%RSD	62.66	.3146	84.59	2.942	1.293	25.23
#1	.0140	6.484	<.0000	.0281	1.458	<.0000
#2	.0106	6.461	<.0000	.0265	1.493	<.0000
#3	.0054	6.508	<.0000	.0265	1.491	<.0000
#4	.0261	6.471	.0000	.0265	1.501	<.0000

Analysis Report

06/05/02 10:53:53 AM

page 1

Method: NYCW Sample Name: 11928

Run Time: 06/05/02 10:52:10

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0073	6.472	<.0000	.0285	1.479	<.0000
SDev	.0122	.042	.0012	.0008	.011	.0013
%RSD	166.5	.6560	83.08	2.753	.7544	12.82
#1	.0097	6.455	<.0000	.0281	1.468	<.0000
#2	.0227	6.433	.0000	.0281	1.472	<.0000
#3	<.0000	6.466	<.0000	.0297	1.493	<.0000
#4	.0037	6.532	<.0000	.0281	1.483	<.0000

Analysis Report

06/05/02 10:55:45 AM

page 1

Method: NYCW Sample Name: 11929

Operator: it

Run Time: 06/05/02 10:54:02

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0014	8.980	<.0000	.0312	2.408	<.0000
SDev	.0030	.052	.0005	.0015	.010	.0011
%RSD	216.5	.5839	33.96	4.831	.3978	8.918
#1	.0051	8.956	<.0000	.0292	2.400	<.0000
#2	.0025	9.037	<.0000	.0323	2.416	<.0000
#3	<.0000	9.006	<.0000	.0308	2.416	<.0000
#4	<.0000	8.919	<.0000	.0323	2.400	<.0000

Analysis Report

06/05/02 10:58:00 AM

page 1

Method: NYCW Sample Name: 11930

Run Time: 06/05/02 10:56:18

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	11.22	<.0000	.0366	3.234	<.0000
SDev	.0038	.03	.0001	.0017	.004	.0013
%RSD	63.31	.2637	5.375	4.637	.1095	10.46
#1	<.0000	11.19	<.0000	.0366	3.229	<.0000
#2	<.0000	11.25	<.0000	.0374	3.233	<.0000
#3	<.0000	11.24	<.0000	.0382	3.238	<.0000
#4	<.0000	11.20	<.0000	.0343	3.235	<.0000

Analysis Report

06/05/02 11:00:40 AM

page 1

Method: NYCW Sample Name: 11931

Operator: it

Run Time: 06/05/02 10:58:57

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.0034	7.004	<.0000	.0283	1.397	<.0000
SDev	.0097	.015	.0009	.0016	.011	.0025
%RSD	282.9	.2147	65.75	5.739	.8094	21.91
#1	<.0000	6.996	<.0000	.0281	1.383	<.0000
#2	<.0000	7.003	<.0000	.0265	1.410	<.0000
#3	.0148	7.025	<.0000	.0281	1.393	<.0000
#4	.0071	6.992	<.0000	.0304	1.400	<.0000

Analysis Report

06/05/02 11:02:40 AM

page 1

Method: NYCW Sample Name: 11932

Run Time: 06/05/02 11:00:58

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	12.99	<.0000	.0340	4.761	<.0000
SDev	.0106	.04	.0008	.0015	.021	.0033
%RSD	265.1	.3129	42.08	4.436	.4450	27.79
#1	.0027	13.03	<.0000	.0344	4.785	<.0000
#2	.0044	13.02	<.0000	.0329	4.754	<.0000
#3	<.0000	12.97	<.0000	.0360	4.735	<.0000
#4	<.0000	12.95	<.0000	.0329	4.769	<.0000

Analysis Report

06/05/02 11:11:04 AM

page 1

Method: NYCW Sample Name: ipc
Run Time: 06/05/02 11:09:21
Comment:
Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	1.034	1.026	1.027	1.037	1.024	.9827
SDev	.005	.005	.008	.008	.012	.0123
%RSD	.5156	.4577	.8123	.7459	1.139	1.248

#1	1.029	1.027	1.020	1.030	1.012	.9682
#2	1.041	1.028	1.033	1.043	1.027	.9825
#3	1.031	1.019	1.019	1.030	1.018	.9820
#4	1.035	1.030	1.035	1.044	1.039	.9982

✓

✓

✓

✓

Analysis Report

06/05/02 11:13:03 AM

page 1

Method: NYCW Sample Name: cb

Run Time: 06/05/02 11:11:20

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	<.0000	.0004	<.0000	<.0000	.0178
SDev	.0070	.0013	.0009	.0016	.0021	.0067
%RSD	216.9	5.787	229.7	118.1	40.01	37.63
#1	.0028	<.0000	.0010	<.0000	<.0000	.0257
#2	<.0000	<.0000	.0010	.0000	<.0000	.0191
#3	.0028	<.0000	<.0000	.0000	<.0000	.0170
#4	<.0000	<.0000	.0004	<.0000	<.0000	.0094

✓

✓

✓

CSO

Analysis Report

06/05/02 11:15:19 AM

page 1

Method: NYCW Sample Name: 12092

Run Time: 06/05/02 11:13:36

Comment:

Mode: CONC Corr. Factor: 1

Operator: it



Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.435	<.0000	.0297	1.480	.0005
SDev	.0022	.027	.0009	.0013	.011	.0013
%RSD	38.47	.4126	31.05	4.312	.7537	247.8
#1	<.0000	6.447	<.0000	.0313	1.491	.0012
#2	<.0000	6.398	<.0000	.0281	1.466	.0017
#3	<.0000	6.459	<.0000	.0297	1.487	.0007
#4	<.0000	6.438	<.0000	.0297	1.476	<.0000

Analysis Report

06/05/02 11:17:45 AM

page 1

Method: NYCW Sample Name: 12092lab dup

Operator: it

Run Time: 06/05/02 11:16:02

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.449	<.0000	.0279	1.480	<.0000
SDev	.0073	.024	.0002	.0013	.018	.0028
%RSD	153.9	.3680	10.67	4.790	1.236	41.83
#1	<.0000	6.432	<.0000	.0273	1.458	<.0000
#2	<.0000	6.427	<.0000	.0265	1.474	<.0000
#3	<.0000	6.479	<.0000	.0297	1.501	<.0000
#4	.0011	6.457	<.0000	.0281	1.485	<.0000

✓

✓

69

✓

Analysis Report

06/05/02 11:32:29 AM

page 1

Method: NYCW Sample Name: 12092spk

Operator: it

Run Time: 06/05/02 11:30:38

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.2529	6.203	.0972	.2703	1.456	.0853
SDev	.0091	.013	.0015	.0030	.006	.0021
%RSD	3.601	.2029	1.541	1.098	.4209	2.478
#1	.2479	6.212	.0953	.2673	1.451	.0827
#2	.2444	6.190	.0974	.2697	1.449	.0858
#3	.2652	6.194	.0971	.2697	1.460	.0878
#4	.2540	6.215	.0990	.2744	1.462	.0848

101

97

96

84.8 ^{sk}

Analysis Report

06/05/02 11:38:23 AM

page 1

Method: NYCW Sample Name: 12093

Run Time: 06/05/02 11:36:40

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.414	<.0000	.0311	1.434	<.0000
SDev	.0048	.014	.0005	.0019	.011	.0010
%RSD	9.712	.2106	8.109	5.973	.7340	8.560
#1	<.0000	6.411	<.0000	.0297	1.422	<.0000
#2	<.0000	6.426	<.0000	.0297	1.437	<.0000
#3	<.0000	6.424	<.0000	.0336	1.447	<.0000
#4	<.0000	6.397	<.0000	.0313	1.431	<.0000

Analysis Report

06/05/02 11:40:21 AM

page 1

Method: NYCW Sample Name: 12094

Operator: it

Run Time: 06/05/02 11:38:39

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mn2020
Avge	<.0000	9.556	<.0000	.0484	2.550	<.0000
SDev	.0054	.057	.0007	.0019	.026	.0024
%RSD	13.12	.5993	9.443	3.869	1.024	20.36
#1	<.0000	9.515	<.0000	.0464	2.529	<.0000
#2	<.0000	9.500	<.0000	.0503	2.529	<.0000
#3	<.0000	9.593	<.0000	.0472	2.558	<.0000
#4	<.0000	9.616	<.0000	.0495	2.583	<.0000

Analysis Report

06/05/02 11:45:20 AM

page 1

Method: NYCW Sample Name: 12095

Run Time: 06/05/02 11:43:37

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	7.894	<.0000	.0404	1.912	<.0000
SDev	.0093	.027	.0016	.0009	.019	.0021
%RSD	26.99	.3418	29.27	2.230	1.011	17.54
#1	<.0000	7.858	<.0000	.0397	1.893	<.0000
#2	<.0000	7.924	<.0000	.0412	1.938	<.0000
#3	<.0000	7.895	<.0000	.0412	1.913	<.0000
#4	<.0000	7.898	<.0000	.0397	1.905	<.0000

Analysis Report

06/05/02 11:47:11 AM

page

Method: NYCW Sample Name: 12096

Operator: it

Run Time: 06/05/02 11:45:29

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.797	<.0000	.0354	1.489	<.0000
SDev	.0044	.038	.0006	.0034	.016	.0011
%RSD	10.66	.5627	10.57	9.485	1.095	8.981
#1	<.0000	6.755	<.0000	.0312	1.472	<.0000
#2	<.0000	6.781	<.0000	.0367	1.493	<.0000
#3	<.0000	6.845	<.0000	.0391	1.510	<.0000
#4	<.0000	6.808	<.0000	.0344	1.481	<.0000

Analysis Report

06/05/02 11:49:19 AM

page 1

Method: NYCW Sample Name: 12097

Run Time: 06/05/02 11:47:36

Comment:

Operator: it

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.369	<.0000	.0297	1.406	<.0000
SDev	.0065	.042	.0004	.0009	.019	.0018
%RSD	13.41	.6621	6.233	3.054	1.326	14.02
#1	<.0000	6.414	<.0000	.0305	1.402	<.0000
#2	<.0000	6.327	<.0000	.0289	1.383	<.0000
#3	<.0000	6.395	<.0000	.0289	1.427	<.0000
#4	<.0000	6.341	<.0000	.0305	1.414	<.0000

Analysis Report

06/05/02 11:51:19 AM

page 1

Method: NYCW Sample Name: 12098

Run Time: 06/05/02 11:49:36

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.681	<.0000	.0391	1.356	<.0000
SDev	.0120	.026	.0005	.0020	.007	.0016
%RSD	30.31	.3887	7.292	5.192	.5519	13.68
#1	<.0000	6.657	<.0000	.0415	1.358	<.0000
#2	<.0000	6.706	<.0000	.0399	1.349	<.0000
#3	<.0000	6.661	<.0000	.0368	1.366	<.0000
#4	<.0000	6.701	<.0000	.0384	1.352	<.0000

Analysis Report

06/05/02 11:53:09 AM

page 1

Method: NYCW Sample Name: 12151

Operator: it

Run Time: 06/05/02 11:51:26

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.351	<.0000	.0234	1.421	<.0000
SDev	.0081	.040	.0010	.0011	.021	.0011
%RSD	18.39	.6315	20.76	4.753	1.481	8.311
#1	<.0000	6.341	<.0000	.0250	1.418	<.0000
#2	<.0000	6.367	<.0000	.0234	1.429	<.0000
#3	<.0000	6.395	<.0000	.0226	1.443	<.0000
#4	<.0000	6.301	<.0000	.0226	1.393	<.0000

Analysis Report

06/05/02 11:55:29 AM

page 1

Method: NYCW Sample Name: 12152
Run Time: 06/05/02 11:53:46
Comment:
Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	6.367	<.0000	.0266	1.413	<.0000
SDev	.0087	.041	.0008	.0020	.011	.0019
%RSD	16.01	.6383	10.43	7.648	.8025	13.73
#1	<.0000	6.331	<.0000	.0289	1.427	<.0000
#2	<.0000	6.344	<.0000	.0258	1.399	<.0000
#3	<.0000	6.370	<.0000	.0273	1.416	<.0000
#4	<.0000	6.423	<.0000	.0242	1.410	<.0000

Analysis Report

06/05/02 11:57:24 AM

page 1

Method: NYCW Sample Name: 12153

Run Time: 06/05/02 11:55:41

Comment:

Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	26.09	<.0000	.0567	8.875	<.0000
SDev	.0086	.04	.0007	.0019	.031	.0022
%RSD	13.69	.1586	9.776	3.285	.3492	14.74
#1	<.0000	26.09	<.0000	.0580	8.903	<.0000
#2	<.0000	26.03	<.0000	.0580	8.836	<.0000
#3	<.0000	26.12	<.0000	.0564	8.897	<.0000
#4	<.0000	26.12	<.0000	.0541	8.863	<.0000

Analysis Report

06/05/02 12:02:20 PM

page 1

Method: NYCW Sample Name: ipc
Run Time: 06/05/02 12:00:38
Comment:
Mode: CONC Corr. Factor: 1

Operator: it

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	.9585	1.014	.9947	1.009	.9711	.9397
SDev	.0036	.008	.0036	.006	.0105	.0151
%RSD	.3720	.7556	.3631	.6195	1.079	1.607
#1	.9590	1.024	.9928	.9996	.9815	.9228
#2	.9590	1.015	.9914	1.011	.9732	.9315
#3	.9624	1.009	.9997	1.014	.9566	.9498
#4	.9538	1.006	.9950	1.010	.9732	.9548

Analysis Report

06/05/02 12:06:09 PM

page 1

Method: NYCW Sample Name: cb

Operator: it

Run Time: 06/05/02 12:04:26

Comment:

Mode: CONC Corr. Factor: 1

Elem	Al3082	Ca3179	Co2286	Fe2599	Mg2790	Mo2020
Avge	<.0000	<.0000	<.0000	<.0000	<.0000	<.0000
SDev	.0080	.0020	.0010	.0033	.0105	.0024
%RSD	12.42	8.860	17.55	139.9	19.94	64.89
#1	<.0000	<.0000	<.0000	<.0000	<.0000	<.0000
#2	<.0000	<.0000	<.0000	<.0000	<.0000	<.0000
#3	<.0000	<.0000	<.0000	<.0000	<.0000	<.0000
#4	<.0000	<.0000	<.0000	.0016	<.0000	<.0000