

DFTPP

Data File : C:\MSDCHEM\1\DATA\050202\DFTPA.D
Acq On : 2 May 2002 10:27 am
Sample :
Misc :
MS Integration Params: EVENTS.E

Vial: 1
Operator: Mclean
Inst : Instrumen
Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\050102.M (Chemstation Integrator)
Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3604

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	42.6	82904	PASS
68	69	0.00	2	1.6	1161	PASS
69	198	0.00	100	37.4	72840	PASS
70	69	0.00	2	0.5	385	PASS
127	198	40	60	48.6	94744	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	194752	PASS
199	198	5	9	6.1	11906	PASS
275	198	10	30	20.1	39056	PASS
365	198	1	99	2.5	4925	PASS
441	443	1	100	76.4	20560	PASS
442	198	40	110	73.1	142400	PASS
443	442	17	23	18.9	26912	PASS

DFTPA.D 050102.M

Thu May 02 11:17:48 2002