

DFTPP

Data File : C:\MSDCHEM\1\DATA\043002\DFTPPA.D
Acq On : 30 Apr 2002 11:46 am
Sample :
Misc :
MS Integration Params: EVENTS.E

Vial: 1
Operator: Mclean
Inst : Instrumen
Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\050102.M (Chemstation Integrator)
Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3604

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	44.5	154368	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	38.1	131968	PASS
70	69	0.00	2	0.6	833	PASS
127	198	40	60	51.4	178240	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	346560	PASS
199	198	5	9	6.4	22320	PASS
275	198	10	30	19.9	69128	PASS
365	198	1	99	2.6	9143	PASS
441	443	1	100	75.5	39560	PASS
442	198	40	110	78.0	270400	PASS
443	442	17	23	19.4	52400	PASS

DFTPPA.D 050102.M Thu May 02 09:52:31 2002

File : C:\MSDCHEM\1\DATA\043002\DFTPPA.D
 Operator : Mclean
 Acquired : 30 Apr 2002 11:46 am using AcqMethod 508SCAN
 Instrument : Instrumen
 Sample Name:
 Misc Info :
 Vial Number: 1

