

Method 508 mod.

Date Extracted 4-30-02

Page 91

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count		
LB		500ml		1m			1ml	8A	7			
Ref					1ml			7A				
Ref					↓			8B				
9093								7B				
10198								6B				
10199								6A				
10200								3A				
10201								9A				
10202								3B				
10203								4A				
10204								9B				
10205								10B				
10206								10A				
Remarks												
Reviewed by <u>JB 8/21/02</u>												
drip rate must be between 5-15 ml/min for liq-liq extractors												
drip rate -												
Extraction type - <u>Slp fwr</u> Solv. lot# <u>X06E27</u> NaCl lot# <u>V39614</u> Na2SO4 lot# <u>V46648</u>												
STANDARD	INT STD LOG#											
CALIBRATION												
REF SPIKE	<u>2279E</u>											
MDL/MDRF												
Q. CHECK												
INT. STD.												
SURROGATE	<u>GCMs</u>	<u>TCM X</u>	<u>4/25</u>									
LPC												
ENDRIN/DDT												

Ref ext
5/10/02
sup
my

Analyst SH
Revision 4/12/01

Verified by _____

Reviewed by DN
Date 5/6/02