

DFTPP

Data File : C:\MSDCHEM\1\DATA\050802\DFTPB.D
Acq On : 8 May 2002 10:45 pm
Sample :
Misc :
MS Integration Params: EVENTS.E

Vial: 1
Operator: Mclean
Inst : Instrumen
Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\050102.M (Chemstation Integrator)
Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3602

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	45.7	138688	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	38.8	117664	PASS
70	69	0.00	2	0.5	641	PASS
127	198	40	60	48.7	147776	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	303424	PASS
199	198	5	9	6.7	20240	PASS
275	198	10	30	19.2	58336	PASS
365	198	1	99	2.1	6356	PASS
441	443	1	100	78.5	29048	PASS
442	198	40	110	64.2	194880	PASS
443	442	17	23	19.0	37016	PASS

DFTPB.D 050102.M

Thu May 09 15:46:41 2002