

Method 508 modDate Extracted 5/03/02Page 100

Lab #	Description	Sample	Drip rate	Surr	Spk	IS	Final	T.V.	pH	Count		
		vol/wt	ml/min	add	add	add	vol.	pos.				
	UB	500ml		1ml			1ml	9A	7			
	Ref 1			1ml				9B				
	Ref 2			1ml				10A				
	10603							10B				
	10604							3A				
	10605							8B				
	10606							9A				
	10609							9B				
	10610							10A				
Remarks												
Reviewed by, <u>CB 8/21/02</u>												
drip rate must be between 5-15 ml/min for liq-liq extractors drip rate -												
Extraction type - <u>SEP-Fnd</u> Solv.lot# <u>X05E10</u> NaCl lot# <u>V39614</u> Na2SO4 lot# <u>V46648</u>												
STANDARD	INT STD LOG#											
CALIBRATION												
REF SPIKE	<u>2279E</u>											
MDL/MDRF												
Q. CHECK												
INT. STD.												
SURROGATE	<u>1ST 440 TCMX</u>											
LPC												
ENDRIN/DDT												

Analyst Feb
revision 4/12/01

Verified by _____

Reviewed by DW
Date 5/16/02

DFTPP

Data File : C:\MSDCHEM\1\DATA\050702\DFTPA.D
Acq On : 7 May 2002 1:04 pm
Sample :
Misc :
MS Integration Params: EVENTS.E

Vial: 1
Operator: Mclean
Inst : Instrumen
Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\050102.M (Chemstation Integrator)
Title : 508 Modified GC/MS scan

Spectrum Information: Scan 3603

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	46.7	114216	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	39.8	97296	PASS
70	69	0.00	2	0.4	390	PASS
127	198	40	60	51.3	125448	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	244736	PASS
199	198	5	9	7.1	17272	PASS
275	198	10	30	20.1	49168	PASS
365	198	1	99	2.8	6932	PASS
441	443	1	100	80.3	27352	PASS
442	198	40	110	72.9	178368	PASS
443	442	17	23	19.1	34048	PASS

DFTPA.D 050102.M

Thu May 09 09:19:33 2002

ion 209.00 (208.70 to 209.70): TCMX.D

1 20.540 BB 0.030 513098 20.480 20.617