

Method 508 mod.

Date Extracted 4-16-02

Page 61

Ref

\$ GCMS-18297

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count		
UB		500ml		1ml	1ml		1ml	8A	7			
Ref								1A				
Ref								6A				
8958 ✓								7B				
8959 ✓								1B				
8960 ✓								1B				
8961 ✓								2A				
8962 ✓								2B				
8963 ✓								4B				
8964 ✓								4A				
8965 ✓	FB							1A				
8969 ✓								7A				
8970 ✓								8B				
Remarks												
										Reviewed by, <u>CB 8/16/02</u>		
drip rate must be between 5-15 ml/min for liq-liq extractors drip rate - Extraction type <u>Sep Fur</u> Solv. lot# <u>X06E27</u> NaCl lot# <u>V39614</u> Na2SO4 lot# <u>V46648</u>												
STANDARD	INT STD LOG#											
CALIBRATION												
REF SPIKE	<u>2193</u>											
MDL/MDRF												
Q. CHECK												
INT. STD.												
SURROGATE	<u>TCMX</u>											
LPC												
ENDRIN/DDT												

Analyst EH.
revision 4/12/01

Verified by _____

Reviewed by DM
Date 3/7/2

Sequence Name: C:\MSDCHEM\1\SEQUENCE\020424.S

Comment: GCMS scan

Operator: Nefliu

Data Path: C:\MSDchem\1\DATA\020424\

Pre-Seq Cmd:

Post-Seq Cmd:

Method Sections To Run	On A Barcode Mismatch
(X) Full Method	(X) Inject Anyway
() Reprocessing Only	() Don't Inject

Line	Type	Vial	DataFile	Method	Sample Name
------	------	------	----------	--------	-------------

1	Sample	1	DFTPP	508FAST	dftpp
2	Sample	2	0424C3	508SCAN1	ccv
3	Sample	3	0416LB	508SCAN1	04/16 lab blk
4	Sample	4	0417LB	508SCAN1	04/17 lab blk
5	Sample	5	0416REF1	508SCAN1	04/16 ref
6	Sample	6	0416REF2	508SCAN1	04/16 ref
7	Sample	7	0417REF1	508SCAN1	04/17 ref
8	Sample	8	0417REF2	508SCAN1	04/17 ref
9	Sample	9	08958	508SCAN1	sample
10	Sample	10	08959	508SCAN1	sample
11	Sample	11	08960	508SCAN1	sample
12	Sample	12	08961	508SCAN1	sample
13	Sample	13	08962	508SCAN1	sample
14	Sample	14	08963	508SCAN1	sample
15	Sample	15	08964	508SCAN1	sample
16	Sample	16	08965	508SCAN1	sample
17	Sample	17	08969	508SCAN1	sample
18	Sample	18	08970	508SCAN1	sample
19	Sample	19	07171	508SCAN1	sample
20	Sample	20	09151	508SCAN1	sample
21	Sample	21	09152	508SCAN1	sample
22	Sample	22	09153	508SCAN1	sample
23	Sample	23	09154	508SCAN1	sample
24	Sample	24	09155	508SCAN1	sample
25	Sample	25	09156	508SCAN1	sample
26	Sample	26	09157	508SCAN1	sample

Method JAPGON2 Scan
Instrument 214

File Directory 020424

Storage Directory _____

Lab #	Filename	Description	Vial #		Inj.	Standards used - Lot #			Int. Std.	Remarks	Initials
			front	rear		Calibration	Reference	Surrogate			
	HTPP	HTPP	1		1						MM
	042403	HTPP	2			22876/2876/2860					
	041643	HTPP	3								
	041743	HTPP	4								
	041644	HTPP	5				2193				
	041645	HTPP	6								
	041744	HTPP	7								
	041745	HTPP	8								
	041746	sample	9								
	041747		10								
	041748		11								
	041749		12								
	041750		13								
	041751		14								
	041752		15								
	041753		16								
	041754		17								
	041755		18								
	041756		19								
	041757		20								
	041758		21								
	041759		22								
	041760		23								
	041761		24								
	041762		25								
	041763		26								

Date 04/24/02

Extraction date 04/16, 17/02

Page 5