NYC 9/11 Public Portal Document

Method 508 mod Date Extracted 1/14/02 Page 27

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk ref add	IS add	Final vol.	T.V. pos.	pH	Count		
LB		500mL		1mL			1mL	1A	7			
Ref		↓		↓	0.5mL		↓	4B	↓			
943.								4A				
944.								9A				
945.								2A				
947.								3B				
948.								3A				
949.		↓		↓			↓	5A	↓			

Remarks

Sample 00945 was lost due to the malfunction of turbo vap 2A sensor. 945 was re-extracted on 1/15/02.

drip rate must be between 5-15 ml/min for liq-liq extractors drip rate -

Extraction type - sepfun Solv.lot# V39466 NaCl lot# V11599 Na2SO4 lot# V39H00

STANDARD	INT STD LOG#
CALIBRATION	
REF/SPIKE	91759
MDL/MDRF	
Q. CHECK	
INT. STD.	
SURROGATE	010302
LPC	
ENDRIN/DDT	

Reviewed by, 5/18/02

Analyst LF
Revision 4/12/01

Verified by_____

Reviewed by
Date 2/28/02

NYC-WTC_000052331