

DFTPP

Data File : C:\MSDCHEM\1\DATA\022202\DFTPPA.D
 Acq On : 22 Feb 2002 10:35 am
 Sample : DFTPP
 Misc :
 MS Integration Params: BENZO.P

Vial: 1
 Operator: McLean
 Inst : Instrumen
 Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\5080202.M (RTE Integrator)
 Title : Quantitation For Method 508gcscan

Spectrum Information: Scan 2292

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	56.6	384512	PASS
68	69	0.00	2	1.5	4630	PASS
69	198	0.00	100	45.7	310528	PASS
70	69	0.00	2	0.6	1911	PASS
127	198	40	60	48.5	329408	PASS
197	198	0.00	2	0.6	4015	PASS
198	198	100	100	100.0	679680	PASS
199	198	5	9	6.3	42832	PASS
275	198	10	30	22.8	155008	PASS
365	198	1	100	2.9	19880	PASS
441	443	0.01	100	75.2	82944	PASS
442	198	40	110	87.8	596736	PASS
443	442	17	23	18.5	110328	PASS

DFTPPA.D 5080202.M Wed Feb 27 06:12:17 2002

File : C:\MSDCHEM\1\DATA\022202\DFTPPA.D
 Operator : McLean
 Acquired : 22 Feb 2002 10:35 am using AcqMethod 508SCAN
 Instrument : Instrumen
 Sample Name: DFTPP
 Misc Info :
 Vial Number: 1

