

Method 508 mol. Date Extracted 01-24-02 Page 45

REF

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	.Spk add	IS add	Final vol.	T.V. pos.	pH	Count
<u>LB</u>		<u>500 ml</u>		<u>1 ml</u>			<u>1 ml</u>	<u>7B</u>	<u>7</u>	
<u>REF</u>					<u>5 ml</u>			<u>7A</u>		
<u>1639</u>	<u>FB</u>							<u>1A</u>		
<u>1640</u>	<u>—</u>							<u>1B</u>		
<u>1641</u>								<u>6B</u>		
<u>1642</u>								<u>6A</u>		
<u>*1643</u>	<u>FB</u>							<u>7A</u>		
<u>1644</u>	<u>—</u>							<u>8B</u>		
<u>1645</u>	<u>—</u>							<u>1A</u>		
<u>1646</u>								<u>8A</u>		
<u>1647</u>	<u>FB</u>							<u>8B</u>		
<u>1648</u>								<u>2A</u>		
<u>1649</u>		<u>↓</u>		<u>↓</u>			<u>↓</u>	<u>5A</u>	<u>↓</u>	

Reviewed by,
JD 5/13/02

Remarks

* sample # 1643 dried out on Turbo Vap.

rest 39, 40, 44, 45, 47 low surrogate recoveries

drip rate must be between 5-15 ml/min for liq-liq extractors drip rate -

Extraction type Sep Fun Solv. lot# V41470 NaCl lot# V11599 Na2SO4 lot# V39400

STANDARD	INT STD LOG#
CALIBRATION	
REF SPIKE	<u>91759</u>
MDL/MDRF	
Q. CHECK	
INT. STD.	
SURROGATE	<u>Gcms 011602</u>
LPC	
ENDRIN/DDT	

Analyst EA revision 4/12/01 Verified by _____ Reviewed by DN Date 3/4/02

DFTPP

Data File : C:\MSDCHEM\1\DATA\022602\DFTPP.D
 Acq On : 26 Feb 2002 9:03 am
 Sample : DFTPP
 Misc :
 MS Integration Params: BENZO.P

Vial: 1
 Operator: McLean
 Inst : Instrumen
 Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\5080202.M (RTE Integrator)
 Title : Quantitation For Method 508gcscan

Spectrum Information: Scan 2291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	51.1	296448	PASS
68	69	0.00	2	1.6	3845	PASS
69	198	0.00	100	42.7	247936	PASS
70	69	0.00	2	0.5	1213	PASS
127	198	40	60	46.2	268416	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	580672	PASS
199	198	5	9	6.5	37720	PASS
275	198	10	30	23.3	135552	PASS
365	198	1	100	2.8	16264	PASS
441	443	0.01	100	76.2	75200	PASS
442	198	40	110	90.0	522624	PASS
443	442	17	23	18.9	98712	PASS

DFTPP.D 5080202.M Thu Feb 28 11:20:13 2002

File : C:\MSDCHEM\1\DATA\022602\DFTPP.D
 Operator : McLean
 Acquired : 26 Feb 2002 9:03 am using AcqMethod 508SCAN
 Instrument : Instrumen
 Sample Name: DFTPP
 Misc Info :
 Vial Number: 1

