

Daily Performance Report

Sample ID: DPC

Sample Date/Time: Friday, February 15, 2002 09:31:18

Sample Description: DAILY PERFORMANCE CHECK

Method File: d:\elandata\Method\Daily.mth

Dataset File: d:\elandata\Dataset\daily performance\DPC.603

Tuning File: d:\elandata\Tuning\default.tun

Optimization File: d:\elandata\Optimize\default.dac

Dual Detector Mode: Pulse

Acq. Dead Time(ns): 55

Current Dead Time (ns): 55

Summary

Analyte	Mass	Meas. Intens.	Mean	Net Intens.	Mean	Net Intens.	SD	Net Intens.	RSD
Mg	24.0		76346.3	720		76346.313	320	331.333	0.4
Rh	102.9		302243.6	710		302243.556	750	1231.700	0.4
In	114.9		402479.9	710		402479.941		948.969	0.2
Pb	208.0		275694.0	710		275693.986	750	874.867	0.3
[> Ba	137.9		346032.7			346032.688		1660.384	0.5
[Ba++	69.0		4734.2			0.014	<0.034	0.000	0.7
[> Ce	139.9		428205.5			428205.485		1805.448	0.4
[CeO	155.9		12100.7			0.028	<0.034	0.000	1.1
Bkgd	220.0		11.8	30		11.800		0.721	6.1

Current Optimization File Data

Current Value	Description
0.89	Nebulizer Gas Flow
6.25	Lens Voltage
1000.00	ICP RF Power
-1887.50	Analog Stage Voltage
1500.00	Pulse Stage Voltage
70.00	Discriminator Threshold
-6.00	AC Rod Offset
60.00	Service DAC 1
0.00	Quadrupole Rod Offset

Current Autolens Data

Analyte	Mass	Num of Pts	DAC Value	Maximum Intensity
Be	9	21	6.0	6293.2
Al	27	21	6.0	139942.3
Sc	45	21	6.3	770852.9
V	51	21	6.3	168904.3
Cr	52	21	6.3	154584.8
Mn	55	21	6.3	225315.2
Ni	58	21	6.3	95327.1
Co	59	21	6.5	164245.0
Cu	63	21	6.5	84487.6
Zn	64	21	6.5	46220.7
Cu	65	21	6.3	41494.9
Zn	66	21	6.5	27506.8
Zn	68	21	6.5	20076.3
Ge	74	21	6.5	182944.2
As	75	21	6.5	24776.0

Sample ID: DPC

Report Date/Time: Friday, February 15, 2002 09:34:44

Page 1

Se	82	21	6.8	3012.5
Mo	98	21	7.0	79113.5
Ag	107	21	7.0	137539.0
Cd	114	21	7.0	77414.0
In	115	21	6.8	1618244.8
Sb	121	21	6.8	103635.4
Sb	123	21	7.0	81363.3
Ba	136	21	7.0	41666.7
Ba	138	21	7.0	376002.9
Ho	165	13	8.8	1991806.4
Tl	205	21	8.0	404616.2
Pb	208	21	8.3	284917.7