

User: Fakhouri, Morgan
 Westchester Dept. of Labs & Research
 Date: 01/29/2002 Time: 10:08:34

original data

Sample: AD27984
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 1/29/02 10:07 Ended: 1/29/02 10:07 Analyst: MF
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		3.76		1.0
20	Surr_Triphenyl phosphate		5.29		1.0
21	Surr_Perylene-d12		4.50		1.0

Quantitation Report (QT Reviewed)

Data File : C:\MSDCHEM\1\DATA\121001B\27984.D

Acq On : 10 Dec 2001 9:39 pm

Sample : sample

Misc :

MS Integration Params: BENZO.P

Quant Time: Dec 10 22:17:02 2001

Vial: 8

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Quant Results File: 120501.RES

Original data

Quant Method : C:\MSDCHEM\1\5973N\120501.M (RTE Integrator)

Title : Quantitation For Method 525.2

Last Update : Thu Dec 06 10:31:04 2001

Response via : Initial Calibration

DataAcq Meth : 120501

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Acenaphthene-d10	18.32	164	6239087	5.00	ug/L	0.00
9) Phenanthrene D-10	22.66	188	9823940	5.00	ug/L	0.02
20) Chrysene D-12	30.44	240	5430888	5.00	ug/L	0.02

System Monitoring Compounds

2) 1,3-Dimethyl-2-nitrobenzen	13.24	134	1245913	3.76	ug/L	0.02
Spiked Amount	5.000		Recovery	=	75.20%	
19) Triphenyl Phosphate surr	29.70	326	1800736	5.29	ug/L	0.00
Spiked Amount	5.000		Recovery	=	105.80%	
21) Perylene D-12 surr	34.46	264	3432643	4.50	ug/L	0.02
Spiked Amount	5.000		Recovery	=	90.00%	

Target Compounds

					Qvalue
3) Hexachlorocyclopentadiene	0.00	237	0	N.D.	
4) Hexachlorobenzene	21.53	284	5875	0.01 ug/L #	67
5) EPTC	0.00	128	0	N.D.	
6) 2,6-Dinitrotoluene	0.00	165	0	N.D. d	
7) 2,4-Dinitrotoluene	18.66	165	10992	0.32 ug/L	98
8) Molinate	19.24	126	6269	0.01 ug/L #	1
10) Simazine	0.00	201	0	N.D.	
11) Atrazine	0.00	200	0	N.D.	
12) Alachlor	0.00	160	0	N.D.	
13) Terbacil	22.81	161	5945	0.01 ug/L #	1
14) Acetochlor	0.00	146	0	N.D.	
15) Metribuzin	0.00	198	0	N.D.	
16) Metolachlor	0.00	162	0	N.D. d	
17) Butachlor	0.00	160	0	N.D. d	
18) 4,4'-DDE	0.00	246	0	N.D. d	
22) Bis (2-Ethylhexyl) adipate	29.62	129	11694	0.12 ug/L #	1
23) Bis (2-Ethylhexyl) phthala	31.07	149	263727	0.36 ug/L	100
24) Benzo (a) pyrene	0.00	252	0	N.D. d	


(#) = qualifier out of range (m) = manual integration

27984.D 120501.M Thu Jan 10 11:58:17 2002

Quantitation report (QT reviewed)

Data File : C:\MSDCHEM\1\DATA\121001B\27984.D Vial: 8
Acq On : 10 Dec 2001 9:39 pm Operator: McLean
Sample : sample Inst : Instrumen
Misc : Multiplr: 1.00
MS Integration Params: BENZO.P
Quant Time: Jan 10 11:58 2002 Quant Results File: 120501.RES

Method : C:\MSDCHEM\1\5973N\120501.M (RTE Integrator)
Title : Quantitation For Method 525.2
Last Update : Thu Dec 06 10:31:04 2001
Response via : Initial Calibration



Quantitation Report

(QT Reviewed)

Data File : C:\MSDCHEM\1\DATA\012502A\27984RR.D

Acq On : 25 Jan 2002 3:24 pm

Sample : REINJ

Misc :

MS Integration Params: rteint.p

Quant Time: Jan 28 11:50:52 2002

Vial: 2

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Quant Results File: 012402B.RES

Rejected data

Quant Method : C:\MSDCHEM\1\5973N\012402B.M (RTE Integrator)

Title : Quantitation For Method 525.2

Last Update : Fri Jan 25 11:19:14 2002

Response via : Initial Calibration

DataAcq Meth : 012402DF

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Acenaphthene-d10	17.96	164	14351741	5.00	ug/L	0.00
9) Phenanthrene D-10	22.27	188	19016621	5.00	ug/L	0.00
20) Chrysene D-12	30.03	240	13407976	5.00	ug/L	0.00

System Monitoring Compounds

2) 1,3-Dimethyl-2-nitrobenzen	12.89	134	3518985	4.78	ug/L	0.00
Spiked Amount	5.000		Recovery	=	95.60%	
19) Triphenyl Phosphate	29.33	326	3075555	5.37	ug/L	0.00
Spiked Amount	5.000		Recovery	=	107.40%	
21) Perylene D-12	34.03	264	8720086	4.58	ug/L	0.00
Spiked Amount	5.000		Recovery	=	91.60%	

Target Compounds

					Qvalue
3) Hexachlorocyclopentadiene	0.00	237	0	N.D. d	
4) Hexachlorobenzene	21.13	284	13525	0.02 ug/L	91
5) EPTC	15.82	128	11298	N.D.	
6) 2,6-Dinitrotoluene	0.00	165	0	N.D. d	
7) 2,4-Dinitrotoluene	18.71	165	23259	0.93 ug/L	21
8) Molinate	18.85	126	5143	Below Cal #	1
10) Simazine	0.00	201	0	N.D. d	
11) Atrazine	0.00	200	0	N.D.	
12) Alachlor	23.31	160	5788	0.15 ug/L	66
13) Terbacil	0.00	161	0	N.D. d	
14) Acetochlor	0.00	146	0	N.D. d	
15) Metribuzin	0.00	198	0	N.D.	
16) Metolachlor	24.58	162	16493	0.23 ug/L	84
17) Butachlor	0.00	160	0	N.D. d	
18) 4,4'-DDE	0.00	246	0	N.D. d	
22) Bis (2-Ethylhexyl) adipate	29.33	129	42521	0.52 ug/L	54
23) Bis (2-Ethylhexyl) phthala	30.70	149	425749	0.53 ug/L	98
24) Benzo (a) pyrene	0.00	252	0	N.D. d	

(#) = qualifier out of range (m) = manual integration

27984RR.D 012402B.M Mon Jan 28 11:53:45 2002

Quantitation Report (Q1 Reviewed)

Data File : C:\MSDCHEM\1\DATA\012502A\27984RR.D

Vial: 2

Acq On : 25 Jan 2002 3:24 pm

Operator: McLean

Sample : REINJ

Inst : Instrumen

Misc :

Multiplr: 1.00

MS Integration Params: rteint.p

Quant Time: Jan 28 11:53 2002


Quant Results File: 012402B.RES

Method : C:\MSDCHEM\1\5973N\012402B.M (RTE Integrator)

Title : Quantitation For Method 525.2

Last Update : Fri Jan 25 11:19:14 2002

Response via : Initial Calibration



•
. Comments for Sample AD27984 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 01-30-2002 Time: 14:36:09

The recoveries for 6 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. The recoveries for 4 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges.