


User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 01/08/2002 Time: 14:59:26


Sample: AD27769
 Analysis: SLP525 Organic Extractables
 Units: ug
 Started: 12/9/01 14:57 Ended: 12/9/01 14:57 Analyst: TB
 Page: 1

	Component Name	Viol	Result
1	DFIPP		PASSED
2			
3			

Chromatogram Plot File: E:\DFTA1209 Date: Dec-09-1991 05:28:13
Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 4096093



Chromatogram Plot File: E:\DFTA1209 Date: Dec-09-1991 05:28:13
 Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN
 Scan No: 1080 Retention Time: 18:00 RIC: 34548 Mass Range: 47 - 207
 Plotted: 1080 to 1120 Range: 1 to 1860 100% = 4096093



Mass List

File: E:\DFTA1209 Date: Dec-09-1991 05:28:13

Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN

Scan No: 1095 Retention Time: 18:15 RIC: 4096093 Mass Range: 47 - 444

Peaks: 196 Mass of Base Peak: 442 Int: 591228 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
47	3,070	0.52	82	2,631	0.45	122	5,994	1.01
48	2,339	0.40	84	3,654	0.62	123	4,093	0.69
49	12,280	2.08	85	3,947	0.67	124	4,824	0.82
50	83,187	14.07	86	9,502	1.61	125	2,046	0.35
51	265,789	44.96	87	4,532	0.77	127	151,900	25.69
52	9,064	1.53	91	2,046	0.35	128	8,771	1.48
55	3,947	0.67	92	1,461	0.25	129	66,959	11.33
56	5,994	1.01	93	30,116	5.09	130	4,532	0.77
57	11,403	1.93	94	1,608	0.27	131	1,900	0.32
61	4,678	0.79	98	22,076	3.73	134	4,093	0.69
62	5,116	0.87	99	8,918	1.51	135	4,385	0.74
63	6,140	1.04	101	5,847	0.99	136	1,900	0.32
65	2,777	0.47	103	3,362	0.57	137	4,385	0.74
68	2,923	0.49	104	4,532	0.77	141	12,573	2.13
69	171,783	29.06	105	4,532	0.77	142	2,923	0.49
72	877	0.15	106	1,169	0.20	143	1,315	0.22
73	3,070	0.52	107	62,134	10.51	147	5,409	0.91
74	28,362	4.80	108	5,994	1.01	148	7,748	1.31
75	29,824	5.04	109	3,654	0.62	149	3,362	0.57
76	7,163	1.21	110	119,005	20.13	151	2,631	0.45
77	187,426	31.70	111	16,812	2.84	152	1,315	0.22
78	14,181	2.40	113	1,754	0.30	153	2,631	0.45
79	24,269	4.10	116	2,485	0.42	154	1,169	0.20
80	15,497	2.62	117	92,982	15.73	155	4,093	0.69
81	10,672	1.81	118	6,140	1.04	156	5,994	1.01

Mass List

File: E:\DFTA1209

Date: Dec-09-1991 05:28:13

Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN

Scan No: 1095

Retention Time: 18:15

RIC: 4096093

Mass Range: 47 - 444

Peaks: 196

Mass of Base Peak: 442

Int: 591228

Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
157	1,169	0.20	195	1,023	0.17	234	3,070	0.52
159	2,046	0.35	196	5,994	1.01	235	1,754	0.30
160	1,608	0.27	197	3,508	0.59	237	3,070	0.52
161	4,678	0.79	198	378,216	63.97	239	1,900	0.32
165	5,555	0.94	199	23,099	3.91	241	1,169	0.20
166	1,900	0.32	203	2,046	0.35	242	3,362	0.57
167	23,976	4.06	204	11,695	1.98	243	2,923	0.49
168	13,157	2.23	205	18,274	3.09	244	26,023	4.40
170	1,169	0.20	206	73,099	12.36	245	6,432	1.09
171	1,900	0.32	207	9,356	1.58	246	5,847	0.99
174	2,192	0.37	208	3,947	0.67	250	1,169	0.20
175	7,309	1.24	211	1,315	0.22	252	1,023	0.17
177	4,824	0.82	215	1,461	0.25	253	2,192	0.37
178	2,631	0.45	216	1,169	0.20	254	1,315	0.22
179	9,649	1.63	217	23,830	4.03	255	252,485	42.71
180	8,187	1.38	218	3,216	0.54	256	28,216	4.77
181	3,362	0.57	221	7,602	1.29	257	1,315	0.22
184	2,339	0.40	222	1,608	0.27	258	14,766	2.50
185	5,701	0.96	223	5,555	0.94	259	1,461	0.25
186	58,333	9.87	224	46,783	7.91	260	1,023	0.17
187	14,619	2.47	225	11,549	1.95	265	5,263	0.89
188	3,070	0.52	226	1,608	0.27	273	7,456	1.26
189	3,508	0.59	227	21,637	3.66	274	29,824	5.04
192	3,508	0.59	229	2,631	0.45	275	141,666	23.96
193	5,116	0.87	233	1,461	0.25	276	16,081	2.72

Mass List

File: E:\DFTA1209 Date: Dec-09-1991 05:28:13

Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN


Scan No: 1095 Retention Time: 18:15 RIC: 4096093 Mass Range: 47 - 444

Peaks: 196 Mass of Base Peak: 442 Int: 591228 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
226	1,608	0.27	273	7,456	1.26	341	1,900	0.32
227	21,637	3.66	274	29,824	5.04	343	2,046	0.35
229	2,631	0.45	275	141,666	23.96	346	1,754	0.30
233	1,461	0.25	276	16,081	2.72	352	5,555	0.94
234	3,070	0.52	277	12,573	2.13	353	5,263	0.89
235	1,754	0.30	278	2,046	0.35	354	5,263	0.89
237	3,070	0.52	281	1,754	0.30	355	1,754	0.30
239	1,900	0.32	285	3,362	0.57	365	25,292	4.28
241	1,169	0.20	291	1,608	0.27	366	5,847	0.99
242	3,362	0.57	293	2,046	0.35	372	8,625	1.46
243	2,923	0.49	294	2,046	0.35	373	3,216	0.54
244	26,023	4.40	295	1,023	0.17	374	1,608	0.27
245	6,432	1.09	296	27,046	4.57	383	4,824	0.82
246	5,847	0.99	297	5,701	0.96	390	2,046	0.35
250	1,169	0.20	302	2,046	0.35	402	3,947	0.67
252	1,023	0.17	303	2,631	0.45	403	3,654	0.62
253	2,192	0.37	304	1,608	0.27	404	2,192	0.37
254	1,315	0.22	314	2,192	0.37	421	3,216	0.54
255	252,485	42.71	315	5,701	0.96	422	1,608	0.27
256	28,216	4.77	323	16,520	2.79	423	27,485	4.65
257	1,315	0.22	324	3,070	0.52	424	6,432	1.09
258	14,766	2.50	327	2,192	0.37	441	82,894	14.02
259	1,461	0.25	333	1,900	0.32	442	591,228	100.00
260	1,023	0.17	334	8,040	1.36	443	115,058	19.46
265	5,263	0.89	335	5,701	0.96	444	8,918	1.51

Passed

Chromatogram Plot File: E:\DFTA1209 Date: Dec-09-1991 05:28:13
 Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN
 Scan No: 1280 Retention Time: 21:20 RIC: 57750 Mass Range: 45 - 355
 Plotted: 1250 to 1280 Range: 1 to 1860 100% = 4096093



Chromatogram Plot

File: E:\DFTA1209

Date: Dec-09-1991 05:28:13

Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN

Scan No: 1300

Retention Time: 21:40


RIC: 57476

Mass Range: 45 - 281

Plotted: 1272 to 1300

Range: 1 to 1860

100% = 4096093



Chromatogram Plot

File: E:\DFTA1209 Date: Dec-09-1991 05:28:13


Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN

Scan No: 1367 Retention Time: 22:47 RIC: 64278 Mass Range: 45 - 355

Plotted: 1337 to 1367

Range: 1 to 1860

100% = 4096093



Chromatogram Plot

File: E:\DDTA1209

Date: Dec-09-1991 06:37:06

Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DDT

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0


Plotted: 1 to 1860

Range: 1 to 1860


100% = 2710005



Chromatogram Plot File: E:\DDTA1209 Date: Dec-09-1991 06:37:06
 Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DDT
 Scan No: 1250 Retention Time: 20:50 RIC: 29018 Mass Range: 45 - 355
 Plotted: 1230 to 1250 Range: 1 to 1860 100% = 2710005



Chromatogram Plot File: E:\DDTA1209 Date: Dec-09-1991 06:37:06
Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DDT
Scan No: 1300 Retention Time: 21:40 RIC: 51846 Mass Range: 47 - 284
Plotted: 1262 to 1300 Range: 1 to 1860 100% = 2710005



Chromatogram Plot

File: E:\DDTA1209

Date: Dec-09-1991 06:37:06

Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DDT

Scan No: 1325

Retention Time: 22:05


RIC: 45982

Mass Range: 45 - 355


Plotted: 1300 to 1325

Range: 1 to 1860


100% = 2710005



Chromatogram Plot File: E:\DFTB1209 Date: Dec-09-1991 18:05:16
Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 3458982



Chromatogram Plot File: E:\DFTB1209 Date: Dec-09-1991 18:05:16
Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN
Scan No: 1080 Retention Time: 18:00 RIC: 29823 Mass Range: 47 - 207
Plotted: 1080 to 1120 Range: 1 to 1860 100% = 3458982
9.07%



Mass List

File: E:\DFTB1209 Date: Dec-09-1991 18:05:16

Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN

Scan No: 1094 Retention Time: 18:14 RIC: 3458982 Mass Range: 49 - 444

Peaks: 203 Mass of Base Peak: 442 Int: 471447 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
49	12,236	2.60	83	1,315	0.28	114	1,315	0.28
50	86,447	18.34	84	1,973	0.42	115	1,710	0.36
51	249,210	52.86	85	2,631	0.56	116	1,052	0.22
52	9,736	2.07	86	7,236	1.53	117	73,289	15.55
55	3,815	0.81	87	2,763	0.59	118	3,684	0.78
56	4,868	1.03	91	1,447	0.31	119	921	0.20
57	10,657	2.26	92	2,894	0.61	122	6,842	1.45
60	1,578	0.33	93	24,605	5.22	123	4,078	0.86
61	4,342	0.92	94	1,842	0.39	124	2,763	0.59
62	5,657	1.20	96	2,631	0.56	125	3,289	0.70
63	5,657	1.20	97	1,315	0.28	127	123,552	26.21
65	3,421	0.73	98	20,657	4.38	128	8,289	1.76
67	1,973	0.42	99	8,947	1.90	129	47,500	10.08
68	2,500	0.53	100	1,315	0.28	130	5,526	1.17
69	151,315	32.10	101	4,868	1.03	131	921	0.20
72	1,447	0.31	103	4,473	0.95	132	1,052	0.22
73	2,631	0.56	104	4,473	0.95	133	1,052	0.22
74	29,078	6.17	105	4,868	1.03	135	5,394	1.14
75	24,605	5.22	106	1,578	0.33	136	3,157	0.67
76	8,552	1.81	107	49,078	10.41	137	1,447	0.31
77	167,894	35.61	108	6,447	1.37	140	2,236	0.47
78	7,763	1.65	109	3,026	0.64	141	5,789	1.23
79	18,421	3.91	110	100,657	21.35	142	2,105	0.45
80	11,447	2.43	111	13,947	2.96	143	2,236	0.47
81	10,000	2.12	112	1,315	0.28	146	1,184	0.25

Mass List

File: E:\DFTB1209 Date: Dec-09-1991 18:05:16

Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN

Scan No: 1094 Retention Time: 18:14 RIC: 3458982 Mass Range: 49 - 444

Peaks: 203 Mass of Base Peak: 442 Int: 471447 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
147	7,368	1.56	179	8,815	1.87	215	1,447	0.31
148	12,105	2.57	180	7,631	1.62	216	1,578	0.33
149	4,342	0.92	181	2,500	0.53	217	16,842	3.57
151	2,368	0.50	183	1,973	0.42	218	1,973	0.42
153	1,578	0.33	185	4,342	0.92	219	1,052	0.22
154	3,684	0.78	186	50,921	10.80	221	7,368	1.56
155	3,421	0.73	187	15,131	3.21	222	3,157	0.67
156	4,210	0.89	188	2,368	0.50	223	2,105	0.45
158	2,500	0.53	189	1,184	0.25	224	33,815	7.17
159	2,105	0.45	192	1,578	0.33	225	5,921	1.26
160	1,184	0.25	193	2,763	0.59	226	1,710	0.36
161	3,684	0.78	196	6,578	1.40	227	15,657	3.32
164	921	0.20	197	2,236	0.47	228	1,578	0.33
165	2,631	0.56	198	313,684	66.54	229	1,842	0.39
166	2,500	0.53	199	25,789	5.47	234	2,105	0.45
167	17,894	3.80	200	1,447	0.31	235	3,947	0.84
168	6,447	1.37	202	2,368	0.50	236	1,973	0.42
169	1,184	0.25	203	1,447	0.31	237	2,236	0.47
172	1,447	0.31	204	10,789	2.29	242	1,052	0.22
173	1,578	0.33	205	15,789	3.35	243	2,763	0.59
174	3,289	0.70	206	59,210	12.56	244	26,578	5.64
175	7,763	1.65	207	10,131	2.15	245	5,526	1.17
176	1,184	0.25	208	1,578	0.33	246	7,500	1.59
177	1,578	0.33	209	1,184	0.25	247	1,447	0.31
178	1,447	0.31	211	2,105	0.45	252	1,447	0.31

Mass List

File: E:\DFTB1209 Date: Dec-09-1991 18:05:16

Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN


Scan No: 1094 Retention Time: 18:14 RIC: 3458982 Mass Range: 49 - 444

Peaks: 203 Mass of Base Peak: 442 Int: 471447 Threshold: 0 %


Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
218	1,973	0.42	258	12,500	2.65	333	1,447	0.31
219	1,052	0.22	259	1,184	0.25	334	6,184	1.31
221	7,368	1.56	265	5,000	1.06	335	2,105	0.45
222	3,157	0.67	272	2,500	0.53	336	1,184	0.25
223	2,105	0.45	273	4,605	0.98	341	1,184	0.25
224	33,815	7.17	274	24,736	5.25	352	2,763	0.59
225	5,921	1.26	275	103,289	21.91	353	2,894	0.61
226	1,710	0.36	276	14,605	3.10	354	6,842	1.45
227	15,657	3.32	277	8,815	1.87	355	2,631	0.56
228	1,578	0.33	284	1,973	0.42	365	25,263	5.36
229	1,842	0.39	285	1,315	0.28	366	1,447	0.31
234	2,105	0.45	290	1,052	0.22	372	7,236	1.53
235	3,947	0.84	293	1,578	0.33	373	1,315	0.28
236	1,973	0.42	294	1,710	0.36	383	1,578	0.33
237	2,236	0.47	296	22,105	4.69	401	1,052	0.22
242	1,052	0.22	297	4,736	1.00	402	1,842	0.39
243	2,763	0.59	303	2,894	0.61	403	3,815	0.81
244	26,578	5.64	314	1,973	0.42	421	3,157	0.67
245	5,526	1.17	315	4,605	0.98	422	3,026	0.64
246	7,500	1.59	316	1,315	0.28	423	21,447	4.55
247	1,447	0.31	321	1,710	0.36	424	6,973	1.48
252	1,447	0.31	322	921	0.20	441	69,868	14.82
255	218,026	46.25	323	8,421	1.79	442	471,447	100.00
256	31,184	6.61	324	1,184	0.25	443	88,815	18.84
257	1,842	0.39	327	1,315	0.28	444	9,868	2.09

Passed

Chromatogram Plot File: E:\DFTB1209 Date: Dec-09-1991 18:05:16
Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN
Scan No: 1280 Retention Time: 21:20 RIC: 54391 Mass Range: 45 - 355
Plotted: 1250 to 1280 Range: 1 to 1860 100% = 3458982



Chromatogram Plot File: E:\DFTB1209 Date: Dec-09-1991 18:05:16
Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN
Scan No: 1302 Retention Time: 21:42 RIC: 67002 Mass Range: 45 - 355
Plotted: 1272 to 1302 Range: 1 to 1860 100% = 3458982



Chromatogram Plot

File: E:\DFTB1209 Date: Dec-09-1991 18:05:16


Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DFTPP/ENDRIN

Scan No: 1367 Retention Time: 22:47 RIC: 67073 Mass Range: 45 - 355

Plotted: 1337 to 1367

Range: 1 to 1860

100% = 3458982



Chromatogram Plot

File: E:\DDTB1209

Date: Dec-09-1991 19:14:00

Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DDT

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0


Plotted: 1 to 1860

Range: 1 to 1860


100% = 1565682



Chromatogram Plot File: E:\DDTB1209 Date: Dec-09-1991 19:14:00
Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DDT
Scan No: 1255 Retention Time: 20:55 RIC: 25890 Mass Range: 47 - 281
Plotted: 1230 to 1255 Range: 1 to 1860 100% = 1565682
6.74%



Chromatogram Plot File: E:\DDTB1209 Date: Dec-09-1991 19:14:00
Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DDT
Scan No: 1300 Retention Time: 21:40 RIC: 47760 Mass Range: 45 - 284
Plotted: 1262 to 1300 Range: 1 to 1860 100% = 1565682



Chromatogram Plot File: E:\DDTB1209 Date: Dec-09-1991 19:14:00
Comment: BOHLK; METHOD 525; INST 204; 12/09/01; DDT
Scan No: 1315 Retention Time: 21:55 RIC: 50847 Mass Range: 45 - 285
Plotted: 1295 to 1315 Range: 1 to 1860 100% = 1565682

