

User: Fakhouri, Morgan
 Westchester Dept. of Labs & Research
 Date: 12/20/2001 Time: 11:49:54

Sample: AD28847
 Analysis: \$LP525 Organic Extractables
 Units: ug
 Started: 12/20/01 11:44 Ended: 12/20/01 11:44 Analyst: MF
 Page: 1

	Component Name	Viol	Result
1	DFTPP		passed
2	ENDRIN		7.6
3	DDT		17.1

DFTPP

Data File : C:\MSDCHEM\1\DATA\121801\DFTPP5.D

Acq On : 19 Dec 2001 2:50 am

Sample :

Misc :

MS Integration Params: BENZO.P

Vial: 10

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\121401.M (RTE Integrator)

Title : Quantitation For Method 525.2

Spectrum Information: Scan 857

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.8	241280	PASS
68	69	0.00	2	1.5	4494	PASS
69	198	0.00	100	31.9	297984	PASS
70	69	0.00	2	0.5	1442	PASS
127	198	10	80	51.0	476032	PASS
197	198	0.00	2	0.0	0	PASS
198	198	50	100	100.0	933952	PASS
199	198	5	9	6.0	55696	PASS
275	198	10	60	19.5	182208	PASS
365	198	1	100	1.7	15765	PASS
441	443	0.01	100	76.3	107632	PASS
442	198	50	100	77.1	719808	PASS
443	442	15	24	19.6	141056	PASS

DFTPP5.D 121401.M

Thu Dec 20 09:33:08 2001

Area Percent Report

Data File : C:\MSDCHEM\1\DATA\121801\DFTPP5.D

Vial: 10

Acq On : 19 Dec 2001 2:50 am

Operator: McLean

Sample :

Inst : Instrumen

Misc :

Multiplr: 1.00

MS Integration Params: BENZO.P

Method : C:\MSDCHEM\1\5973N\121401.M (RTE Integrator)

Title : Quantitation For Method 525.2

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 2500 Area counts

Start Thrs: 0.15

Max Peaks: 10

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	27.876	1296	1301	1305	rM 2	3935921	7296763	100.00%	92.133%
2	28.466	1362	1366	1370	rM 7	106411	254695	3.49%	3.216%
3	28.992	1421	1424	1426	rM 2	194556	368341	5.05%	4.651%

Sum of corrected areas: 7919799

Area Percent Report

Data File : C:\MSDCHEM\1\DATA\121801\DDT5.D

Acq On : 19 Dec 2001 3:29 am

Sample :

Misc :

MS Integration Params: BENZO.P

Vial: 11

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\121401.M (RTE Integrator)

Title : Quantitation For Method 525.2

Smoothing : ON

Sampling : 1

Start Thrs: 0.15

Stop Thrs : 0

Filtering: 5

Min Area: 2500 Area counts

Max Peaks: 10

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	28.266	1342	1344	1347	rM	667983	1239729	15.14%	12.541%
2	28.992	1419	1424	1427	rM 2	211488	458428	5.60%	4.637%
3	29.146	1433	1441	1449	rM	4249424	8187352	100.00%	82.822%

Sum of corrected areas: 9885509