

User: Fakhouri, Morgan
 Westchester Dept. of Labs & Research
 Date: 12/12/2001 Time: 15:58:15

Sample: AD27232
 Analysis: \$LP525 Organic Extractables
 Units: ug
 Started: 11/15/01 15:57 Ended: 12/6/01 15:57 Analyst: MF
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result
1	DFTPP		passes
2	ENDRIN		17.8
3	DDT		19.8

DFTPP

Data File : C:\MSDCHEM\1\DATA\120601\DFTPP.D
Acq On : 6 Dec 2001 4:02 pm
Sample : dftpp +ddt+end
Misc :
MS Integration Params: BENZO.P

Vial: 1
Operator: McLean
Inst : Instrumen
Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\120501.M (RTE Integrator)
Title : Quantitation For Method 525.2

Spectrum Information: Scan 868

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.4	282816	PASS
68	69	0.00	2	1.4	4623	PASS
69	198	0.00	100	34.9	335552	PASS
70	69	0.00	2	0.6	2155	PASS
127	198	10	80	52.5	505216	PASS
197	198	0.00	2	0.3	2800	PASS
198	198	50	100	100.0	961984	PASS
199	198	5	9	5.8	55744	PASS
275	198	10	60	16.9	162176	PASS
365	198	1	100	1.9	18104	PASS
441	443	0.01	100	75.0	92472	PASS
442	198	50	100	61.8	594496	PASS
443	442	15	24	20.7	123216	PASS

DFTPP.D 120501.M Wed Dec 12 11:33:01 2001

Area Percent Report

Data File : C:\MSDCHEM\1\DATA\120601\DFTPP.D

Acq On : 6 Dec 2001 4:02 pm

Sample : dftpp +ddt+end

Misc :

MS Integration Params: BENZO.P

Vial: 1

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\120501.M (RTE Integrator)

Title : Quantitation For Method 525.2

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 2500 Area counts

Start Thrs: 0.15

Max Peaks: 10

Stop Thrs : 0

Peak Location: TOP


If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	27.994	1307	1314	1320	rM 2	3506155	7155032	100.00%	82.122%
2	28.384	1354	1357	1361	rM	470199	943183	13.18%	10.825%
3	28.593	1376	1380	1384	rM 3	305531	614524	8.59%	7.053%

Sum of corrected areas: 8712739



Area Percent Report

Data File : C:\MSDCHEM\1\DATA\120601\DFTPP.D Vial: 1
 Acq On : 6 Dec 2001 4:02 pm Operator: McLean
 Sample : dftpp +ddt+end Inst : Instrumen
 Misc : Multiplr: 1.00
 MS Integration Params: BENZO.P

Method : C:\MSDCHEM\1\5973N\120501.M (RTE Integrator)
 Title : Quantitation For Method 525.2
 Smoothing : ON Filtering: 5
 Sampling : 1 Min Area: 2500 Area counts
 Start Thrs: 0.15 Max Peaks: 10
 Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
 Peak separation: 5

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	29.255	1447	1453	1467	rM	5763207	11406333	100.00%	80.219%
2	30.225	1557	1560	1563	rM 2	1388682	2812649	24.66%	19.781%

Sum of corrected areas: 14218982

DFTPP

Data File : C:\MSDCHEM\1\DATA\120601\DFTPPB.D
 Acq On : 7 Dec 2001 1:46 am
 Sample : big 3 dftpp
 Misc :
 MS Integration Params: BENZO.P

Vial: 1
 Operator: McLean
 Inst : Instrumen
 Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\120501.M (RTE Integrator)
 Title : Quantitation For Method 525.2

Spectrum Information: Scan 868

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.3	281472	PASS
68	69	0.00	2	1.3	4637	PASS
69	198	0.00	100	35.3	363776	PASS
70	69	0.00	2	0.5	1969	PASS
127	198	10	80	53.9	556224	PASS
197	198	0.00	2	0.0	0	PASS
198	198	50	100	100.0	1031936	PASS
199	198	5	9	6.3	64680	PASS
275	198	10	60	18.5	191040	PASS
365	198	1	100	1.3	13163	PASS
441	443	0.01	100	71.3	93968	PASS
442	198	50	100	65.5	675648	PASS
443	442	15	24	19.5	131712	PASS

DFTPPB.D 120501.M Wed Dec 12 11:36:54 2001

Area Percent Report

Data File : C:\MSDCHEM\1\DATA\120601\DFTPPB.D

Acq On : 7 Dec 2001 1:46 am

Sample : big 3 dftpp

Misc :

MS Integration Params: BENZO.P

Vial: 1

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\120501.M (RTE Integrator)

Title : Quantitation For Method 525.2

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 2500 Area counts

Start Thrs: 0.15

Max Peaks: 10

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	27.985	1308	1313	1320	rM 2	2151423	4228772	100.00%	73.526%
2	28.384	1355	1357	1361	rM	422704	735435	17.39%	12.787%
3	28.584	1377	1379	1382	rM 2	471581	787156	18.61%	13.686%

Sum of corrected areas: 5751363

14

Handwritten: 

Area Percent Report

Data File : C:\MSDCHEM\1\DATA\120601\DFTPPB.D Vial: 1
 Acq On : 7 Dec 2001 1:46 am Operator: McLean
 Sample : big 3 dftpp Inst : Instrumen
 Misc : Multiplr: 1.00
 MS Integration Params: BENZO.P

Method : C:\MSDCHEM\1\5973N\120501.M (RTE Integrator)
 Title : Quantitation For Method 525.2
 Smoothing : ON Filtering: 5
 Sampling : 1 Min Area: 2500 Area counts
 Start Thrs: 0.15 Max Peaks: 10
 Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
 Peak separation: 5

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	29.255	1444	1453	1463	rM	5667754	10624699	100.00%	76.126%
2	30.225	1557	1560	1562	rM 2	1839474	3331951	31.36%	23.874%

Sum of corrected areas: 13956650