

User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 01/04/2002 Time: 11:24:45

Sample: AD27118
Analysis: \$LP525 Organic Extractables
Units: ug
Started: 12/4/01 11:22 Ended: 12/4/01 11:22 Analyst: TB
Page: 1

	Component Name	Viol	Result
1	DFTPP		PASSED
2	ENDRN		
3	DDT		4.9

Chromatogram Plot

File: E:\DFTA1204

Date: Dec-04-1991 17:10:35

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0

Plotted: 1 to 1859

Range: 1 to 1859

100% = 4482104



Chromatogram Plot

File: E:\DFTA1204 Date: Dec-04-1991 17:10:35

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN


Scan No: 1080 Retention Time: 18:00 RIC: 27423 Mass Range: 47 - 86

Plotted: 1080 to 1120

Range: 1 to 1859

100% = 4482104

9.43%



Mass List

File: E:\DFTA1204 Date: Dec-04-1991 17:10:35

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN

Scan No: 1093 Retention Time: 18:13 RIC: 4482104 Mass Range: 49 - 444

Peaks: 201 Mass of Base Peak: 442 Int: 621000 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
49	12,333	1.99	80	14,833	2.39	111	19,833	3.19
50	103,333	16.64	81	16,833	2.71	112	2,500	0.40
51	304,333	49.01	82	3,333	0.54	116	4,833	0.78
52	13,833	2.23	83	3,000	0.48	117	81,000	13.04
55	5,333	0.86	84	2,666	0.43	118	4,500	0.72
56	8,333	1.34	85	4,000	0.64	122	3,666	0.59
57	12,666	2.04	86	11,500	1.85	123	9,666	1.56
58	2,166	0.35	88	1,833	0.30	124	3,333	0.54
60	1,166	0.19	91	2,166	0.35	126	1,000	0.16
61	3,500	0.56	92	3,666	0.59	127	155,500	25.04
62	4,666	0.75	93	37,333	6.01	128	13,666	2.20
63	12,833	2.07	94	3,500	0.56	129	64,333	10.36
64	1,333	0.21	96	1,333	0.21	130	8,166	1.31
65	4,166	0.67	97	1,500	0.24	134	2,333	0.38
67	5,166	0.83	98	27,833	4.48	135	10,500	1.69
68	3,833	0.62	99	15,000	2.42	136	4,500	0.72
69	200,000	32.21	100	2,666	0.43	137	2,333	0.38
72	2,166	0.35	101	3,166	0.51	139	2,000	0.32
73	1,666	0.27	103	4,833	0.78	141	10,166	1.64
74	29,166	4.70	104	4,500	0.72	142	2,000	0.32
75	28,833	4.64	105	3,833	0.62	143	2,000	0.32
76	14,500	2.33	107	63,333	10.20	147	3,166	0.51
77	199,500	32.13	108	9,166	1.48	148	11,166	1.80
78	10,500	1.69	109	1,500	0.24	149	2,333	0.38
79	21,333	3.44	110	135,166	21.77	150	3,166	0.51

Mass List
 File: E:\DFT\1204 Date: Dec-04-1991 17:10:35
 Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN
 Scan No: 1093 Retention Time: 18:13 RIC: 4482104 Mass Range: 49 - 444
 # Peaks: 201 Mass of Base Peak: 442 Int: 621000 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
152	1,833	0.30	189	3,166	0.51	229	5,000	0.81
154	2,666	0.43	192	2,333	0.38	231	1,666	0.27
155	3,000	0.48	193	2,333	0.38	233	1,833	0.30
156	4,666	0.75	196	5,333	0.86	234	1,500	0.24
158	2,833	0.46	197	4,000	0.64	235	1,666	0.27
159	2,500	0.40	198	422,500	68.04	236	1,333	0.21
160	5,333	0.86	199	35,666	5.74	240	1,333	0.21
161	6,000	0.97	200	1,833	0.30	241	1,833	0.30
162	2,000	0.32	201	1,333	0.21	243	3,833	0.62
165	4,500	0.72	204	7,333	1.18	244	36,666	5.90
166	4,333	0.70	205	21,000	3.38	245	7,333	1.18
167	24,333	3.92	206	80,500	12.96	246	7,166	1.15
168	12,833	2.07	207	14,000	2.25	247	2,500	0.40
170	1,833	0.30	208	2,666	0.43	249	2,333	0.38
174	1,833	0.30	209	1,833	0.30	251	2,500	0.40
175	7,000	1.13	210	2,666	0.43	252	1,666	0.27
177	2,333	0.38	216	2,333	0.38	255	283,833	45.71
179	12,000	1.93	217	15,166	2.44	256	39,000	6.28
180	5,333	0.86	221	8,333	1.34	257	3,333	0.54
181	2,833	0.46	223	7,500	1.21	258	15,833	2.55
182	1,500	0.24	224	61,000	9.82	259	1,500	0.24
185	10,500	1.69	225	10,500	1.69	260	2,000	0.32
186	63,333	10.20	226	2,500	0.40	265	5,166	0.83
187	15,500	2.50	227	20,500	3.30	273	7,333	1.18
188	3,166	0.51	228	3,833	0.62	274	27,166	4.37

Mass List

File: E:\DFTA1204 Date: Dec-04-1991 17:10:35

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN

Scan No: 1093 Retention Time: 18:13 RIC: 4482104 Mass Range: 49 - 444

Peaks: 201 Mass of Base Peak: 442 Int: 621000 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
231	1,666	0.27	276	19,333	3.11	354	5,166	0.83
233	1,833	0.30	277	12,666	2.04	355	3,666	0.59
234	1,500	0.24	278	2,166	0.35	361	1,333	0.21
235	1,666	0.27	283	2,333	0.38	365	21,000	3.38
236	1,333	0.21	285	3,666	0.59	366	3,166	0.51
240	1,333	0.21	293	2,500	0.40	368	1,500	0.24
241	1,833	0.30	296	29,166	4.70	372	8,500	1.37
243	3,833	0.62	297	3,166	0.51	373	2,500	0.40
244	36,666	5.90	303	2,666	0.43	374	2,333	0.38
245	7,333	1.18	304	2,333	0.38	383	4,333	0.70
246	7,166	1.15	308	2,166	0.35	384	1,500	0.24
247	2,500	0.40	315	4,833	0.78	391	2,500	0.40
249	2,333	0.38	316	3,000	0.48	398	1,166	0.19
251	2,500	0.40	321	2,333	0.38	402	2,833	0.46
252	1,666	0.27	322	2,166	0.35	403	5,666	0.91
255	283,833	45.71	323	14,000	2.25	404	4,666	0.75
256	39,000	6.28	324	3,500	0.56	421	5,500	0.89
257	3,333	0.54	325	1,000	0.16	422	4,000	0.64
258	15,833	2.55	327	4,666	0.75	423	24,500	3.95
259	1,500	0.24	334	17,833	2.87	424	7,500	1.21
260	2,000	0.32	335	4,166	0.67	425	2,666	0.43
265	5,166	0.83	341	1,500	0.24	441	94,833	15.27
273	7,333	1.18	346	3,666	0.59	442	621,000	100.00
274	27,166	4.37	352	2,500	0.40	443	129,500	20.85
275	131,000	21.10	353	6,666	1.07	444	13,500	2.17

Chromatogram Plot

File: E:\DFTA1204

Date: Dec-04-1991 17:10:35

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN

Scan No: 1280

Retention Time: 21:20


RIC: 59146

Mass Range: 45 - 281


Plotted: 1250 to 1280

Range: 1 to 1859


100% = 4482104



Chromatogram Plot File: E:\DFTA1204 Date: Dec-04-1991 17:10:35
 Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN
 Scan No: 1302 Retention Time: 21:42 RIC: 48109 Mass Range: 45 - 281
 Plotted: 1272 to 1302 Range: 1 to 1859 100% = 4482104



Chromatogram Plot File: E:\DFTA1204 Date: Dec-04-1991 17:10:35
Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN
Scan No: 1367 Retention Time: 22:47 RIC: 57473 Mass Range: 45 - 281
Plotted: 1337 to 1367 Range: 1 to 1859 100% = 4482104



Chromatogram Plot

File: E:\DDTA1204

Date: Dec-04-1991 17:48:22

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DDT

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0

Plotted: 1 to 1860

Range: 1 to 1860

100% = 4225927



Chromatogram Plot

File: E:\DDTA1204 Date: Dec-04-1991 17:48:22


Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DDT

Scan No: 1260 Retention Time: 21:00 RIC: 30444 Mass Range: 45 - 355

Plotted: 1229 to 1260

Range: 1 to 1860

100% = 4225927



Chromatogram Plot

File: E:\DDTA1204

Date: Dec-04-1991 17:48:22

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DDT

Scan No: 1302

Retention Time: 21:42


RIC: 124923

Mass Range: 47 - 284

Plotted: 1262 to 1302

Range: 1 to 1860

100% = 4225927



Chromatogram Plot

File: E:\DDTA1204

Date: Dec-04-1991 17:48:22

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DDT

Scan No: 1310

Retention Time: 21:50


RIC: 46288

Mass Range: 47 - 281

Plotted: 1290 to 1310

Range: 1 to 1860

100% = 4225927



Chromatogram Plot

File: E:\DFTC1204

Date: Dec-05-1991 08:21:53

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0

Plotted: 1 to 1860

Range: 1 to 1860

100% = 3443458



Chromatogram Plot


File: E:\DFTC1204 Date: Dec-05-1991 08:21:53

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN

Scan No: 1080 Retention Time: 18:00 RIC: 31494 Mass Range: 47 - 207

Plotted: 1080 to 1120

Range: 1 to 1860 100% = 3443458



Mass List

File: E:\DFTC1204 Date: Dec-05-1991 08:21:53

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN

Scan No: 1094 Retention Time: 18:14 RIC: 3443458 Mass Range: 47 - 444

Peaks: 189 Mass of Base Peak: 442 Int: 426754 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
47	3,801	0.89	83	3,362	0.79	122	4,385	1.03
48	1,754	0.41	84	4,093	0.96	123	3,801	0.89
49	16,081	3.77	85	2,339	0.55	124	1,754	0.41
50	87,719	20.55	86	6,578	1.54	125	2,777	0.65
51	253,947	59.51	87	1,169	0.27	127	111,111	26.04
52	10,233	2.40	91	2,046	0.48	128	8,625	2.02
55	2,777	0.65	92	4,824	1.13	129	69,152	16.20
56	7,163	1.68	93	30,555	7.16	130	7,894	1.85
57	9,210	2.16	98	23,830	5.58	135	5,409	1.27
61	1,900	0.45	99	12,134	2.84	136	1,315	0.31
62	3,508	0.82	101	3,801	0.89	138	1,461	0.34
63	7,894	1.85	103	4,532	1.06	139	877	0.21
67	1,461	0.34	104	5,701	1.34	141	4,385	1.03
69	151,023	35.39	105	6,578	1.54	142	3,070	0.72
70	1,461	0.34	107	71,198	16.68	143	1,023	0.24
73	6,578	1.54	108	8,918	2.09	145	2,777	0.65
74	31,140	7.30	109	1,608	0.38	146	2,777	0.65
75	23,684	5.55	110	83,918	19.66	147	5,409	1.27
76	11,257	2.64	111	13,596	3.19	148	11,842	2.77
77	155,409	36.42	112	2,631	0.62	149	1,315	0.31
78	10,380	2.43	116	2,485	0.58	151	1,315	0.31
79	18,859	4.42	117	78,654	18.43	152	3,654	0.86
80	8,771	2.06	118	3,947	0.92	153	2,631	0.62
81	12,426	2.91	119	2,192	0.51	154	1,169	0.27
82	1,461	0.34	121	1,169	0.27	155	1,461	0.34

Mass List

File: E:\DFTC1204 Date: Dec-05-1991 08:21:53

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN

Scan No: 1094 Retention Time: 18:14 RIC: 3443458 Mass Range: 47 - 444

Peaks: 189 Mass of Base Peak: 442 Int: 426754 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
155	1,461	0.34	192	1,315	0.31	235	1,608	0.38
156	2,923	0.68	193	4,532	1.06	236	1,315	0.31
158	1,315	0.31	196	3,654	0.86	237	2,192	0.51
159	2,046	0.48	197	2,485	0.58	238	1,315	0.31
160	1,315	0.31	198	356,871	83.62	241	3,216	0.75
161	2,339	0.55	199	17,543	4.11	242	2,046	0.48
165	2,923	0.68	203	4,970	1.16	244	21,198	4.97
166	4,970	1.16	204	10,380	2.43	245	3,070	0.72
167	16,374	3.84	205	21,345	5.00	246	7,602	1.78
168	11,988	2.81	206	75,877	17.78	247	1,315	0.31
169	2,046	0.48	207	7,309	1.71	249	1,900	0.45
172	1,315	0.31	208	2,192	0.51	255	206,140	48.30
174	3,654	0.86	210	2,631	0.62	256	29,532	6.92
175	4,385	1.03	215	1,754	0.41	257	3,216	0.75
179	7,017	1.64	217	12,134	2.84	258	11,842	2.77
180	4,093	0.96	221	7,602	1.78	262	1,461	0.34
181	2,631	0.62	222	3,216	0.75	265	4,970	1.16
184	1,900	0.45	223	3,508	0.82	273	7,748	1.82
185	6,140	1.44	224	32,309	7.57	274	16,959	3.97
186	43,128	10.11	225	8,625	2.02	275	111,257	26.07
187	12,280	2.88	226	2,777	0.65	276	15,350	3.60
188	2,485	0.58	227	17,690	4.15	277	8,333	1.95
189	1,900	0.45	228	2,923	0.68	283	1,900	0.45
190	2,192	0.51	229	5,409	1.27	284	1,900	0.45
191	1,169	0.27	230	1,608	0.38	285	2,485	0.58

Mass List

File: E:\DFTC1204 Date: Dec-05-1991 08:21:53

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN

Scan No: 1094 Retention Time: 18:14 RIC: 3443458 Mass Range: 47 - 444

Peaks: 189 Mass of Base Peak: 442 Int: 426754 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
221	7,602	1.78	262	1,461	0.34	335	4,239	0.99
222	3,216	0.75	265	4,970	1.16	346	3,508	0.82
223	3,508	0.82	273	7,748	1.82	352	5,701	1.34
224	32,309	7.57	274	16,959	3.97	353	3,654	0.86
225	8,625	2.02	275	111,257	26.07	354	5,701	1.34
226	2,777	0.65	276	15,350	3.60	365	21,637	5.07
227	17,690	4.15	277	8,333	1.95	366	1,608	0.38
228	2,923	0.68	283	1,900	0.45	372	8,771	2.06
229	5,409	1.27	284	1,900	0.45	373	1,315	0.31
230	1,608	0.38	285	2,485	0.58	374	1,169	0.27
235	1,608	0.38	286	2,046	0.48	375	877	0.21
236	1,315	0.31	293	1,169	0.27	383	1,023	0.24
237	2,192	0.51	296	22,953	5.38	390	1,169	0.27
238	1,315	0.31	297	3,947	0.92	391	1,754	0.41
241	3,216	0.75	303	3,362	0.79	402	2,923	0.68
242	2,046	0.48	306	1,315	0.31	403	2,192	0.51
244	21,198	4.97	314	3,508	0.82	404	1,315	0.31
245	3,070	0.72	315	2,485	0.58	421	2,923	0.68
246	7,602	1.78	316	2,192	0.51	422	5,701	1.34
247	1,315	0.31	321	2,485	0.58	423	22,076	5.17
249	1,900	0.45	323	9,064	2.12	424	3,362	0.79
255	206,140	48.30	324	1,608	0.38	441	61,842	14.49
256	29,532	6.92	327	3,654	0.86	442	426,754	100.00
257	3,216	0.75	328	1,461	0.34	443	70,906	16.62
258	11,842	2.77	334	11,257	2.64	444	2,631	0.62

Passed

Chromatogram Plot

File: E:\DFTC1204 Date: Dec-05-1991 08:21:53


Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN

Scan No: 1270 Retention Time: 21:10 RIC: 68548 Mass Range: 45 - 357

Plotted: 1255 to 1270

Range: 1 to 1860

100% = 3443458



Chromatogram Plot

File: E:\DFTC1204 Date: Dec-05-1991 08:21:53


Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFIPP/ENDRIN

Scan No: 1290 Retention Time: 21:30 RIC: 77281 Mass Range: 45 - 430

Plotted: 1270 to 1290

Range: 1 to 1860

100% = 3443458



Chromatogram Plot

File: E:\DFTC1204 Date: Dec-05-1991 08:21:53


Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DFTPP/ENDRIN

Scan No: 1367 Retention Time: 22:47 RIC: 33536 Mass Range: 45 - 281

Plotted: 1337 to 1367

Range: 1 to 1860

100% = 3443458



Chromatogram Plot

File: E:\DDTC1204

Date: Dec-05-1991 09:01:42

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DDT

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0


Plotted: 1 to 1859

Range: 1 to 1859

100% = 4937973



Chromatogram Plot File: E:\DDTC1204 Date: Dec-05-1991 09:01:42
Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DDT
Scan No: 1260 Retention Time: 21:00 RIC: 26945 Mass Range: 47 - 281
Plotted: 1230 to 1260 Range: 1 to 1859 100% = 4937973



Chromatogram Plot

File: E:\DDTC1204 Date: Dec-05-1991 09:01:42


Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DDT

Scan No: 1300 Retention Time: 21:40 RIC: 39685 Mass Range: 45 - 282

Plotted: 1262 to 1300

Range: 1 to 1859

100% = 4937973



Chromatogram Plot

File: E:\DDTC1204

Date: Dec-05-1991 09:01:42

Comment: BOHLK; METHOD 525; INST 204; 12/04/01; DDT

Scan No: 1310

Retention Time: 21:50


RIC: 68568

Mass Range: 45 - 340

Plotted: 1300 to 1310

Range: 1 to 1859

100% = 4937973



Chromatogram Plot

File: E:\DFTA1205

Date: Dec-05-1991 16:45:33

Comment: BOHLK; METHOD 525; INST 204; 12/05/01; DFTPP/ENDRIN

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0

Plotted: 1 to 1860

Range: 1 to 1860

100% = 6055137



Chromatogram Plot

File: E:\DFTA1205 Date: Dec-05-1991 16:45:33

Comment: BOHLK; METHOD 525; INST 204; 12/05/01; DFTPP/ENDRIN


Scan No: 1080 Retention Time: 18:00 RIC: 55372 Mass Range: 47 - 281

Plotted: 1080 to 1120

Range: 1 to 1860

100% = 6055137

8.71%



Mass List

File: E:\DFTA1205 Date: Dec-05-1991 16:45:33

Comment: BOHLK; METHOD 525; INST 204; 12/05/01; DFTPP/ENDRIN

Scan No: 1094 Retention Time: 18:14 RIC: 6055137 Mass Range: 48 - 445

Peaks: 191 Mass of Base Peak: 442 Int: 834285 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
48	5,238	0.63	83	3,809	0.46	117	110,714	13.27
49	24,761	2.97	84	5,714	0.68	118	3,809	0.46
50	149,285	17.89	85	9,285	1.11	122	4,285	0.51
51	409,285	49.06	86	7,857	0.94	123	4,761	0.57
52	15,000	1.80	87	5,238	0.63	125	4,285	0.51
55	5,476	0.66	91	7,619	0.91	126	2,857	0.34
56	7,142	0.86	92	2,619	0.31	127	247,857	29.71
57	18,333	2.20	93	38,809	4.65	128	13,333	1.60
60	2,142	0.26	96	2,619	0.31	129	90,000	10.79
61	3,333	0.40	97	2,857	0.34	130	5,000	0.60
63	10,476	1.26	98	30,000	3.60	133	2,380	0.29
65	4,285	0.51	99	14,285	1.71	134	4,523	0.54
68	4,047	0.49	101	5,952	0.71	135	8,333	1.00
69	302,857	36.30	103	10,238	1.23	137	3,571	0.43
70	1,904	0.23	104	17,142	2.05	141	20,000	2.40
73	5,000	0.60	105	6,904	0.83	142	6,666	0.80
74	40,952	4.91	106	2,142	0.26	144	4,047	0.49
75	43,095	5.17	107	77,142	9.25	147	8,333	1.00
76	9,285	1.11	108	16,904	2.03	148	22,619	2.71
77	244,523	29.31	109	5,000	0.60	149	4,523	0.54
78	14,523	1.74	110	160,714	19.26	150	3,095	0.37
79	39,047	4.68	111	27,142	3.25	152	2,142	0.26
80	13,571	1.63	112	2,142	0.26	153	1,904	0.23
81	18,095	2.17	115	1,666	0.20	154	2,380	0.29
82	4,047	0.49	116	6,904	0.83	155	6,666	0.80

Mass List

File: E:\DFTA1205

Date: Dec-05-1991 16:45:33

Comment: BOHLK; METHOD 525; INST 204; 12/05/01; DFTPP/ENDRIN

Scan No: 1094

Retention Time: 18:14

RIC: 6055137

Mass Range: 48 - 445

Peaks: 191

Mass of Base Peak: 442

Int: 834285

Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
155	6,666	0.80	196	13,333	1.60	241	1,666	0.20
156	8,809	1.06	197	4,523	0.54	242	4,047	0.49
158	5,714	0.68	198	527,142	63.18	243	2,142	0.26
161	4,047	0.49	199	31,428	3.77	244	53,333	6.39
165	5,714	0.68	200	2,380	0.29	245	6,428	0.77
166	5,714	0.68	203	2,857	0.34	246	10,476	1.26
167	34,523	4.14	204	14,285	1.71	249	6,190	0.74
168	14,761	1.77	205	34,761	4.17	251	1,428	0.17
169	1,428	0.17	206	155,476	18.64	254	5,476	0.66
172	2,857	0.34	207	14,523	1.74	255	371,428	44.52
174	1,904	0.23	208	4,285	0.51	256	54,761	6.56
175	10,714	1.28	209	3,095	0.37	257	6,666	0.80
176	3,095	0.37	210	3,571	0.43	258	18,571	2.23
177	6,428	0.77	211	5,476	0.66	259	3,095	0.37
179	16,666	2.00	217	31,190	3.74	265	7,380	0.88
180	11,428	1.37	218	2,380	0.29	273	14,523	1.74
181	6,428	0.77	221	14,761	1.77	274	49,523	5.94
185	13,095	1.57	222	7,857	0.94	275	216,666	25.97
186	73,809	8.85	223	3,809	0.46	276	22,142	2.65
187	24,285	2.91	224	61,190	7.33	277	16,666	2.00
188	5,238	0.63	225	25,952	3.11	278	2,857	0.34
189	8,095	0.97	227	27,380	3.28	282	1,666	0.20
191	2,380	0.29	228	2,142	0.26	283	4,047	0.49
192	2,380	0.29	229	4,047	0.49	285	3,809	0.46
193	4,761	0.57	237	2,142	0.26	290	2,380	0.29

Mass List

File: E:\DFTA1205 Date: Dec-05-1991 16:45:33

Comment: BOHLK; METHOD 525; INST 204; 12/05/01; DFIPP/ENDRIN

Scan No: 1094 Retention Time: 18:14 RIC: 6055137 Mass Range: 48 - 445

Peaks: 191 Mass of Base Peak: 442 Int: 834285 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
222	7,857	0.94	275	216,666	25.97	352	8,571	1.03
223	3,809	0.46	276	22,142	2.65	353	6,428	0.77
224	61,190	7.33	277	16,666	2.00	354	11,428	1.37
225	25,952	3.11	278	2,857	0.34	355	3,333	0.40
227	27,380	3.28	282	1,666	0.20	356	2,857	0.34
228	2,142	0.26	283	4,047	0.49	365	34,285	4.11
229	4,047	0.49	285	3,809	0.46	366	4,523	0.54
237	2,142	0.26	290	2,380	0.29	370	2,857	0.34
241	1,666	0.20	291	1,904	0.23	372	19,047	2.28
242	4,047	0.49	293	7,619	0.91	373	1,904	0.23
243	2,142	0.26	296	51,666	6.19	384	2,619	0.31
244	53,333	6.39	297	10,714	1.28	402	5,000	0.60
245	6,428	0.77	302	2,142	0.26	403	9,761	1.17
246	10,476	1.26	303	5,714	0.68	404	2,857	0.34
249	6,190	0.74	306	2,142	0.26	413	2,857	0.34
251	1,428	0.17	315	4,761	0.57	421	4,285	0.51
254	5,476	0.66	316	2,142	0.26	422	4,523	0.54
255	371,428	44.52	323	16,428	1.97	423	35,000	4.20
256	54,761	6.56	324	3,809	0.46	424	4,761	0.57
257	6,666	0.80	327	6,904	0.83	425	2,142	0.26
258	18,571	2.23	334	13,571	1.63	441	133,333	15.98
259	3,095	0.37	335	3,809	0.46	442	834,285	100.00
265	7,380	0.88	336	2,380	0.29	443	146,428	17.55
273	14,523	1.74	341	2,142	0.26	444	11,190	1.34
274	49,523	5.94	346	5,476	0.66	445	1,428	0.17

Passed

Chromatogram Plot

File: E:\DFTA1205 Date: Dec-05-1991 16:45:33


Comment: BOHLK; METHOD 525; INST 204; 12/05/01; DFTPP/ENDRIN

Scan No: 1270 Retention Time: 21:10 RIC: 187173 Mass Range: 45 - 429


Plotted: 1255 to 1270

Range: 1 to 1860


100% = 6055137



Chromatogram Plot File: E:\DFTA1205 Date: Dec-05-1991 16:45:33
Comment: BOHLK; METHOD 525; INST 204; 12/05/01; DFTPP/ENDRIN
Scan No: 1300 Retention Time: 21:40 RIC: 140836 Mass Range: 45 - 355
Plotted: 1272 to 1300 Range: 1 to 1860 100% = 6055137



Chromatogram Plot File: E:\DFTA1205 Date: Dec-05-1991 16:45:33
Comment: BOHLK; METHOD 525; INST 204; 12/05/01; DFTPP/ENDRIN
Scan No: 1367 Retention Time: 22:47 RIC: 151130 Mass Range: 45 - 355
Plotted: 1337 to 1367 Range: 1 to 1860 100% = 6055137



Chromatogram Plot

File: E:\DDTA1205

Date: Dec-05-1991 17:20:39

Comment: BOHLK; METHOD 525; INST 204; 12/05/01; DDT

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0

Plotted: 1 to 1859

Range: 1 to 1859

100% = 3432219



Chromatogram Plot

File: E:\DDTA1205 Date: Dec-05-1991 17:20:39


Comment: BOHLK; METHOD 525; INST 204; 12/05/01; DDT

Scan No: 1269 Retention Time: 21:09 RIC: 84098 Mass Range: 45 - 356

Plotted: 1229 to 1269

Range: 1 to 1859

100% = 3432219



Chromatogram Plot

File: E:\DDTA1205 Date: Dec-05-1991 17:20:39


Comment: BOHLK; METHOD 525; INST 204; 12/05/01; DDT

Scan No: 1292 Retention Time: 21:32 RIC: 107895 Mass Range: 45 - 355

Plotted: 1262 to 1292

Range: 1 to 1859

100% = 3432219



Chromatogram Plot

File: E:\DDTA1205

Date: Dec-05-1991 17:20:39

Comment: BOHLK; METHOD 525; INST 204; 12/05/01; DDT

Scan No: 1320

Retention Time: 22:00

RIC: 99063

Mass Range: 45 - 355

Plotted: 1290 to 1320

Range: 1 to 1859

100% = 3432219

